

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Continuing Calib Date: 12/01/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 11:40

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.32       | 6.33      | 6.23      | 6.43  | 0.01       |
| Aroclor-1254-2       | (2) | 6.54       | 6.55      | 6.45      | 6.65  | 0.02       |
| Aroclor-1254-3       | (3) | 6.90       | 6.91      | 6.81      | 7.01  | 0.02       |
| Aroclor-1254-4       | (4) | 7.18       | 7.19      | 7.09      | 7.29  | 0.01       |
| Aroclor-1254-5       | (5) | 7.59       | 7.60      | 7.50      | 7.70  | 0.01       |
| Tetrachloro-m-xylene |     | 4.34       | 4.35      | 4.25      | 4.45  | 0.01       |
| Decachlorobiphenyl   |     | 9.91       | 9.93      | 9.83      | 10.03 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Continuing Calib Date: 12/01/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 11:40

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.51       | 5.53      | 5.43      | 5.63 | 0.02       |
| Aroclor-1254-2 (2)   | 5.66       | 5.67      | 5.57      | 5.77 | 0.01       |
| Aroclor-1254-3 (3)   | 6.06       | 6.08      | 5.98      | 6.18 | 0.02       |
| Aroclor-1254-4 (4)   | 6.29       | 6.30      | 6.20      | 6.40 | 0.01       |
| Aroclor-1254-5 (5)   | 6.71       | 6.72      | 6.62      | 6.82 | 0.01       |
| Tetrachloro-m-xylene | 3.64       | 3.65      | 3.55      | 3.75 | 0.01       |
| Decachlorobiphenyl   | 8.62       | 8.64      | 8.54      | 8.74 | 0.02       |