

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Continuing Calib Date: 12/01/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 17:13

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.32	6.33	6.23	6.43	0.01
Aroclor-1254-2	(2)	6.54	6.55	6.45	6.65	0.02
Aroclor-1254-3	(3)	6.90	6.91	6.81	7.01	0.02
Aroclor-1254-4	(4)	7.18	7.19	7.09	7.29	0.01
Aroclor-1254-5	(5)	7.59	7.60	7.50	7.70	0.01
Tetrachloro-m-xylene		4.34	4.35	4.25	4.45	0.01
Decachlorobiphenyl		9.91	9.93	9.83	10.03	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Continuing Calib Date: 12/01/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 17:13

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.51	5.53	5.43	5.63	0.02
Aroclor-1254-2 (2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1254-3 (3)	6.06	6.08	5.98	6.18	0.02
Aroclor-1254-4 (4)	6.29	6.30	6.20	6.40	0.01
Aroclor-1254-5 (5)	6.71	6.72	6.62	6.82	0.01
Tetrachloro-m-xylene	3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl	8.62	8.64	8.54	8.74	0.02