

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 01:00

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.48	5.49	5.39	5.59	0.02
Aroclor-1016-2	(2)	5.50	5.51	5.41	5.61	0.01
Aroclor-1016-3	(3)	5.56	5.57	5.47	5.67	0.01
Aroclor-1016-4	(4)	5.65	5.67	5.57	5.77	0.02
Aroclor-1016-5	(5)	5.95	5.96	5.86	6.06	0.01
Aroclor-1260-1	(1)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-2	(2)	7.31	7.33	7.23	7.43	0.02
Aroclor-1260-3	(3)	7.67	7.68	7.58	7.78	0.01
Aroclor-1260-4	(4)	7.89	7.90	7.80	8.00	0.01
Aroclor-1260-5	(5)	8.20	8.21	8.11	8.31	0.01
Tetrachloro-m-xylene		4.34	4.35	4.25	4.45	0.01
Decachlorobiphenyl		9.91	9.93	9.83	10.03	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3694

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 01:00

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1016-2	(2)	4.73	4.75	4.65	4.85	0.02
Aroclor-1016-3	(3)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-4	(4)	4.95	4.96	4.86	5.06	0.01
Aroclor-1016-5	(5)	5.16	5.18	5.08	5.28	0.02
Aroclor-1260-1	(1)	6.19	6.21	6.11	6.31	0.02
Aroclor-1260-2	(2)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-3	(3)	6.53	6.55	6.45	6.65	0.02
Aroclor-1260-4	(4)	7.00	7.02	6.92	7.12	0.02
Aroclor-1260-5	(5)	7.24	7.26	7.16	7.36	0.02
Tetrachloro-m-xylene		3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.62	8.64	8.54	8.74	0.02