

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115340.D **Time Analyzed:** 20:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	560.610	500.000	12.1
Aroclor-1016-2	5.501	5.408	5.608	553.190	500.000	10.6
Aroclor-1016-3	5.562	5.470	5.670	581.530	500.000	16.3
Aroclor-1016-4	5.658	5.566	5.766	568.810	500.000	13.8
Aroclor-1016-5	5.949	5.858	6.058	569.090	500.000	13.8
Aroclor-1260-1	7.063	6.971	7.171	561.040	500.000	12.2
Aroclor-1260-2	7.317	7.225	7.425	588.440	500.000	17.7
Aroclor-1260-3	7.674	7.582	7.782	574.140	500.000	14.8
Aroclor-1260-4	7.896	7.803	8.003	556.800	500.000	11.4
Aroclor-1260-5	8.202	8.110	8.310	555.290	500.000	11.1
Decachlorobiphenyl	9.919	9.830	10.030	56.520	50.000	13.0
Tetrachloro-m-xylene	4.340	4.248	4.448	55.950	50.000	11.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115340.D **Time Analyzed:** 20:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.628	4.828	534.270	500.000	6.9
Aroclor-1016-2	4.739	4.647	4.847	532.290	500.000	6.5
Aroclor-1016-3	4.913	4.821	5.021	537.270	500.000	7.5
Aroclor-1016-4	4.955	4.864	5.064	514.290	500.000	2.9
Aroclor-1016-5	5.168	5.076	5.276	583.870	500.000	16.8
Aroclor-1260-1	6.197	6.106	6.306	513.330	500.000	2.7
Aroclor-1260-2	6.385	6.294	6.494	492.990	500.000	-1.4
Aroclor-1260-3	6.537	6.446	6.646	499.440	500.000	-0.1
Aroclor-1260-4	7.008	6.917	7.117	495.150	500.000	-1.0
Aroclor-1260-5	7.249	7.158	7.358	500.640	500.000	0.1
Decachlorobiphenyl	8.630	8.541	8.741	54.680	50.000	9.4
Tetrachloro-m-xylene	3.644	3.548	3.748	53.810	50.000	7.6