

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115355.D **Time Analyzed:** 02:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	567.650	500.000	13.5
Aroclor-1016-2	5.501	5.408	5.608	561.100	500.000	12.2
Aroclor-1016-3	5.562	5.470	5.670	593.380	500.000	18.7
Aroclor-1016-4	5.658	5.566	5.766	585.130	500.000	17.0
Aroclor-1016-5	5.950	5.858	6.058	590.730	500.000	18.1
Aroclor-1260-1	7.063	6.971	7.171	585.730	500.000	17.1
Aroclor-1260-2	7.316	7.225	7.425	591.500	500.000	18.3
Aroclor-1260-3	7.674	7.582	7.782	586.850	500.000	17.4
Aroclor-1260-4	7.896	7.803	8.003	564.790	500.000	13.0
Aroclor-1260-5	8.202	8.110	8.310	563.830	500.000	12.8
Decachlorobiphenyl	9.920	9.830	10.030	57.680	50.000	15.4
Tetrachloro-m-xylene	4.341	4.248	4.448	56.970	50.000	13.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115355.D **Time Analyzed:** 02:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.628	4.828	541.660	500.000	8.3
Aroclor-1016-2	4.739	4.647	4.847	540.850	500.000	8.2
Aroclor-1016-3	4.914	4.821	5.021	543.770	500.000	8.8
Aroclor-1016-4	4.955	4.864	5.064	517.340	500.000	3.5
Aroclor-1016-5	5.167	5.076	5.276	524.620	500.000	4.9
Aroclor-1260-1	6.197	6.106	6.306	518.180	500.000	3.6
Aroclor-1260-2	6.385	6.294	6.494	496.140	500.000	-0.8
Aroclor-1260-3	6.537	6.446	6.646	489.450	500.000	-2.1
Aroclor-1260-4	7.008	6.917	7.117	499.390	500.000	-0.1
Aroclor-1260-5	7.249	7.158	7.358	502.960	500.000	0.6
Decachlorobiphenyl	8.630	8.541	8.741	55.920	50.000	11.8
Tetrachloro-m-xylene	3.644	3.548	3.748	54.950	50.000	9.9