

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115365.D **Time Analyzed:** 07:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	589.180	500.000	17.8
Aroclor-1016-2	5.500	5.408	5.608	582.400	500.000	16.5
Aroclor-1016-3	5.562	5.470	5.670	599.490	500.000	19.9
Aroclor-1016-4	5.657	5.566	5.766	579.130	500.000	15.8
Aroclor-1016-5	5.948	5.858	6.058	575.320	500.000	15.1
Aroclor-1260-1	7.062	6.971	7.171	549.570	500.000	9.9
Aroclor-1260-2	7.315	7.225	7.425	608.590	500.000	21.7
Aroclor-1260-3	7.673	7.582	7.782	584.510	500.000	16.9
Aroclor-1260-4	7.896	7.803	8.003	587.980	500.000	17.6
Aroclor-1260-5	8.202	8.110	8.310	584.600	500.000	16.9
Decachlorobiphenyl	9.918	9.830	10.030	59.370	50.000	18.7
Tetrachloro-m-xylene	4.340	4.248	4.448	58.810	50.000	17.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115365.D **Time Analyzed:** 07:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.628	4.828	561.220	500.000	12.2
Aroclor-1016-2	4.738	4.647	4.847	551.430	500.000	10.3
Aroclor-1016-3	4.913	4.821	5.021	557.300	500.000	11.5
Aroclor-1016-4	4.954	4.864	5.064	520.050	500.000	4.0
Aroclor-1016-5	5.167	5.076	5.276	628.340	500.000	25.7
Aroclor-1260-1	6.196	6.106	6.306	535.060	500.000	7.0
Aroclor-1260-2	6.384	6.294	6.494	518.730	500.000	3.7
Aroclor-1260-3	6.536	6.446	6.646	523.220	500.000	4.6
Aroclor-1260-4	7.007	6.917	7.117	513.500	500.000	2.7
Aroclor-1260-5	7.248	7.158	7.358	516.970	500.000	3.4
Decachlorobiphenyl	8.629	8.541	8.741	57.750	50.000	15.5
Tetrachloro-m-xylene	3.643	3.548	3.748	56.810	50.000	13.6