

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115388.D **Time Analyzed:** 09:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.476	5.387	5.587	499.290	500.000	-0.1
Aroclor-1016-2	5.498	5.408	5.608	490.880	500.000	-1.8
Aroclor-1016-3	5.560	5.470	5.670	508.310	500.000	1.7
Aroclor-1016-4	5.655	5.566	5.766	496.900	500.000	-0.6
Aroclor-1016-5	5.946	5.858	6.058	489.140	500.000	-2.2
Aroclor-1260-1	7.060	6.971	7.171	482.060	500.000	-3.6
Aroclor-1260-2	7.315	7.225	7.425	527.210	500.000	5.4
Aroclor-1260-3	7.672	7.582	7.782	509.680	500.000	1.9
Aroclor-1260-4	7.893	7.803	8.003	495.720	500.000	-0.9
Aroclor-1260-5	8.199	8.110	8.310	492.250	500.000	-1.6
Decachlorobiphenyl	9.914	9.830	10.030	56.740	50.000	13.5
Tetrachloro-m-xylene	4.337	4.248	4.448	55.780	50.000	11.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115388.D **Time Analyzed:** 09:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.628	4.828	468.150	500.000	-6.4
Aroclor-1016-2	4.736	4.647	4.847	473.080	500.000	-5.4
Aroclor-1016-3	4.910	4.821	5.021	460.820	500.000	-7.8
Aroclor-1016-4	4.952	4.864	5.064	453.950	500.000	-9.2
Aroclor-1016-5	5.165	5.076	5.276	500.340	500.000	0.1
Aroclor-1260-1	6.194	6.106	6.306	429.990	500.000	-14.0
Aroclor-1260-2	6.381	6.294	6.494	413.520	500.000	-17.3
Aroclor-1260-3	6.533	6.446	6.646	443.430	500.000	-11.3
Aroclor-1260-4	7.004	6.917	7.117	426.740	500.000	-14.7
Aroclor-1260-5	7.245	7.158	7.358	427.770	500.000	-14.4
Decachlorobiphenyl	8.626	8.541	8.741	55.970	50.000	11.9
Tetrachloro-m-xylene	3.641	3.548	3.748	53.620	50.000	7.2