

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115403.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.477	5.387	5.587	479.830	500.000	-4.0
Aroclor-1016-2	5.498	5.408	5.608	473.500	500.000	-5.3
Aroclor-1016-3	5.560	5.470	5.670	488.200	500.000	-2.4
Aroclor-1016-4	5.655	5.566	5.766	473.400	500.000	-5.3
Aroclor-1016-5	5.947	5.858	6.058	473.890	500.000	-5.2
Aroclor-1260-1	7.060	6.971	7.171	463.500	500.000	-7.3
Aroclor-1260-2	7.314	7.225	7.425	509.080	500.000	1.8
Aroclor-1260-3	7.672	7.582	7.782	489.930	500.000	-2.0
Aroclor-1260-4	7.893	7.803	8.003	476.400	500.000	-4.7
Aroclor-1260-5	8.199	8.110	8.310	477.550	500.000	-4.5
Decachlorobiphenyl	9.913	9.830	10.030	56.490	50.000	13.0
Tetrachloro-m-xylene	4.338	4.248	4.448	53.990	50.000	8.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115403.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.628	4.828	457.180	500.000	-8.6
Aroclor-1016-2	4.737	4.647	4.847	461.940	500.000	-7.6
Aroclor-1016-3	4.911	4.821	5.021	452.860	500.000	-9.4
Aroclor-1016-4	4.953	4.864	5.064	446.600	500.000	-10.7
Aroclor-1016-5	5.165	5.076	5.276	489.020	500.000	-2.2
Aroclor-1260-1	6.193	6.106	6.306	423.900	500.000	-15.2
Aroclor-1260-2	6.381	6.294	6.494	404.140	500.000	-19.2
Aroclor-1260-3	6.533	6.446	6.646	432.610	500.000	-13.5
Aroclor-1260-4	7.004	6.917	7.117	416.430	500.000	-16.7
Aroclor-1260-5	7.245	7.158	7.358	418.550	500.000	-16.3
Decachlorobiphenyl	8.625	8.541	8.741	54.730	50.000	9.5
Tetrachloro-m-xylene	3.642	3.548	3.748	52.230	50.000	4.5