

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL07 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115418.D **Time Analyzed:** 19:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.477	5.387	5.587	491.110	500.000	-1.8
Aroclor-1016-2	5.499	5.408	5.608	477.610	500.000	-4.5
Aroclor-1016-3	5.560	5.470	5.670	500.200	500.000	0.0
Aroclor-1016-4	5.656	5.566	5.766	486.130	500.000	-2.8
Aroclor-1016-5	5.948	5.858	6.058	483.490	500.000	-3.3
Aroclor-1260-1	7.061	6.971	7.171	476.420	500.000	-4.7
Aroclor-1260-2	7.315	7.225	7.425	521.470	500.000	4.3
Aroclor-1260-3	7.672	7.582	7.782	502.740	500.000	0.5
Aroclor-1260-4	7.894	7.803	8.003	485.260	500.000	-2.9
Aroclor-1260-5	8.200	8.110	8.310	486.050	500.000	-2.8
Decachlorobiphenyl	9.915	9.830	10.030	56.130	50.000	12.3
Tetrachloro-m-xylene	4.339	4.248	4.448	54.060	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL07 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115418.D **Time Analyzed:** 19:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.628	4.828	466.160	500.000	-6.8
Aroclor-1016-2	4.737	4.647	4.847	462.590	500.000	-7.5
Aroclor-1016-3	4.911	4.821	5.021	459.870	500.000	-8.0
Aroclor-1016-4	4.954	4.864	5.064	452.240	500.000	-9.6
Aroclor-1016-5	5.165	5.076	5.276	490.940	500.000	-1.8
Aroclor-1260-1	6.194	6.106	6.306	428.960	500.000	-14.2
Aroclor-1260-2	6.382	6.294	6.494	412.720	500.000	-17.5
Aroclor-1260-3	6.534	6.446	6.646	439.690	500.000	-12.1
Aroclor-1260-4	7.005	6.917	7.117	422.580	500.000	-15.5
Aroclor-1260-5	7.247	7.158	7.358	425.590	500.000	-14.9
Decachlorobiphenyl	8.627	8.541	8.741	55.170	50.000	10.3
Tetrachloro-m-xylene	3.643	3.548	3.748	52.250	50.000	4.5