

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL09 **Date Analyzed:** 12/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115479.D **Time Analyzed:** 16:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.473	5.387	5.587	484.180	500.000	-3.2
Aroclor-1016-2	5.496	5.408	5.608	474.620	500.000	-5.1
Aroclor-1016-3	5.557	5.470	5.670	478.340	500.000	-4.3
Aroclor-1016-4	5.653	5.566	5.766	463.340	500.000	-7.3
Aroclor-1016-5	5.944	5.858	6.058	482.710	500.000	-3.5
Aroclor-1260-1	7.057	6.971	7.171	443.840	500.000	-11.2
Aroclor-1260-2	7.311	7.225	7.425	492.280	500.000	-1.5
Aroclor-1260-3	7.669	7.582	7.782	480.450	500.000	-3.9
Aroclor-1260-4	7.890	7.803	8.003	468.020	500.000	-6.4
Aroclor-1260-5	8.197	8.110	8.310	460.700	500.000	-7.9
Decachlorobiphenyl	9.910	9.830	10.030	57.050	50.000	14.1
Tetrachloro-m-xylene	4.336	4.248	4.448	57.490	50.000	15.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3694
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL09 **Date Analyzed:** 12/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115479.D **Time Analyzed:** 16:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.715	4.628	4.828	470.130	500.000	-6.0
Aroclor-1016-2	4.734	4.647	4.847	480.510	500.000	-3.9
Aroclor-1016-3	4.908	4.821	5.021	479.840	500.000	-4.0
Aroclor-1016-4	4.951	4.864	5.064	478.130	500.000	-4.4
Aroclor-1016-5	5.162	5.076	5.276	485.820	500.000	-2.8
Aroclor-1260-1	6.191	6.106	6.306	422.330	500.000	-15.5
Aroclor-1260-2	6.379	6.294	6.494	392.810	500.000	-21.4
Aroclor-1260-3	6.531	6.446	6.646	402.610	500.000	-19.5
Aroclor-1260-4	7.001	6.917	7.117	401.900	500.000	-19.6
Aroclor-1260-5	7.243	7.158	7.358	402.560	500.000	-19.5
Decachlorobiphenyl	8.622	8.541	8.741	52.890	50.000	5.8
Tetrachloro-m-xylene	3.640	3.548	3.748	54.200	50.000	8.4