



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                        |                           |                   |
|-----------------|------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3694</u>           | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>Tetra Tech, EMI</u> | <b>Datafile :</b>         | <u>PO115331.D</u> |

| Lab Sample ID            | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits |      | RPD |
|--------------------------|-----------|-------|--------|-------|-----|-----|------|------|--------|------|-----|
|                          |           |       |        |       |     |     |      | Qual | Low    | High |     |
| PB170688BS<br>(Column 1) | AR1016    | 166.6 | 149    | ug/kg | 89  |     |      |      | 71     | 120  |     |
|                          | AR1260    | 166.6 | 135    | ug/kg | 81  |     |      |      | 65     | 130  |     |
| PB170688BS<br>(Column 2) | AR1016    | 166.6 | 138    | ug/kg | 83  |     |      |      | 71     | 120  |     |
|                          | AR1260    | 166.6 | 125    | ug/kg | 75  |     |      |      | 65     | 130  |     |