

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
 Data File : VL043279.D
 Acq On : 24 Nov 2025 13:11
 Operator : SY/MD
 Sample : VL1124ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1124ABL01

Quant Time: Nov 24 23:39:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	106137	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	277570	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	209566	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	151293	9.849	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.500%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043279.D
Acq On : 24 Nov 2025 13:11
Operator : SY/MD
Sample : VL1124ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1124ABL01

Quant Time: Nov 24 23:39:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043279.D
Acq: 24 Nov 2025 13:11




Instrument :
MSVOA_L
ClientSampleId :
VL1124ABL01


Tgt Ion: 49 Resp: 106137
Ion Ratio Lower Upper
49 100
128 49.2 23.0 69.0
130 63.9 29.9 89.7



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043279.D
Acq: 24 Nov 2025 13:11

Tgt Ion: 114 Resp: 277570
Ion Ratio Lower Upper
114 100
63 25.1 21.0 31.6
88 17.7 14.9 22.3









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 21
Delta R.T. 0.003 min
Lab File: VL043279.D
Acq: 24 Nov 2025 13:11

Instrument :
MSVOA_1
ClientSampleId :
VL1124ABL01

Tgt Ion:117 Resp: 209566
Ion Ratio Lower Upper
117 100
82 67.0 56.1 84.1
119 32.1 25.0 37.6



#68
1-Bromo-4-Fluorobenzene
Concen: 9.849 ppbv
RT: 10.370 min Scan# 3177
Delta R.T. 0.003 min
Lab File: VL043279.D
Acq: 24 Nov 2025 13:11

Tgt Ion: 95 Resp: 151293
Ion Ratio Lower Upper
95 100
174 65.8 50.2 75.2
175 5.1 4.1 6.1

