

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3698</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>GFE LLC</u>	Datafile :	<u>VL043260.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1120ABS01	Vinyl Chloride	10	9.50	ppbv	95			70	130	
	Heptane	10	8.90	ppbv	89			70	130	
	1,1-Dichloroethane	10	11.0	ppbv	110			70	130	
	Cyclohexane	10	9.60	ppbv	96			70	130	
	cis-1,2-Dichloroethene	10	10.0	ppbv	100			70	130	
	1,1,1-Trichloroethane	10	11.2	ppbv	112			70	130	
	2,2,4-Trimethylpentane	10	9.60	ppbv	96			70	130	
	Benzene	10	10.0	ppbv	100			70	130	
	Trichloroethene	10	10.6	ppbv	106			70	130	
	Toluene	10	9.60	ppbv	96			70	130	
	Tetrachloroethene	10	10.4	ppbv	104			70	130	
	Ethyl Benzene	10	10.4	ppbv	104			70	130	
	m/p-Xylene	20	21.4	ppbv	107			70	130	
	o-Xylene	10	10.7	ppbv	107			70	130	
	1,3,5-Trimethylbenzene	10	10.6	ppbv	106			70	130	
	1,2,4-Trimethylbenzene	10	10.6	ppbv	106			70	130	
	Naphthalene	10	9.80	ppbv	98			70	130	
	Hexane	10	9.50	ppbv	95			70	130	