

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

## LAB CHRONICLE

**OrderID:** Q3700  
**Client:** Ardmore Chemical  
**Contact:** Michael Sharphouse

**OrderDate:** 11/20/2025 4:11:00 PM  
**Project:** PVSC Monthly 2025  
**Location:** E11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3700-01	EFF-WW	WATER			11/20/25 13:00	11/24/25	11/24/25 13:03	11/20/25
			Cyanide	SM4500-CN C,E				
Q3700-04	EFF-WW	WATER			11/20/25 13:00		11/21/25 15:20	11/20/25
			BOD5	SM5210 B				
			TSS	SM2540 D			11/21/25 15:30	



# SAMPLE DATA

## Report of Analysis

Client:	Ardmore Chemical	Date Collected:	11/20/25 13:00
Project:	PVSC Monthly 2025	Date Received:	11/20/25
Client Sample ID:	EFF-WW	SDG No.:	Q3700
Lab Sample ID:	Q3700-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Cyanide	0.0012	U	1	0.0012	0.0050	mg/L	11/24/25 10:15	11/24/25 13:03	SM 4500-CN C-21 plus E-21

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client:	Ardmore Chemical	Date Collected:	11/20/25 13:00
Project:	PVSC Monthly 2025	Date Received:	11/20/25
Client Sample ID:	EFF-WW	SDG No.:	Q3700
Lab Sample ID:	Q3700-04	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
BOD5	432		1	0.20	2.00	mg/L		11/21/25 15:20	SM 5210 B-16
TSS	79.1		1	1.00	4.00	mg/L		11/21/25 15:30	SM 2540 D-20

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



# QC RESULT SUMMARY

### Initial and Continuing Calibration Verification

**Client:** Ardmore Chemical

**SDG No.:** Q3700

**Project:** PVSC Monthly 2025

**RunNo.:** LB138026

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.099	0.099	100	85-115	11/24/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	11/24/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	11/24/2025

### Initial and Continuing Calibration Blank Summary

**Client:** Ardmore Chemical

**SDG No.:** Q3700

**Project:** PVSC Monthly 2025

**RunNo.:** LB138026

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>ICB1</b> Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/24/2025
Sample ID: <b>CCB1</b> Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/24/2025
Sample ID: <b>CCB2</b> Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/24/2025



## Preparation Blank Summary

**Client:** Ardmore Chemical

**SDG No.:** Q3700

**Project:** PVSC Monthly 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: BOD5	<b>LB138012BL</b> mg/L	< 0.2000	0.2000	U	0.20	2.0	11/21/2025
Sample ID: TSS	<b>LB138018BL</b> mg/L	1	2.0000	J	1	4	11/21/2025
Sample ID: Cyanide	<b>PB170708BL</b> mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/24/2025

### Matrix Spike Summary

<b>Client:</b>	Ardmore Chemical	<b>SDG No.:</b>	Q3700
<b>Project:</b>	PVSC Monthly 2025	<b>Sample ID:</b>	Q3700-01
<b>Client ID:</b>	EFF-WWMS	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.041		0.0012	U	0.04	1	102		11/24/2025

## Matrix Spike Summary

<b>Client:</b>	Ardmore Chemical	<b>SDG No.:</b>	Q3700
<b>Project:</b>	PVSC Monthly 2025	<b>Sample ID:</b>	Q3700-01
<b>Client ID:</b>	EFF-WWMSD	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.041		0.0012	U	0.04	1	102		11/24/2025

## Duplicate Sample Summary

<b>Client:</b> Ardmore Chemical	<b>SDG No.:</b> Q3700
<b>Project:</b> PVSC Monthly 2025	<b>Sample ID:</b> Q3676-04
<b>Client ID:</b> OUTFALL-DSN-002DUP	<b>Percent Solids for Spike Sample:</b> 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	2890		2890		1	0.05		11/21/2025

## Duplicate Sample Summary

<b>Client:</b> Ardmore Chemical	<b>SDG No.:</b> Q3700
<b>Project:</b> PVSC Monthly 2025	<b>Sample ID:</b> Q3700-01
<b>Client ID:</b> EFF-WWDUP	<b>Percent Solids for Spike Sample:</b> 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0012	U	0.0012	U	1	0		11/24/2025

## Duplicate Sample Summary

<b>Client:</b> Ardmore Chemical	<b>SDG No.:</b> Q3700
<b>Project:</b> PVSC Monthly 2025	<b>Sample ID:</b> Q3700-01
<b>Client ID:</b> EFF-WWMSD	<b>Percent Solids for Spike Sample:</b> 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.041		0.041		1	0		11/24/2025

## Duplicate Sample Summary

<b>Client:</b> Ardmore Chemical <b>Project:</b> PVSC Monthly 2025 <b>Client ID:</b> EFF-WWDUP	<b>SDG No.:</b> Q3700 <b>Sample ID:</b> Q3700-04 <b>Percent Solids for Spike Sample:</b> 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	432		417		1	3.49		11/21/2025

### Laboratory Control Sample Summary

**Client:** Ardmore Chemical

**SDG No.:** Q3700

**Project:** PVSC Monthly 2025

**Run No.:** LB138012

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138012BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	11/21/2025



### Laboratory Control Sample Summary

**Client:** Ardmore Chemical

**SDG No.:** Q3700

**Project:** PVSC Monthly 2025

**Run No.:** LB138018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138018BS							
TSS	mg/L	550	591		107	1	90-110	11/21/2025

### Laboratory Control Sample Summary

**Client:** Ardmore Chemical

**SDG No.:** Q3700

**Project:** PVSC Monthly 2025

**Run No.:** LB138026

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170708BS							
Cyanide	mg/L	0.1	0.099		99	1	85-115	11/24/2025



# RAW DATA

# BOD5 LOG

ANALYST: rubin

SUPERVISOR: Iwona

Analysis Date: 11/21/2025

MANGANOUS SULFATE SOLUTION: W3103

Alkaline Iodide Azide: W3109

Sodium Thiosulfate, 0.025N: W3248

NaOH, 1N: WP113878

IncubatorID: INCUBATOR #3

GuageID: 0511064

Zero DO: WP115341

QC BATCH ID: LB138012

BOD Water: WP115797

Starch: W3149

Sulfuric acid, 1N: WP115342

POLYSEED: WP115800

GGA: WP115798

Chlorine Strips: W3155

pH Strips: W3241

Lab SampleID	Client ID	Bottle No.	VOL. ML	Initial Reading (ML)	Final Reading (ML)	Difference	Average
WINKLER 1	WINKLER 1	1	300	0.0	9.6	9.6	9.6
WINKLER 2	WINKLER 2	2	300	9.8	19.4	9.6	9.6

Meter Calibration1: 9.06 Zero DO Reading1: 0.10 mg/L (<=0.2 Criteria)

Barometric Pressure1: 760 mmHg DO Meter BOD fluid reading for winkler comparison: 9.68

## After Incubation

Meter Calibration2: 8.74 Zero DO Reading2: 0.10 mg/L (<=0.2 Criteria)

Barometric Pressure2: 755 mmHg

QC BATCH ID: LB138012

INCUBATOR TEMP IN(C): 20.0

INCUBATOR TEMP OUT(C): 20.0

TIME IN: 15:20

TIME OUT: 10:00

DATE IN: 11/21/2025

DATE OUT: 11/26/2025

Lab SampleID	Bottle No.	Check CL	Initial PH	Final PH	Temp °C	Sam Vol. (mL)	D.O.1 Initial	D.O.2 Final	Depletion	BOD Result (mg/L)	Avg Result (mg/L)	Comment
LB138012BL	1	No	6.56	N/A	20.70	300	9.67	9.65	0.02	0.02	0.02	
POLYSEED	1					10	9.55	6.57	2.98	0.6	0.61	
POLYSEED	2					15	9.53	4.99	4.54	0.61		
POLYSEED	3					20	9.50	3.18	6.32	0.63		
GGA	1					6	9.54	5.47	4.07	173	178	
GGA	2					6	9.56	5.41	4.15	177		
GGA	3					6	9.54	5.25	4.29	184		
Q3700-04	1	No	6.37	6.98	20.10	5	9.57	8.21	-	0	432	pH Adjusted
Q3700-04	2					10	9.55	7.36	2.19	474		
Q3700-04	3					20	9.50	6.25	3.25	396		
Q3700-04	4					30	9.48	4.61	4.87	426		
Q3700-04DUP	1	No	6.37	6.98	20.10	5	9.58	8.02	-	0	417.17	pH Adjusted
Q3700-04DUP	2					10	9.56	7.44	2.12	453		
Q3700-04DUP	3					20	9.52	6.36	3.16	382.5		
Q3700-04DUP	4					30	9.48	4.71	4.77	416		
Q3701-01	1	No	5.41	6.77	20.20	1	9.62	8.84	-	0	15330	pH Adjusted
Q3701-01	2					5	9.59	6.22	3.37	16560		
Q3701-01	3					10	9.56	4.25	5.31	14100		
Q3701-01	4					50	9.45	0.56	-	0		
Q3701-01	5					100	9.26	0.26	-	0		
Q3701-05	1	No	5.02	6.91	20.00	1	9.65	8.62	-	0	20445	pH Adjusted
Q3701-05	2					5	9.60	5.45	4.15	21240		
Q3701-05	3					10	9.54	2.38	7.16	19650		
Q3701-05	4					50	9.46	0.69	-	0		
Q3701-05	5					100	9.22	0.29	-	0		

NOTE: 2ml POLYSEED added to GGA and all the Samples, but not in Blank.

NOTE (For, CBOD5): 0.16 g Nitrification Inhibitor added to GGA and all the Samples, but not in Blank.

# WORKLIST(Hardcopy Internal Chain)

16138012

WorkList Name : BOD5-11-21.

WorkList ID : 193288

Department : Wet-Chemistry

Date : 11-21-2025 10:18:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3700-04	EFF-WW	Water	BOD5	Cool 4 deg C	ARDM01	E11	11/20/2025	SM5210 B
Q3701-01	EFFLUENT	Water	BOD5	Cool 4 deg C	HOLL01	D41	11/20/2025	SM5210 B
Q3701-05	INFLUENT	Water	BOD5	Cool 4 deg C	HOLL01	D41	11/20/2025	SM5210 B

Date/Time 11/21/2025 13:30  
 Raw Sample Received by: RM CWB  
 Raw Sample Relinquished by: [Signature]

Date/Time 11/21/2025 15:30  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: RM [Signature]

**TOTAL SUSPENDED SOLIDS - SM2540D**

**TEMP1 IN:** 104 °C 11/20/2025 13:30 **TEMP1 OUT:** 103 °C 11/20/2025 14:30  
**TEMP2 IN:** 104 °C 11/20/2025 15:30 **TEMP2 OUT:** 104 °C 11/20/2025 16:30  
**TEMP3 IN:** 104 °C 11/21/2025 15:30 **TEMP3 OUT:** 103 °C 11/21/2025 17:00  
**TEMP4 IN:** 104 °C 11/21/2025 17:30 **TEMP4 OUT:** 103 °C 11/21/2025 18:37

**SUPERVISOR:** Iwona  
**ANALYST:** jignesh  
**Date:** 11/20/2025  
**Run Number:** LB138018  
**BalanceID:** WC SC-5  
**OvenID:** WC OVEN-1  
**FilterID:** 17416528  
**ThermometerID:** WET OVEN#1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB138018BL	LB138018BL	1.4863	1.4863	100	1.4864	1.4864	1.4864	0.0001	1
2	LB138018BS	LB138018BS	1.3523	1.3523	100	1.4114	1.4114	1.4114	0.0591	591
3	Q3675-02	COMP	1.4834	1.4835	200	1.5110	1.5110	1.5110	0.0275	137.5
4	Q3676-01	OUTFALL-DSN-001	1.4880	1.4881	1000	1.5979	1.5980	1.5980	0.1099	109.9
5	Q3676-04	OUTFALL-DSN-002	1.4946	1.4946	300	2.3626	2.3626	2.3626	0.8680	2893.3
6	Q3676-04DUP	OUTFALL-DSN-002DUP	1.4835	1.4835	300	2.3519	2.3519	2.3519	0.8684	2894.7
7	Q3690-01	RW8-SP100-20251118	1.4791	1.4791	1800	1.4799	1.4799	1.4799	0.0008	0.4
8	Q3690-02	RW8-SP303-20251118	1.4987	1.4987	1200	1.4989	1.4989	1.4989	0.0002	0.2
9	Q3700-04	EFF-WW	1.4737	1.4737	1000	1.5528	1.5528	1.5528	0.0791	79.1
10	Q3701-01	EFFLUENT	1.4977	1.4977	40	1.5391	1.5391	1.5391	0.0414	1035
11	Q3701-04	AERATION	1.5013	1.5013	30	1.5472	1.5472	1.5472	0.0459	1530
12	Q3703-01	SW-2	1.4802	1.4803	650	1.5178	1.5178	1.5178	0.0375	57.7
13	Q3704-01	SW-2	1.4103	1.4104	950	1.4139	1.4139	1.4139	0.0035	3.7

**TOTAL SUSPENDED SOLIDS - SM2540D**

**SUPERVISOR:** Iwona

**ANALYST:** jignesh

**Date:** 11/20/2025

**Run Number:** LB138018

**BalanceID:** WC SC-5

**OvenID:** WC OVEN-1

**FilterID:** 17416528

**ThermometerID:** WET OVEN#1

**TEMP1 IN:** 104 °C 11/20/2025 13:30 **TEMP1 OUT:** 103 °C 11/20/2025 14:30  
**TEMP2 IN:** 104 °C 11/20/2025 15:30 **TEMP2 OUT:** 104 °C 11/20/2025 16:30  
**TEMP3 IN:** 104 °C 11/21/2025 15:30 **TEMP3 OUT:** 103 °C 11/21/2025 17:00  
**TEMP4 IN:** 104 °C 11/21/2025 17:30 **TEMP4 OUT:** 103 °C 11/21/2025 18:37

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)  
 B = Final Empty Dish Weight (g)  
 C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)  
 D = Weight (g)

Weight (g) = C - B

Result mg/L =  $\frac{D}{A} \times 1000 \times 1000$



# WORKLIST(Hardcopy Internal Chain)

11/21/25 13:10

WorkList Name : tss q3703      WorkList ID : 193293      Department : Wet-Chemistry      Date : 11-21-2025 12:53:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3675-02	COMP	Water	TSS	Cool 4 deg C	ARAM01	E11	11/19/2025	SM2540 D
Q3676-01	OUTFALL-DSN-001	Water	TSS	Cool 4 deg C	TRIS02	E11	11/18/2025	SM2540 D
Q3676-04	OUTFALL-DSN-002	Water	TSS	Cool 4 deg C	TRIS02	E11	11/18/2025	SM2540 D
Q3690-01	RW8-SP100-20251118	Water	TSS	Cool 4 deg C	TETR06	A11	11/18/2025	SM2540 D
Q3690-02	RW8-SP303-20251118	Water	TSS	Cool 4 deg C	TETR06	A11	11/18/2025	SM2540 D
Q3700-04	EFF-WW	Water	TSS	Cool 4 deg C	ARDM01	E11	11/20/2025	SM2540 D
Q3701-01	EFFLUENT	Water	TSS	Cool 4 deg C	HOLL01	D41	11/20/2025	SM2540 D
Q3701-04	AERATION	Water	TSS	Cool 4 deg C	HOLL01	D41	11/20/2025	SM2540 D
Q3703-01	SW-2	Water	TSS	Cool 4 deg C	ATGG01	D41	11/18/2025	SM2540 D
Q3704-01	SW-2	Water	TSS	Cool 4 deg C	ATGG01	D31	11/18/2025	SM2540 D
Q3705-01	MH 11242025	Water	TSS	Cool 4 deg C	EURO03	D41	11/21/2025	SM2540 D

11/21/25 13:10

11/22/2025

Date/Time 11/21/25 13:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 11/21/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

16138026

Test results

Aquakem 7.2AQ1

Page:

Alliance Technical Group  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

11/24/2025 13:03

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.590	0.0	0.083	
ICB1	0.836	0.0	0.002	
CCV1	246.140	0.0	0.205	
CCB1	0.413	0.0	0.001	
RL CHECK	4.948	0.0	0.005	
PB170708BL	0.472	0.0	0.001	98% (50-150)
PB170708BS	99.438	0.0	0.083	
MIDPB170708	247.155	0.0	0.206	98% (90-110) 11/24/2025 RM
Q3700-01	0.573	0.0	0.001	
Q3700-01DUP	0.561	0.0	0.001	
Q3700-01MS	41.084	0.0	0.035	
Q3700-01MSD	41.314	0.0	0.035	
CCV2	260.373	0.0	0.217	
CCB2	0.856	0.0	0.002	

N 14  
Mean 74.482  
SD 101.8397  
CV% 136.73

Aquakem v. 7.2AQ1

Results from time period:

Mon Nov 24 12:28:18 2025

Mon Nov 24 13:03:06 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.8561	µg/l	11/24/2025 10:00:18	
5.0PPBCN	A	Total CN	P	5.4993	µg/l	11/24/2025 10:00:19	
10PPBCN	A	Total CN	P	10.1742	µg/l	11/24/2025 10:00:20	
50PPBCN	A	Total CN	P	47.955	µg/l	11/24/2025 10:00:21	
100PPBCN	A	Total CN	P	100.4082	µg/l	11/24/2025 10:00:22	
250PPBCN	A	Total CN	P	249.9857	µg/l	11/24/2025 10:00:23	
500PPBCN	A	Total CN	P	500.1216	µg/l	11/24/2025 10:00:24	
ICV1	S	Total CN	P	98.5898	µg/l	11/24/2025 12:28:18	
ICB1	S	Total CN	P	0.8359	µg/l	11/24/2025 12:28:21	
CCV1	S	Total CN	P	246.1404	µg/l	11/24/2025 12:28:22	
CCB1	S	Total CN	P	0.4128	µg/l	11/24/2025 12:28:25	
RL CHECK	S	Total CN	P	4.9485	µg/l	11/24/2025 12:28:27	
PB170708BL	S	Total CN	P	0.4723	µg/l	11/24/2025 12:35:51	
PB170708BS	S	Total CN	P	99.4378	µg/l	11/24/2025 12:35:53	
MIDPB170708	S	Total CN	P	247.155	µg/l	11/24/2025 12:35:55	
Q3700-01	S	Total CN	P	0.5726	µg/l	11/24/2025 13:03:01	
Q3700-01DUP	S	Total CN	P	0.5613	µg/l	11/24/2025 13:03:02	
Q3700-01MS	S	Total CN	P	41.0843	µg/l	11/24/2025 13:03:03	
Q3700-01MSD	S	Total CN	P	41.3143	µg/l	11/24/2025 13:03:04	
CCV2	S	Total CN	P	260.3726	µg/l	11/24/2025 13:03:05	
CCB2	S	Total CN	P	0.8564	µg/l	11/24/2025 13:03:06	

=====

Calibration results      Aquakem 7.2AQ1      Page:      11/24/2025 10:01

Alliance Technical Group  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM      Instrument ID : Konelab

11/24/2025 10:01


Test      Total CN

Accepted      11/24/2025 10:01

Factor      1206  
Bias      0.001

Coeff. of det.      0.999974

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	0.8561	0.0000	-
2	5.0PPBCN	0.005	5.4993	5.0000	10.0
3	10PPBCN	0.009	10.1742	10.0000	1.7
4	50PPBCN	0.041	47.9550	50.0000	-4.1
5	100PPBCN	0.084	100.4082	100.0000	0.4
6	250PPBCN	0.208	249.9857	250.0000	0.0
7	500PPBCN	0.416	500.1216	500.0000	0.0

11/24/2025  
RM

SOP ID : MSM4500-CN C,E-Cyanide-13

SDG No : N/A

Start Digest Date: 11/24/2025 Time : 10:15 Temp : 124 °C

Matrix : WATER

End Digest Date: 11/24/2025 Time : 11:45 Temp : 128 °C

Pippete ID : WC

Balance ID : N/A

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : MC-1,MC-2

Filter paper ID : N/A

 Prep Technician Signature: 

Weigh By : N/A

pH Meter ID : N/A

 Supervisor Signature: 


Standard Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP113838
MS/MSD SPIKE SOL.	0.40ML	WP113837
PBW	50.0ML	W3112
RL CHECK	50.0ML	WP115812
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP113836
50% v/v H2SO4	5.0ML	WP112826
51% w/v MgCL2	2.0ML	WP112827
pH Paper 0-14	N/A	W3241
Nitrate/Nitrite Strip	N/A	W3182
Lead Acetate strip	N/A	W3134
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
S0	S0	N/A	N/A
S5.0	S5.0	N/A	N/A
S10.0	S10.0	N/A	N/A
S100.0	S100.0	N/A	N/A
S250.0	S250.0	N/A	N/A
S500.0	S500.0	N/A	N/A
ICV	ICV	0.5ML	W3012
ICB	ICB	N/A	N/A
CCV	CCV	N/A	N/A
CCB	CCB	N/A	N/A
Midrange	Midrange	2.5ML	WP113837
HIGHSTD	HIGHSTD	N/A	N/A
LOWSTD	LOWSTD	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/24/2025 11:55		RM (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB170708BL	PBW708	50	50	>12	Negative	Negative	Negative	N/A	N/A
PB170708BS	LCS708	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3700-01DUP	EFF-WWDUP	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3700-01MS	EFF-WWMS	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3700-01MSD	EFF-WWMSD	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3700-01	EFF-WW	50	50	>12	Negative	Negative	Negative	N/A	N/A

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : cn w q3700

WorkList ID : 193305

Department : Distillation

Date : 11-24-2025 08:05:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3700-01	EFF-WW	Water	Cyanide	1:1 NaOH to pH >12	ARDM01	E11	11/20/2025	SM4500-CN C

Date/Time 11/24/2025 08:30  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

Date/Time 11/24/2025 10:45  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

**Instrument ID:** DO METER

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138012**

Review By	rubina	Review On	11/26/2025 10:37:20 AM
Supervise By	Iwona	Supervise On	11/26/2025 10:37:45 AM
SubDirectory	LB138012	Test	BOD5
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP115797,W3149,WP115342,W3103,W3109,W3248,WP115800,WP115798,WP113878		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB138012BL	LB138012BL	MB	11/21/25 15:20		RUBINA	OK
2	LB138012BS	LB138012BS	LCS	11/21/25 15:20		RUBINA	OK
3	Q3700-04	EFF-WW	SAM	11/21/25 15:20		RUBINA	OK
4	Q3700-04DUP	EFF-WWDUP	DUP	11/21/25 15:20		RUBINA	OK
5	Q3701-01	EFFLUENT	SAM	11/21/25 15:20		RUBINA	OK
6	Q3701-05	INFLUENT	SAM	11/21/25 15:20		RUBINA	OK



**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138018**

Review By	jignesh	Review On	11/25/2025 11:13:12 AM
Supervise By	Iwona	Supervise On	11/25/2025 11:37:09 AM
SubDirectory	LB138018	Test	TSS
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB138018BL	LB138018BL	MB	11/21/25 15:30		jignesh	OK
2	LB138018BS	LB138018BS	LCS	11/21/25 15:30	55 MG W3186 + 100 ML W3112	jignesh	OK
3	Q3675-02	COMP	SAM	11/21/25 15:30		jignesh	OK
4	Q3676-01	OUTFALL-DSN-001	SAM	11/21/25 15:30		jignesh	OK
5	Q3676-04	OUTFALL-DSN-002	SAM	11/21/25 15:30		jignesh	OK
6	Q3676-04DUP	OUTFALL-DSN-002D	DUP	11/21/25 15:30		jignesh	OK
7	Q3690-01	RW8-SP100-2025111	SAM	11/21/25 15:30		jignesh	OK
8	Q3690-02	RW8-SP303-2025111	SAM	11/21/25 15:30		jignesh	OK
9	Q3700-04	EFF-WW	SAM	11/21/25 15:30		jignesh	OK
10	Q3701-01	EFFLUENT	SAM	11/21/25 15:30		jignesh	OK
11	Q3701-04	AERATION	SAM	11/21/25 15:30		jignesh	OK
12	Q3703-01	SW-2	SAM	11/21/25 15:30		jignesh	OK
13	Q3704-01	SW-2	SAM	11/21/25 15:30		jignesh	OK

**Instrument ID:** KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138026**

Review By	rubina	Review On	11/24/2025 4:13:03 PM
Supervise By	Iwona	Supervise On	11/25/2025 9:45:28 AM
SubDirectory	LB138026	Test	Cyanide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	WP115807,WP115808,WP115809,WP115810,WP115811,WP115812,WP115813		
ICV Standard	W3012		
CCV Standard	WP115808		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115815		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	11/24/25 10:00		rubina	OK
2	5.0PPBCN	5.0PPBCN	CAL2	11/24/25 10:00		rubina	OK
3	10PPBCN	10PPBCN	CAL3	11/24/25 10:00		rubina	OK
4	50PPBCN	50PPBCN	CAL4	11/24/25 10:00		rubina	OK
5	100PPBCN	100PPBCN	CAL5	11/24/25 10:00		rubina	OK
6	250PPBCN	250PPBCN	CAL6	11/24/25 10:00		rubina	OK
7	500PPBCN	500PPBCN	CAL7	11/24/25 10:00		rubina	OK
8	ICV1	ICV1	ICV	11/24/25 12:28		rubina	OK
9	ICB1	ICB1	ICB	11/24/25 12:28		rubina	OK
10	CCV1	CCV1	CCV	11/24/25 12:28		rubina	OK
11	CCB1	CCB1	CCB	11/24/25 12:28		rubina	OK
12	RL	RL	LOQ	11/24/25 12:28		rubina	OK
13	PB170708BL	PB170708BL	MB	11/24/25 12:35		rubina	OK
14	PB170708BS	PB170708BS	LCS	11/24/25 12:35		rubina	OK
15	MIDPB170708	MIDPB170708	SAM	11/24/25 12:35		rubina	OK
16	Q3700-01	EFF-WW	SAM	11/24/25 13:03		rubina	OK
17	Q3700-01DUP	EFF-WWDUP	DUP	11/24/25 13:03		rubina	OK
18	Q3700-01MS	EFF-WWMS	MS	11/24/25 13:03		rubina	OK

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138026**

Review By	rubina	Review On	11/24/2025 4:13:03 PM
Supervise By	Iwona	Supervise On	11/25/2025 9:45:28 AM
SubDirectory	LB138026	Test	Cyanide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	WP115807,WP115808,WP115809,WP115810,WP115811,WP115812,WP115813		
ICV Standard	W3012		
CCV Standard	WP115808		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115815		

19	Q3700-01MSD	EFF-WWMSD	MSD	11/24/25 13:03		rubina	OK
20	CCV2	CCV2	CCV	11/24/25 13:03		rubina	OK
21	CCB2	CCB2	CCB	11/24/25 13:03		rubina	OK

## Prep Standard - Chemical Standard Summary

**Order ID :** Q3700

**Test :** BOD5,Cyanide,TSS

**Prepbatch ID :** PB170708,

**Sequence ID/Qc Batch ID:** LB138012, LB138018, LB138026,

**Standard ID :**

WP112826, WP112827, WP113836, WP113837, WP113838, WP113878, WP114324, WP115157, WP115342, WP115797, WP115798, WP115800, WP115806, WP115807, WP115808, WP115809, WP115810, WP115811, WP115812, WP115813, WP115815,

**Chemical ID :**

M6041, M6151, M6186, W2653, W2654, W2668, W3012, W3019, W3103, W3109, W3112, W3113, W3139, W3149, W3152, W3182, W3203, W3214, W3224, W3241, W3248, W3252, W3253,

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1714	Sulfuric Acid, 50% (v/v)	<a href="#">WP112826</a>	04/25/2025	10/25/2025	Rubina Mughal	None	None	Iwona Zarych
								04/25/2025

**FROM** 1000.00000ml of M6041 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3214	Magnesium Chloride For Cyanide 2.5M(51%W/V)	<a href="#">WP112827</a>	04/25/2025	10/25/2025	Rubina Mughal	WETCHEM_S CALE_8 (WC SC-7)	None	Iwona Zarych
								04/25/2025

**FROM** 500.00000ml of W3112 + 510.00000gram of W3152 = Final Quantity: 1000.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
11	Sodium hydroxide absorbing solution 0.25 N	<a href="#">WP113836</a>	07/08/2025	12/31/2025	Rubina Mughal	WETCHEM_S CALE_8 (WC SC-7)	None	Iwona Zarych  07/08/2025
<b><u>FROM</u></b> 21.00000L of W3112 + 210.00000gram of W3113 = Final Quantity: 21.000 L								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3850	Cyanide MS-MSD spiking solution, 5PPM	<a href="#">WP113837</a>	07/08/2025	11/30/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3	Iwona Zarych
(WC)								
<u>FROM</u>	1.00000ml of W3214 + 199.00000ml of WP113836 = Final Quantity: 200.000 ml							

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3371	Cyanide LCS Spike Solution, 5PPM	<a href="#">WP113838</a>	07/08/2025	12/24/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 07/08/2025

**FROM** 1.00000ml of W3224 + 199.00000ml of WP113836 = Final Quantity: 200.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1571	Sodium hydroxide, 1N	<a href="#">WP113878</a>	07/09/2025	12/31/2025	Iwona Zarych	WETCHEM_SCALE_7 (WC SC-6)	None	Jignesh Parikh 07/09/2025

**FROM** 4.00000gram of W3113 + 96.00000ml of W3112 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
607	PYRIDINE-BARBITURIC ACID	<a href="#">WP114324</a>	08/19/2025	02/17/2026	Rubina Mughal	WETCHEM_SCALE_5 (WCS-5)	Glass Pipette-A	Jignesh Parikh 08/19/2025
<b><u>FROM</u></b>	145.00000ml of W3112 + 15.00000gram of W3203 + 15.00000ml of M6151 + 75.00000ml of W3019 = Final Quantity: 250.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
539	CN BUFFER	<a href="#">WP115157</a>	10/10/2025	12/03/2025	Rubina Mughal	WETCHEM_SCALE_8 (WC SC-7)	None	Iwona Zarych 10/14/2025
<u>FROM</u>	138.00000gram of W2668 + 862.00000ml of W3112 = Final Quantity: 1000.000 ml							





<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1841	Sulfuric Acid, 1N	<a href="#">WP115342</a>	10/27/2025	04/27/2026	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 10/27/2025
<b><u>FROM</u></b> 2.80000ml of M6186 + 97.20000ml of W3112 = Final Quantity: 100.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
127	BOD Dilution fluid	<a href="#">WP115797</a>	11/21/2025	11/22/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 11/21/2025
<b><u>FROM</u></b> 18.00000L of W3112 + 3.00000PILLOW of W3253 = Final Quantity: 18.000 L								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
129	Glutamic acid-glucose mix for BOD	<a href="#">WP115798</a>	11/21/2025	11/22/2025	Rubina Mughal	WETCHEM_SCALE_7 (WCS-6)	None	Iwona Zarych 11/21/2025
<b><u>FROM</u></b> 0.15000gram of W2653 + 0.15000gram of W2654 + 1000.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
128	polyseed seed control	<a href="#">WP115800</a>	11/21/2025	11/22/2025	Rubina Mughal	None	None	Iwona Zarych 11/21/2025
<b><u>FROM</u></b> 1.00000PILLOW of W3252 + 300.00000ml of WP115797 = Final Quantity: 300.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3456	Cyanide Intermediate Working Std, 5PPM	<a href="#">WP115806</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Iwona Zarych
<b>FROM</b> 0.25000ml of W3214 + 49.75000ml of WP113836 = Final Quantity: 50.000 ml <div></div>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4	Calibration standard 500 ppb	<a href="#">WP115807</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 11/24/2025
<b><u>FROM</u></b> 45.00000ml of WP113836 + 5.00000ml of WP115806 = Final Quantity: 50.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3761	Calibration-CCV CN Standard 250 ppb	<a href="#">WP115808</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Iwona Zarych
<p><b><u>FROM</u></b>      2.50000ml of WP115806 + 47.50000ml of WP113836 = Final Quantity: 50.000 ml</p>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
6	Calibration Standard 100 ppb	<a href="#">WP115809</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Iwona Zarych
(WC)								
<u>FROM</u>	1.00000ml of WP115806 + 49.00000ml of WP113836 = Final Quantity: 50.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
7	Calibration Standard 50 ppb	<a href="#">WP115810</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Iwona Zarych
<p><b>FROM</b> 0.50000ml of WP115806 + 49.50000ml of WP113836 = Final Quantity: 50.000 ml</p>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
8	Calibration Standard 10 ppb	<a href="#">WP115811</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 11/24/2025
<u>FROM</u>	1.00000ml of WP115807 + 49.00000ml of WP113836 = Final Quantity: 50.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
9	Calibration Standard 5 ppb	<a href="#">WP115812</a>	11/24/2025	11/25/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Iwona Zarych
<b>FROM</b> 0.50000ml of WP115807 + 49.50000ml of WP113836 = Final Quantity: 50.000 ml <div></div>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipettelD</u>	<u>Supervised By</u>
167	0 ppb CN calibration std	<a href="#">WP115813</a>	11/24/2025	11/25/2025	Rubina Mughal	None	None	Iwona Zarych 11/24/2025
<b><u>FROM</u></b> 50.00000ml of WP113836 = Final Quantity: 50.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1582	Chloramine T solution, 0.014M	<a href="#">WP115815</a>	11/24/2025	11/25/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	Glass Pipette-A	Iwona Zarych  11/24/2025
<u>FROM</u>	0.08000gram of W3139 + 20.00000ml of W3112 = Final Quantity: 20.000 ml							

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	23D2462010	03/20/2028	08/16/2024 / mohan	08/16/2024 / mohan	M6041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	22G2862015	02/17/2026	02/18/2025 / Sagar	01/15/2025 / Sagar	M6151

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	23D2462010	07/12/2026	08/13/2025 / Sagar	08/06/2025 / Sagar	M6186

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AC156212500 / GLUTAMIC ACID BIOCHEM REG, 250G	A0405990	01/24/2030	01/24/2020 / apatel	01/24/2020 / apatel	W2653

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	D16-500 / DEXTROSE ANHYDROUS ACS REAGENT, 500G(New)	186122A	01/24/2030	01/24/2020 / apatel	01/24/2020 / apatel	W2654

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3818-5 / SODIUM PHOSPHATE, MONOBAS/HYD, CRYST, ACS, 2.5 KG	0000225799	12/03/2025	04/05/2021 / Alexander	02/10/2020 / apatel	W2668



## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
EPA	/ ICV-CN	ICV6-400	12/31/2025	01/08/2025 / lwona	02/20/2020 / lwona	W3012

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
SIGMA ALDRICH	270970-1L / Pyridine 1L	SHBQ2113	04/03/2028	04/03/2023 / lwona	04/03/2023 / lwona	W3019

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	4620-32 / MANGANOUS SULFATE SOLUTION-364	2403J02	03/31/2026	04/22/2024 / lwona	04/22/2024 / lwona	W3103

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL04100-4 / Alkaline Iodide Azide, 1 L	1405D67	04/30/2026	05/23/2024 / lwona	05/23/2024 / lwona	W3109

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-7 / Sodium Hydroxide Pellets 12 Kg	23B1556310	12/31/2025	07/08/2024 / lwona	07/08/2024 / lwona	W3113

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	JTE494-6 / CHLORAMINE-T BAKER 250GM	10239484	09/09/2029	09/09/2024 / lwona	09/09/2024 / lwona	W3139

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL70850-8 / Starch Solution, 4L	4408P62	08/31/2026	10/16/2024 / lwona	10/16/2024 / lwona	W3149

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	01237-10KG / Megnasium Chloride Hexahydrate ACS 10KG	002126-2019-201	11/25/2029	11/25/2024 / lwona	11/25/2024 / lwona	W3152

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	470112-662 / TEST STRIPES, NITRATE/NITRITE, PK50	436101	04/30/2027	08/05/2025 / lwona	02/26/2025 / lwona	W3182

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EM-BX0035-3 / Barbituric Acid, 100 gms	WXBF3271V	05/16/2029	04/21/2025 / lwona	04/21/2025 / lwona	W3203

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	RC2543-4 / CYANIDE STD 1000PPM 4OZ	1505H73	11/30/2025	05/21/2025 / lwona	05/21/2025 / lwona	W3214

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	LC135457 / Cyanide Standard, 1000 PPM, Second Source	45060288	12/24/2025	07/07/2025 / lwona	07/07/2025 / lwona	W3224

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140444 / TEST PAPERS,PH 0-14,.5 SENSI,100PK	10BDH15251	04/30/2029	10/02/2025 / lwona	10/02/2025 / lwona	W3241

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE	250904J	02/28/2027	10/03/2025 / lwona	10/03/2025 / lwona	W3248

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	136742-80 / POLYSEED	072505	05/31/2027	10/31/2025 / lwona	10/31/2025 / lwona	W3252

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
HACH	1486266 / BOD Nutrient Buffer Pillows, 6 mL concentrate to make 6 L, 50/pk	A5219	08/31/2030	11/19/2025 / lwona	11/19/2025 / lwona	W3253



**ACROS ORGANICS** part of Thermo Fisher Scientific





**Version** 0  
**Molecular weight** 147.13  
**Molecular formula** C5 H9 N O4  
**CAS No** 56-86-0  
**Linear formula** HO2CCH2CH2CH(NH2)CO2H  
**Flash point (°C)**

## Certificate of Analysis

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Acros Organics expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to human or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

<b>Catalog Number</b>	15621	<b>Quality Test / Release Date</b>	13 March 2019
<b>Lot Number</b>	A0405990	<b>Suggested Retest Date</b>	March 2022
<b>Description</b>	L(+)-Glutamic acid, 99%		
<b>Country of Origin</b>	CHINA		
<b>Declaration of Origin</b>	plant		

<b>Origin Comment</b>	The product is made by fermentation of sugar molasses
-----------------------	---

Result Name	Specifications	Test Value
Appearance (Color)	White	White
Appearance (Form)	Powder	Powder
Infrared spectrum	Conforms	Conforms
Titration with NaOH	98.5 to 100.5 % (On dried substance)	99.32 % (On dried substance)
Loss on drying	≤0.5 % (105°C, 3 hrs)	0.002 % (105°C, 3 hrs)
Heavy metals (as Pb)	≤10 ppm	≤10 ppm
Sulfated ash	≤0.1 %	0.08 %
Other amino acids	not detectable	not detectable
Specific optical rotation	+30.5° to +32.5° (20°C, 589 nm) (on dried substance)	+32° (20°C, 589 nm) (on dried substance)
Specific optical rotation	(c=10, 2N HCl)	(c=10, 2N HCl)
Chloride (Cl)	≤200 ppm	≤200 ppm
Iron (Fe)	≤30 ppm	≤10 ppm
Sulfate (SO4)	≤300 ppm	≤200 ppm
Ammonium (NH4)	≤200 ppm	≤200 ppm
Arsenic oxide (As2O3)	≤1 ppm	≤1 ppm



A handwritten signature in black ink, which appears to read "L. Van den Broek".

L. Van den Broek, QA Manager

Issued: 24 January 2020

Acros Organics

ENA23, zone 1, nr 1350, Janssen Pharmaceuticaaan 3a, B-2440 Geel, Belgium

Tel +32 14/57.52.11 - Fax +32 14/59.34.34 Internet: <http://www.acros.com>

1 Reagent Lane, Fair Lawn, NJ 07410, USA Fax 201-796-1329

W3019  
rec 4/3/23

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: [www.sigmaaldrich.com](http://www.sigmaaldrich.com)Email USA: [techserv@sial.com](mailto:techserv@sial.com)Outside USA: [eurtechserv@sial.com](mailto:eurtechserv@sial.com)

## Certificate of Analysis

Product Name:

Pyridine - anhydrous, 99.8%

Product Number:

270970

Batch Number:

SHBQ2113

Brand:

SIAL

CAS Number:

110-86-1

MDL Number:

MFCD00011732

Formula:


C<sub>5</sub>H<sub>5</sub>N

Formula Weight:

79.10 g/mol

Quality Release Date:

15 DEC 2022



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Purity (GC)	≥ 99.75 %	99.99 %
Water (by Karl Fischer)	≤ 0.003 %	0.002 %
Residue on Evaporation	≤ 0.0005 %	< 0.0001 %

  
Larry Coers, Director  
Quality Control  
Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





## Certificate of Analysis

1 Reagent Lane  
Fair Lawn, NJ 07410  
201.796.7100 tel  
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System  
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	D16	Quality Test / Release Date	03/19/2019
Lot Number	186122A		
Description	DEXTROSE, ANHYDROUS, A.C.S.		
Country of Origin	United States	Suggested Retest Date	Mar/2022
Chemical Origin	Organic - Plant		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		
Chemical Comment			

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	White, granular powder
TITRATABLE ACID	MEQ/G	<= 0.002	<0.002
STARCH		= PASS TEST	pass test
SPECIFIC ROTATION @ 25 C	DEGREES (+ OR -)	Inclusive Between +52.5 - +53.0	53.0
SULFATE & SULFITE	%	<= 0.005	<0.005
IRON (Fe)	ppm	<= 5	<5
CHLORIDE	%	<= 0.01	<0.01
IGNITION RESIDUE	%	<= 0.02	<0.02
IDENTIFICATION	PASS/FAIL	= PASS TEST	pass test
HEAVY METALS (as Pb)	ppm	<= 5	<5
LOSS ON DRYING @ 105 C	%	<= 0.2	<0.2
INSOLUBLE MATTER	%	<= 0.005	0.002

*Jerisa Bailey-Wyche*

Quality Assurance Specialist - Certificate of Analysis Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.  
If there are any questions with this certificate, please call at (800) 227-6701.

\*Based on suggested storage condition.



R: 02/20/20  
53

Instructions for QATS Reference Material: *Inorganic ICV Solutions*

For ICP-MS use: dilute the ICV1 concentrate 50-fold with 1% (v/v) nitric acid; pipet 2 mL of the concentrate into a 100 mL volumetric flask and dilute to volume with 1% (v/v) nitric acid.

**ICV5-0415**

For the cold vapor analysis of mercury by AA: dilute the ICV5 concentrate 100-fold with 2% (v/v) nitric acid; pipet 1 mL of the concentrate into a 100 mL volumetric flask and dilute to volume with 2% (v/v) nitric acid. The ICV5 concentrate is prepared in 0.05% (w/v)  $K_2Cr_2O_7$  and 5% (v/v) nitric acid.

**ICV6-0400**

For the analysis of cyanide: dilute the ICV6 concentrate 100-fold with Type II water; pipet 1 mL of the concentrate into a 100 mL volumetric flask and dilute to volume with Type II water. Distill this solution along with the samples before analysis. The cyanide concentrate is prepared from  $K_3Fe(CN)_6$ , Type II water, and 0.1 % sodium hydroxide, and will decompose rapidly if exposed to light.

**NOTE:** USE TYPE II WATER AND HIGH-PURITY ACIDS FOR ALL DILUTIONS.

**(D) CERTIFIED CONCENTRATIONS OF QATS ICV1, ICV5, AND ICV6 SOLUTIONS**

ICV1-1014		
Element	Concentration (µg/L) (after 10-fold dilution)	Concentration (µg/L) (after 50-fold dilution)
Al	2520	504
Sb	1010	202
As	997	199
Ba	518	104
Be	514	103
Cd	514	103
Ca	10000	2000
Cr	517	103
Co	521	104
Cu	505	101
Fe	10100	2020
Pb	1030	206
Mg	5990	1198
Mn	524	105
Ni	525	105
K	9940	1988
Se	1030	206
Ag	252	50
Na	10100	2020
Tl	1040	208
V	504	101
Zn	1010	202

ICV5-0415		ICV6-0400	
Element	Concentration (µg/L) (after 100-fold dilution)	Analyte	Concentration (µg/L) (after 100-fold dilution)
Hg	4.0	CN <sup>-</sup>	99

W3011  
W3012  
W3013  
W3014  
W3015



Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

avantor™



M 6041-4b  
MS

Material No.: 9673-33  
Batch No.: 23D2462010  
Manufactured Date: 2023-03-22  
Retest Date: 2028-03-20  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS – Assay (H <sub>2</sub> SO <sub>4</sub> )	95.0 – 98.0 %	96.1 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Substances Reducing Permanganate (as SO <sub>2</sub> )	≤ 2 ppm	< 2 ppm
Ammonium (NH <sub>4</sub> )	≤ 1 ppm	1 ppm
Chloride (Cl)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO <sub>3</sub> )	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities – Aluminum (Al)	≤ 30.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	8.5 ppb
Trace Impurities – Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities – Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities – Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gold (Au)	≤ 10.0 ppb	0.5 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities – Iron (Fe)	≤ 50.0 ppb	1.3 ppb
Trace Impurities – Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities – Magnesium (Mg)	≤ 7.0 ppb	0.8 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities – Nickel (Ni)	≤ 2.0 ppb	0.3 ppb
Trace Impurities – Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se)	≤ 50.0 ppb	< 0.1 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	31.5 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

 **avantor™**



Material No.: 9673-33  
Batch No.: 23D2462010

Test	Specification	Result
Trace Impurities – Sodium (Na)	$\leq 500.0$ ppb	5.4 ppb
Trace Impurities – Strontium (Sr)	$\leq 5.0$ ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	$\leq 5.0$ ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	$\leq 5.0$ ppb	0.4 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantor™**



M6151

R → 11/15/25

Material No.: 9530-33  
Batch No.: 22G2862015  
Manufactured Date: 2022-06-15  
Retest Date: 2027-06-14  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.9 %
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.191
ACS – Bromide (Br)	≤ 0.005 %	< 0.005 %
ACS – Extractable Organic Substances	≤ 5 ppm	< 1 ppm
ACS – Free Chlorine (as Cl <sub>2</sub> )	≤ 0.5 ppm	< 0.5 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.05 ppm	< 0.03 ppm
Sulfate (SO <sub>4</sub> )	≤ 0.5 ppm	< 0.3 ppm
Sulfite (SO <sub>3</sub> )	≤ 0.8 ppm	0.3 ppm
Ammonium (NH <sub>4</sub> )	≤ 3 ppm	< 1 ppm
Trace Impurities – Arsenic (As)	≤ 0.010 ppm	< 0.003 ppm
Trace Impurities – Aluminum (Al)	≤ 10.0 ppb	1.3 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 3.0 ppb
Trace Impurities – Barium (Ba)	≤ 1.0 ppb	0.2 ppb
Trace Impurities – Beryllium (Be)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Bismuth (Bi)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 20.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	163.0 ppb
Trace Impurities – Chromium (Cr)	≤ 1.0 ppb	0.7 ppb
Trace Impurities – Cobalt (Co)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gallium (Ga)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0 ppb
Trace Impurities – Gold (Au)	≤ 4.0 ppb	0.6 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 15 ppb	6 ppb

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantorsm**



Material No.: 9530-33  
Batch No.: 22G2862015

Test	Specification	Result
Trace Impurities – Lead (Pb)	≤ 1.0 ppb	< 0.5 ppb
Trace Impurities – Lithium (Li)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Magnesium (Mg)	≤ 10.0 ppb	2.9 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	0.1 ppb
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 3.0 ppb
Trace Impurities – Nickel (Ni)	≤ 4.0 ppb	< 0.3 ppb
Trace Impurities – Niobium (Nb)	≤ 1.0 ppb	0.8 ppb
Trace Impurities – Potassium (K)	≤ 9.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se), For Information Only		< 1.0 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	0.5 ppb
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	2.3 ppb
Trace Impurities – Strontium (Sr)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Tantalum (Ta)	≤ 1.0 ppb	1.6 ppb
Trace Impurities – Thallium (Tl)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	4.0 ppb
Trace Impurities – Titanium (Ti)	≤ 1.0 ppb	1.5 ppb
Trace Impurities – Vanadium (V)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.8 ppb
Trace Impurities – Zirconium (Zr)	≤ 1.0 ppb	0.3 ppb

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA–ANALYZED® Reagent  
For Trace Metal Analysis



Material No.: 9530-33  
Batch No.: 22G2862015

Test	Specification	Result
------	---------------	--------

For Laboratory, Research, or Manufacturing Use  
Product Information (not specifications):  
Appearance (clear, fuming liquid)  
Meets ACS Specifications  
Storage Condition: Store below 25 °C.

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.  
Jamie Ethier  
Vice President Global Quality

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

avantor™



M6186

Recieve Date :- 08/06/25

Material No.: 9673-33  
Batch No.: 23D2462010  
Manufactured Date: 2023-03-22  
Retest Date: 2028-03-20  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS - Assay (H <sub>2</sub> SO <sub>4</sub> )	95.0 - 98.0 %	96.1 %
Appearance	Passes Test	Passes Test
ACS - Color (APHA)	≤ 10	5
ACS - Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS - Substances Reducing Permanganate (as SO <sub>2</sub> )	≤ 2 ppm	< 2 ppm
Ammonium (NH <sub>4</sub> )	≤ 1 ppm	1 ppm
Chloride (Cl)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO <sub>3</sub> )	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities - Aluminum (Al)	≤ 30.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	8.5 ppb
Trace Impurities - Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities - Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities - Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities - Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities - Gold (Au)	≤ 10.0 ppb	0.5 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities - Iron (Fe)	≤ 50.0 ppb	1.3 ppb
Trace Impurities - Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities - Magnesium (Mg)	≤ 7.0 ppb	0.8 ppb
Trace Impurities - Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities - Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities - Nickel (Ni)	≤ 2.0 ppb	0.3 ppb
Trace Impurities - Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities - Selenium (Se)	≤ 50.0 ppb	< 0.1 ppb
Trace Impurities - Silicon (Si)	≤ 100.0 ppb	31.5 ppb
Trace Impurities - Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium



Material No.: 9673-33  
Batch No.: 23D2462010

Test	Specification	Result
Trace Impurities – Sodium (Na)	$\leq 500.0$ ppb	5.4 ppb
Trace Impurities – Strontium (Sr)	$\leq 5.0$ ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	$\leq 5.0$ ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	$\leq 5.0$ ppb	0.4 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.  
Jamie Ethier  
Vice President Global Quality

Sodium Phosphate, Monobasic, Monohydrate,  
Crystal  
BAKER ANALYZED® A.C.S. Reagent

(sodium dihydrogen phosphate, monohydrate)



Material No.: 3818-05  
Batch No.: 0000225799  
Manufactured Date: 2018/12/05  
Retest Date: 2025/12/03  
Revision No: 1

## Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay ( $\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$ )	98.0 – 102.0 %	99.5
pH of 5% Solution at 25°C	4.1 – 4.5	4.3
Insoluble Matter	$\leq 0.01$ %	$< 0.01$
Chloride (Cl)	$\leq 5$ ppm	$< 5$
ACS – Sulfate ( $\text{SO}_4$ )	$\leq 0.003$ %	$< 0.003$
Calcium (Ca)	$\leq 0.005$ %	$< 0.005$
Potassium (K)	$\leq 0.01$ %	$< 0.01$
Heavy Metals (as Pb)	$\leq 0.001$ %	$< 0.001$
Trace Impurities – Iron (Fe)	$\leq 0.001$ %	$< 0.001$

For Laboratory, Research or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: IN  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700





# Certificate of Analysis

**Manganous Sulfate Solution, 364 g/L****Lot Number:** 2403J02**Product Number:** 4620**Manufacture Date:** MAR 15, 2024**Expiration Date:** MAR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Manganous Sulfate Monohydrate	10034-96-5	Reagent
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result
Appearance	Pink liquid	Passed
Assay (by Refractive Index)	360-368 g/L	367 g/L

Specification	Reference
Manganous Sulfate Solution	ASTM (D 888 A)
Manganous Sulfate Solution	ASTM (D 888 A)
Manganous Sulfate Solution	APHA (4500-O E)
Manganous Sulfate Solution	APHA (4500-O F)
Manganous Sulfate Solution	APHA (4500-O D)
Manganous Sulfate Solution	APHA (4500-O E)
Manganous Sulfate Solution	APHA (4500-O F)
Manganous Sulfate Solution	APHA (4500-O D)
Manganous Sulfate Solution	APHA (4500-O C)
Manganous Sulfate Solution	APHA (4500-O C)
Manganous Sulfate Solution	EPA (360.2)
Manganous Sulfate Solution	EPA (360.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4620-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (03/15/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

# Certificate of Analysis

## Alkaline-Iodide-Azide, Pomeroy Formulation for Dissolved Oxygen (DO) Analysis

**Lot Number:** 1405D67

**Product Number:** 535

**Manufacture Date:** APR 05, 2024

**Expiration Date:** APR 2026

This solution is intended for use with samples with high Dissolved Oxygen content (above 15 mg/L) and for samples with high concentrations of organic material.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Iodide	7681-82-5	ACS
Sodium Hydroxide	1310-73-2	ACS
Sodium Azide	26628-22-8	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed
Free Iodine	To Pass Test	Passed

Specification	Reference
Alkaline Iodide-Sodium Azide Solution II	ASTM (D 888 A)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
535-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/05/2024)  
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# Certificate of Analysis



## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

Chemical Formula: NaOH  
Molecular Weight: 40  
CAS #: 1310-73-2  
Appearance:

Manufacture Date: 12/14/2022  
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

Chemical Formula: NaOH  
Molecular Weight: 40  
CAS #: 1310-73-2  
Appearance:

Manufacture Date: 12/14/2022  
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

W3139 Received on 9/9/24 by IZ

Product No.: A12044  
Product: Chloramine-T trihydrate, 98%  
Lot No.: 10239484

Appearance:	White powder
Melting Point:	166°C(dec)
Assay (Iodometric titration):	100.5%
Identification (FTIR):	Conforms

Order our products online [thermofisher.com/chemicals](https://thermofisher.com/chemicals)

**This document has been electronically generated and does not require a signature.**

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.



# Certificate of Analysis

## Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Starch, soluble	9005-84-9	ACS
Salicylic Acid	69-72-7	ACS

Test	Specification	Result
Appearance	White translucent liquid	Passed
Suitability for Use	Colorless (Iodine absent) - Blue (Iodine present)	Passed

Specification	Reference
Starch Solution	APHA (4500-S2- F)
Starch Indicator Solution	APHA (4500-CI B)
Starch Indicator	APHA (4500-SO32- B)
Starch indicator solution	APHA (2350 B)
Starch indicator solution	APHA (2350 E)
Starch Solution	APHA (510 B)
Starch Solution	APHA (5530 C)
Starch Indicator	APHA (4500-CI C)
Starch Indicator	EPA (345.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8000-1	4 L natural poly	24 months
8000-16	500 mL natural poly	24 months
8000-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)  
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# *Chem-Impex International, Inc.*

---

**Tel: (630) 766-2112****E-mail: sales@chemimpex.com****Shipping and Correspondence:**

935 Dillon Drive

Wood Dale, IL 60191

**Fax: (630) 766-2218****Web site: www.chemimpex.com****Manufacturing site:**

825 Dillon Drive

Wood Dale, IL 60191

---

## *Certificate of Analysis*

<b>Catalogue Number</b>	01237
<b>Lot Number</b>	002126-2019-201
<b>Product</b>	<b>Magnesium chloride hexahydrate</b>

Magnesium chloride•6H<sub>2</sub>O

<b>CAS Number</b>	7791-18-6
<b>Molecular Formula</b>	MgCl <sub>2</sub> •6H <sub>2</sub> O

<b>Molecular Weight</b>	203.3
-------------------------	-------

---

<b>Appearance</b>	White crystals
<b>Solubility</b>	167 g in 100 mL water
<b>Melting Point</b>	~ 115 °C
<b>Heavy Metals</b>	4.393 ppm
<b>Anion</b>	Nitrate (NO <sub>3</sub> ) : < 0.001% Phosphate (PO <sub>4</sub> ) : < 5 ppm Sulfate (SO <sub>4</sub> ) : < 0.002%
<b>Cation</b>	Ammonium (NH <sub>4</sub> ) : < 0.002% Barium (Ba) : 0.005% Calcium (Ca) : 0.01% Iron (Fe) : 4.5 ppm Manganese (Mn) : 0.624 ppm Potassium (K) : 0.004% Sodium (Na) : 0.000003% Strontium (Sr) : 0.005%
<b>Insoluble material</b>	0.0021%
<b>Assay by titration</b>	100.83%
<b>Grade</b>	ACS reagent
<b>Storage</b>	Store at RT

## ***Certificate of Analysis***

**Catalog Number: 01237**

**Lot Number: 002126-2019-201**

---

**Remarks**

See material safety data sheet for additional information

For laboratory use only

**The foregoing is a copy of the Certificate of Analysis as provided by our supplier**

A handwritten signature in black ink, appearing to read 'Bala Kumar', with a stylized flourish at the end.


**Bala Kumar**  
**Quality Control Manager**

## Certificate of Analysis

Product Name:

Barbituric acid - ReagentPlus®, 99%

Product Number: 185698  
Batch Number: WXBFB3271V  
Brand: SIAL  
CAS Number: 67-52-7  
Formula: C<sub>4</sub>H<sub>4</sub>N<sub>2</sub>O<sub>3</sub>  
Formula Weight: 128.09 g/mol  
Quality Release Date: 16 MAY 2024



Test	Specification	Result
Appearance (Colour)	White to Off-White	White
Appearance (Form)	Powder	Powder
Infrared spectrum	Conforms to Structure	Conforms
Purity (Titration by NaOH)	98.5 - 101.5 %	100.4 %
GC (area %)	≥ 98 %	100 %
VPCT		



Kang Chen  
Quality Manager  
Wuxi, China CN

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



# Certificate of Analysis

## Cyanide Standard, 1000 ppm CN<sup>-</sup>

**Lot Number:** 1505H73

**Product Number:** 2543

**Manufacture Date:** MAY 08, 2025

**Expiration Date:** NOV 2025

This standard is prepared using accurate volumetric techniques from material that has been assayed against Silver Nitrate solution certified traceable to NIST Standard Reference Material 999. The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is the combined uncertainty based on the stability of the assayed Potassium Cyanide, and the uncertainty in the mass and volume measurements.

Use 0.16% (w/v) (0.04 N) Sodium Hydroxide or 0.225 % (w/v) (0.04 N) Potassium Hydroxide to make dilutions of this standard. Restandardize weekly if extreme accuracy is required.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Cyanide	151-50-8	ACS
Sodium Hydroxide	1310-73-2	Reagent (from ACS)

Test	Specification	Result
Appearance	Colorless liquid	Passed
Cyanide (CN <sup>-</sup> )	995-1005 ppm	1000 ppm

Specification	Reference
Stock Standard Cyanide Solution	APHA (4500-CN- F)
Stock Cyanide Solution	APHA (4500-CN- E)
Stock Cyanide Solution	APHA (4500-CN- K)
Stock Cyanide Solution	APHA (4500-CN- H)
Cyanide Reference Solution (1000 mg/L)	EPA (SW-846) (7.3.3.2)
Cyanide Calibration Stock Solution (1,000 mg/L CN <sup>-</sup> )	EPA (SW-846) (9213)
Stock Cyanide Solution	EPA (335.3)
Stock Cyanide Solution	EPA (335.2)
Cyanide Solution Stock	ASTM (D 4282)
Simple Cyanide Solution, Stock (1.0 g/L CN <sup>-</sup> )	ASTM (D 4374)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2543-16	500 mL amber poly	6 months
2543-32	1 L amber poly	6 months
2543-4	120 mL amber poly	6 months

**Recommended Storage:** 2°C - 8°C (36°F - 46°F)



Ernest Mahan (05/08/2025)  
Plant Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Part of TCP Analytical Group

Jackson's Pointe Commerce Park- Building 1000  
1010 Jackson's Pointe Court, Zelienople, PA 16063

## Certificate of Analysis

### Cyanide Standard 1000 ppm (1ml = 1mg CN)

Product Code: **LC13545**

Manufacture Date: June 25, 2025

Lot Number: **45060288**

Expiration Date: December 24, 2025

Test	Specification	Result
Appearance (clarity)	clear solution	clear solution
Appearance (color)	colorless	colorless
Concentration (CN)	0.990 - 1.010mg/mL	1.000mg/mL
Concentration (CN)	990 - 1,010ppm	1,000ppm
Traceable to NIST SRM	Report	999b

**Intended Use** - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

**Storage Information** - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

**Instructions for Handling and Use** - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

**Preparation** - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST

\*The suffix of the product code may differ from what is on your product label. The suffix will designate the size and be associated with a numeric digit(s). Visit [LabChem.com](http://LabChem.com) for more information\*

Suffix	1	2	3/3S/36/36S	4/4C	5	6	7	8	9	20	44	200	246	486
Size	500mL org	1L or 1kg	2.5L/2.5L Coated/6x2.5L/6x2.5L Coated	4L	20L	10L	125mL	25g	100g	20x20mL	4x4L	200L	24x6mL	48x6mL

*Michael Monteleone*

Michael Monteleone  
Chemistry Supervisor - Quality Control  
20250703 15:30:45ahoffman-0-0

ISO9001:2015 Registration #0306-01

# Certificate of Analysis

## Sodium Thiosulfate, 0.0250 Normal (N/40)

**Lot Number:** 250904J

**Product Number:** 7900

**Manufacture Date:** SEP 03, 2025

**Expiration Date:** FEB 2027

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Organic Preservative	Proprietary	
Sodium Carbonate	497-19-8	ACS
Sodium Thiosulfate Pentahydrate	10102-17-7	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.02499-0.02501 N at 20°C	0.02501 N at 20°C	136

Specification	Reference
Standard Sodium Thiosulfate Solution, 0.0250 N	APHA (4500-S2- F)
Standard Sodium Thiosulfate Titrant	APHA (4500-O D)
Standard Sodium Thiosulfate Titrant	APHA (4500-O E)
Standard Sodium Thiosulfate Titrant	APHA (4500-O F)
Standard Sodium Thiosulfate Titrant, 0.025 N	APHA (4500-CI B)
Standard Sodium Thiosulfate Titrant	APHA (4500-O C)
Standard Sodium Thiosulfate Titrant, 0.025 M	APHA (5530 C)
Standard Sodium Thiosulfate Solution (0.025 N)	EPA (SW-846) (9031)
Standard Sodium Thiosulfate solution (0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7900-1	4 L natural poly	18 months
7900-16	500 mL natural poly	18 months
7900-32	1 L natural poly	18 months


**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (09/03/2025)  
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.





## CERTIFICATE OF ANALYSIS

PO BOX 130549 Spring, TX 77393  
Phone: (281) 298-9410 Fax: (281) 298-9411

**FINISHED PRODUCT, LOT NUMBER, MFG. /EXP DATE:**

PolySeed® • Part No. P-110 • Lot 072505 • Mfg. Date: 05/2025 • Exp. Date: 05/2027

**FORMULATION:**

The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.

**VIABLE COUNT, FINAL TEST RESULT:**

The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of  $4.00 \times 10^9$  cfu/g.

**GLUCOSE/GLUTAMIC-ACID RESULTS:**

Tested results within acceptable range 198 +/- 30.5 mg/L (167.5 - 228.5 mg/L). GGA Lot# 43100020 – Average Test Result: 203

See [www.polyseed.com](http://www.polyseed.com) for details.

**SEED CONTROL FACTOR:**

Tested results within acceptable range 0.6 – 1.0 see [www.polyseed.com](http://www.polyseed.com) for details

**SALMONELLA TEST RESULT:**

The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.

The purpose of this document is to ensure that the Finished Product conforms to the above specifications.

Signature: \_\_\_\_\_

*Quality Control Department*

Date: 05/07/2025

POLYSEED.Ref.1.19

Revised Jan 25



An ISO 9001 Certified Company

P.O. Box 389  
Loveland, CO 80539  
(970) 669-3050

## *Certificate of Analysis*

*This is a Component of 1486266 / LOT A5219*

**PRODUCT:** BOD Nutrient Buffer Pillows

**PRODUCT NUMBER:** 1486227

**LOT NUMBER:** A5219

**MANUFACTURE DATE:** 08/26/2025

**DATE OF ANALYSIS:** 09/15/2025

TEST	SPECIFICATIONS	RESULTS
Ammonia Concentration of a diluted pillow	0.57 to 0.79 ppm	0.581
Calcium Concentration of a diluted pillow	0.93 to 1.29 ppm	1.050
Iron Concentration of a diluted pillow	0.27 to 0.36 ppm	0.323
Magnesium Concentration of a diluted pillow	0.35 to 0.48 ppm	0.400
Phosphorus Concentration of a diluted pillow	7.6 to 10.3 ppm	8.85
pH in a 6 L of DI water	7.1 to 7.6 ph	7.20
Five Day Change in Dissolved Oxygen Concentration	-0.2 to 0.2 ppm	0.15
Sterility	To Pass	Passed

The expiration date is Aug 2030

Certified by: *Scott Als*

Analytical Service Chemist



# SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: ARMORE INC  
ADDRESS: 29 RIVERSIDE AVE, Bldg #14  
CITY: Newark, NJ STATE: NJ ZIP: 07104  
ATTENTION: M Sharpshaw  
PHONE: 973 481 2406 FAX: 973 481 2637

CLIENT PROJECT INFORMATION

PROJECT NAME: PUSC MONTHLY  
PROJECT NO.: LOCATION:  
PROJECT MANAGER:  
e-mail:  
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:  
ADDRESS:  
CITY STATE: ZIP:  
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS\*  
HARDCOPY (DATA PACKAGE): DAYS\*  
EDD: STANDARD DAYS\*  
\*TO BE APPROVED BY CHEMTECH  
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP  
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other  
☐ EDD FORMAT

VOC  
CN  
BOD/TSS  
SUOC  
METALS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	EFF WASTE WATER	WW		X	11/29/25	1:00 PM		X	X									
2.	EFF WASTE WATER	WW	X		11/29/25	1:00 PM				X	X	X						
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>Ant Sharpshaw</u>	DATE/TIME: <u>11/20/25</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP <u>2.5</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME: <u>11/20/25 16:00</u>	RECEIVED BY: 2. <u>[Signature]</u>	Comments: <u>METALS - LEAD, ZINC</u> <u>ONLY METALS keep</u>
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page ____ of CLIENT: <input type="checkbox"/> Hand Delivered <input type="checkbox"/> Other Shipment Complete <input type="checkbox"/> YES <input type="checkbox"/> NO

### Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

**LOGIN REPORT/SAMPLE TRANSFER**

<b>Order ID :</b> Q3700	ARDM01	<b>Order Date :</b> 11/20/2025 4:11:00 PM	<b>Project Mgr :</b>
<b>Client Name :</b> Ardmore Chemical		<b>Project Name :</b> PVSC Monthly 2025	<b>Report Type :</b> Level 1
<b>Client Contact :</b> Michael Sharphouse		<b>Receive DateTime :</b> 11/20/2025 4:00:00 PM	<b>EDD Type :</b> NONE
<b>Invoice Name :</b> Ardmore Chemical		<b>Purchase Order :</b>	<b>Hard Copy Date :</b>
<b>Invoice Contact :</b> Michael Sharphouse			<b>Date Signoff :</b>

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3700-01	EFF-WW	Water	11/20/2025	13:00	VOC-PP		624.1	10 Bus. Days	
Q3700-02	Q3700-01MS	Water	11/20/2025	13:00	VOC-PP		624.1	10 Bus. Days	
Q3700-03	Q3700-01MSD	Water	11/20/2025	13:00	VOC-PP		624.1	10 Bus. Days	

Relinquished By : OPDate / Time : 11/21/25 9:55Received By : SausDate / Time : 11/21/25 9:55

Storage Area : VOA Refridgerator Room