

Initial and Continuing Calibration Verification

Client: Ardmore Chemical

SDG No.: Q3700

Project: PVSC Monthly 2025

RunNo.: LB138026

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.099	0.099	100	85-115	11/24/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	11/24/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	11/24/2025