

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3700

Project: PVSC Monthly 2025

Run No.: LB138012

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138012BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	11/21/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3700

Project: PVSC Monthly 2025

Run No.: LB138018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138018BS							
TSS	mg/L	550	591		107	1	90-110	11/21/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3700

Project: PVSC Monthly 2025

Run No.: LB138026

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170708BS							
Cyanide	mg/L	0.1	0.099		99	1	85-115	11/24/2025