
## LB138004

245.1 INSTRUMENT ID: CV1



A=	0.0000e+000
B=	2.3614e-004
C=	3.2107e-004
Rho=	0.9996677
Accept	t=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	80	0.000	80				
0.05	0.050	0.053	0.003	224	0.0 %	224				
0.2	0.200	0.250	0.050	1059	0.0 %	1059				
2.5	2.500	2.511	0.011	10632	0.0 %	10632				
5.0	5.000	4.945	-0.055	20938	0.0 %	20938				
7.5	7.500	7.316	-0.184	30980	0.0 %	30980				
10.0	10.000	10.155	0.155	43004	0.0 %	43004				

## LB138004 INSTRUMENMT ID: CV1

	d: 245.1	Operator: Admin		alysis: 21 N			WI 1D . CV I	
Type	Sample ID 0.0 - 1	Extended ID	Conc	μ Abs.		Type	Stnd CondMethod	Date
;	0.05 - 1	36:00	-	80	PPB	Std	0.0000245.1	21 Nov 2025 11:57:55
	0.05 - 1	502	-	224	PPB	Std	0.0500245.1	21 Nov 2025 12:00:23
	2.5 - 1	52.5	-1	1059	PPB	Std	0.2000245.1	21 Nov 2025 12:02:40
	5.0 - 1	55	-1	10632	PPB	Std	2.5000245.1	21 Nov 2025 12:04:57
	7.5 - 1	13.5	-1	20938	PPB	Std	5.0000245.1	21 Nov 2025 12:07:14
	10.0 - 1	lin	-1	30980	PPB	Std	7.5000245.1	21 Nov 2025 12:09:30
	ICV124 - 1			<b>430</b> 04	PPB	Std	10.0000245.1	21 Nov 2025 12:24:50
	ICB124 - 1	ICV124	3.9416	16690	PPB	SMPL	-245.1	21 Nov 2025 12:29:55
	ICCV22 - 1	ICB124	-0.0124	-54	PPB	SMPL	-245.1	21 Nov 2025 12:32:29
	CCB22 - 1	CCV22 CCB22	5.0302	21300	PPB	SMPL	-245.1	21 Nov 2025 12:34:44
	CRA - 1	CRA	-0.0167	-72	PPB	SMPL	-245.1	21 Nov 2025 12:36:59
	HighStd - 1	1	0.0457	192	PPB	SMPL	-245.1	21 Nov 2025 12:39:14
	ChkStd - 1	HighStd ChkStd	10.3942	44015	PPB	SMPL	-245.1	21 Nov 2025 12:41:29
	PB170687BL - 1	PBW	7.1014	30071	PPB	SMPL	-245.1	21 Nov 2025 12:43:44
	PB170687BS - 1	LCSW	-0.0512	-218	PPB	SMPL	-245.1	21 Nov 2025 12:46:01
	Q3603-01 - 1	SHORT-HILLS-DW	4.4058	18656	PPB	SMPL	-245.1	21 Nov 2025 12:48:18
	Q3603-01DUP - 1	SHORT-HILLS-DWDUP	-0.0231	-99	PPB	SMPL	-245.1	21 Nov 2025 12:50:35
	Q3603-01MS - 1	SHORT-HILLS-DWMS	0.0010	45000	PPB	SMPL	-245.1	21 Nov 2025 12:52:51
	Q3603-01MSD - 1	SHORT-HILLS-DWMSD	3.6322		PPB	SMPL	-245.1	21 Nov 2025 12:55:08
	Q3700-04 - 1	EFF-WW	4.2084	17820	PPB	SMPL	-245.1	21 Nov 2025 12:57:25
- 1	CCV23 - 1	CCV23	-0.0198	-85	PPB	SMPL	-245.1	21 Nov 2025 12:59:40
	CCB23 - 1	CCB23	4.9674	21034	PPB	SMPL	-245.1	21 Nov 2025 13:01:56
	Q3603-01LX5 - 1	55525	-0.0202		PPB	SMPL	-245.1	21 Nov 2025 13:04:11
	Q3603-01A - 1		-0.0013 4.0814		PPB	SMPL	-245.1	21 Nov 2025 13:06:27
- 1	CCV24 - 1	CCV24	5.0760			SMPL	-245.1	21 Nov 2025 13:08:42
	CCB24 - 1	CCB24			- 1	SMPL	-245.1	21 Nov 2025 13:13:57
r		00024	-0.0327	-140	PPB	SMPL	-245.1	21 Nov 2025 13:16:12