



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB124	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	12:32	LB138004

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB22	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	12:36	LB138004
CCB23	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	13:04	LB138004
CCB24	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	13:16	LB138004

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Lead	2.30	+/-6	U	12.0	P	11/21/2025	13:13	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	13:13	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Lead	2.30	+/-6	U	12.0	P	11/21/2025	13:48	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	13:48	LB138017
CCB02	Lead	2.30	+/-6	U	12.0	P	11/21/2025	14:52	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	14:52	LB138017
CCB03	Lead	2.30	+/-6	U	12.0	P	11/21/2025	15:42	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	15:42	LB138017
CCB04	Lead	2.30	+/-6	U	12.0	P	11/21/2025	16:32	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	16:32	LB138017
CCB05	Lead	2.30	+/-6	U	12.0	P	11/21/2025	17:55	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	17:55	LB138017
CCB06	Lead	2.30	+/-6	U	12.0	P	11/21/2025	18:45	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	18:45	LB138017
CCB07	Lead	2.30	+/-6	U	12.0	P	11/21/2025	19:35	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	19:35	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------