

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	8.32			-12	12	11/21/2025	13:22	LB138017
	Zinc	6.33			-40	40	11/21/2025	13:22	LB138017
ICSAB01	Lead	52.8	49.0	108	37	61	11/21/2025	13:26	LB138017
	Zinc	1020	952	107	809	1095	11/21/2025	13:26	LB138017
ICSA	Lead	-6.10			-12	12	11/21/2025	13:30	LB138017
	Zinc	14.5			-40	40	11/21/2025	13:30	LB138017
ICSAB	Lead	16.9	49.0	34	37	61	11/21/2025	13:34	LB138017
	Zinc	1000	952	105	809	1095	11/21/2025	13:34	LB138017