

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3700</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3603-01</u>	Client ID:	<u>SHORT-HILLS-DWDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3603-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3700</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3603-01MS</u>	Client ID:	<u>SHORT-HILLS-DWMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3603-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.63		4.21		15		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3700</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3700-04</u>	Client ID:	<u>EFF-WWDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3700-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Lead	ug/L	20	6.00	U	6.00	U			P
Zinc	ug/L	20	206		204		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Ardmore Chemical</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3700</u>
Contract:	<u>ARDM01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3700-05</u>	Client ID:	<u>EFF-WWMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3700-06	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Lead	ug/L	20	453		440		3		P
Zinc	ug/L	20	287		290		1		P