

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170685BS							
Lead	ug/L	500	474		95	85 - 115	P
Zinc	ug/L	100	103		103	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical

SDG No.: Q3700

Contract: ARDM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170687BS Mercury	ug/L	4.0	4.41		110	85 - 115	CV