

Surrogate Summary

SW-846

SDG No.: Q3700

Client: Ardmore Chemical

Analytical Method: 625.1

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170730BL	PB170730BL	2-Fluorophenol	100	84.6	85		60	140
		Phenol-d6	100	87.9	88		60	140
		Nitrobenzene-d5	100	84.2	84		60	140
		2-Fluorobiphenyl	100	83.9	84		60	140
		2,4,6-Tribromophenol	100	65.0	65		60	140
		Terphenyl-d14	100	79.5	79		60	140
PB170730BS	PB170730BS	2-Fluorophenol	100	88.4	88		60	140
		Phenol-d6	100	88.6	89		60	140
		Nitrobenzene-d5	100	86.0	86		60	140
		2-Fluorobiphenyl	100	90.3	90		60	140
		2,4,6-Tribromophenol	100	81.2	81		60	140
		Terphenyl-d14	100	83.5	84		60	140
Q3700-04	EFF-WW	2-Fluorophenol	100	52.0	52	*	60	140
		Phenol-d6	100	35.9	36	*	60	140
		Nitrobenzene-d5	100	106	106		60	140
		2-Fluorobiphenyl	100	76.1	76		60	140
		2,4,6-Tribromophenol	100	73.8	74		60	140
		Terphenyl-d14	100	65.5	65		60	140
Q3700-05MS	EFF-WWMS	2-Fluorophenol	100	56.5	56	*	60	140
		Phenol-d6	100	40.9	41	*	60	140
		Nitrobenzene-d5	100	84.0	84		60	140
		2-Fluorobiphenyl	100	79.2	79		60	140
		2,4,6-Tribromophenol	100	70.2	70		60	140
		Terphenyl-d14	100	68.0	68		60	140
Q3700-06MSD	EFF-WWMSD	2-Fluorophenol	100	57.5	58	*	60	140
		Phenol-d6	100	40.1	40	*	60	140
		Nitrobenzene-d5	100	87.3	87		60	140
		2-Fluorobiphenyl	100	81.7	82		60	140
		2,4,6-Tribromophenol	100	71.5	72		60	140
		Terphenyl-d14	100	70.8	71		60	140