

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3700

Client ID : SSTDCCC040

Date Analyzed: 11/25/2025

Lab File ID: BF144348.D

Time Analyzed: 13:54

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	48785	6.863	180679	8.15	87905	9.90
UPPER LIMIT	97570	7.363	361358	8.645	175810	10.398
LOWER LIMIT	24392.5	6.363	90339.5	7.645	43952.5	9.398
EPA SAMPLE NO.						
01 PB170730BL	60817	6.86	237169	8.14	123302	9.90
02 PB170730BS	61843	6.86	232918	8.15	113683	9.90
03 EFF-WW	52716	6.86	150908	8.15	98746	9.90
04 EFF-WWMS	53861	6.86	190607	8.15	99198	9.90
05 EFF-WWMSD	51254	6.86	175810	8.15	91917	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.