

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112525\
 Data File : VN088353.D
 Acq On : 25 Nov 2025 11:03
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 26 00:16:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111325W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 14 00:49:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.00
2 M	Dichlorodifluoromethane	2.109	2.104	0.2	91	0.00
3 M	Chloromethane	2.174	2.232	-2.7	99	0.00
4 M	Vinyl Chloride	2.299	2.254	2.0	95	0.00
5 M	Bromomethane	1.226	1.093	10.8	92	0.00
6 M	Chloroethane	1.421	1.384	2.6	95	0.00
7 M	Trichlorofluoromethane	3.223	3.191	1.0	95	0.00
8 T	Diethyl Ether	1.205	1.135	5.8	97	0.00
9	1,1,2-Trichlorotrifluoroeth	1.691	1.640	3.0	93	0.00
10 M	1,1-Dichloroethene	1.650	1.634	1.0	95	0.00
11	Methyl Iodide	1.756	1.491	15.1	87	0.00
12	Methyl Acetate	2.661	2.802	-5.3	105	0.00
13 M	Acrolein	0.392	0.620	-58.2#	179#	-0.01
14 M	Acrylonitrile	1.209	1.180	2.4	98	0.00
15 M	Acetone	0.354	0.261	26.3#	73	0.00
16 M	Carbon Disulfide	5.564	5.370	3.5	92	0.00
17	Allyl chloride	3.206	2.821	12.0	87	0.00
18 M	Methylene Chloride	2.050	2.016	1.7	98	0.02
19 M	trans-1,2-Dichloroethene	1.921	1.870	2.7	94	0.00
20 T	Diisopropyl ether	6.850	6.750	1.5	99	0.00
21 M	1,1-Dichloroethane	3.859	3.714	3.8	95	0.00
22 M	cis-1,2-Dichloroethene	2.233	2.147	3.9	97	0.00
23 M	tert-Butyl Alcohol	0.445	0.405	9.0	95	0.00
24 M	Methyl tert-Butyl Ether	6.651	6.208	6.7	95	0.00
25 M	Chloroform	3.824	3.782	1.1	100	0.00
26	Cyclohexane	3.260	3.084	5.4	89	0.00
27 s	1,2-Dichloroethane-d4	2.740	2.594	5.3	97	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
29	1,1-Dichloropropene	0.501	0.502	-0.2	94	0.00
30 M	2-Butanone	0.324	0.298	8.0	90	0.00
31	2,2-Dichloropropane	0.617	0.599	2.9	91	0.00
32 M	1,1,1-Trichloroethane	0.617	0.634	-2.8	97	0.00
33 M	Carbon Tetrachloride	0.527	0.512	2.8	91	0.00
34 M	Benzene	1.496	1.546	-3.3	98	0.00
35	Methacrylonitrile	0.338	0.330	2.4	97	0.00
36 M	1,2-Dichloroethane	0.582	0.583	-0.2	96	0.00
37 M	Trichloroethene	0.339	0.362	-6.8	101	0.00
38	Methylcyclohexane	0.574	0.542	5.6	87	0.00
39 M	1,2-Dichloropropane	0.386	0.401	-3.9	101	0.00
40	Dibromomethane	0.277	0.282	-1.8	97	0.00
41 M	Bromodichloromethane	0.558	0.561	-0.5	99	0.00
42 M	Vinyl Acetate	1.081	1.050	2.9	95	0.00
43	Ethyl Acetate	0.651	0.686	-5.4	101	0.00
44	Isopropyl Acetate	1.004	1.010	-0.6	98	0.00
45 T	1,4-Dioxane	0.006	0.006	0.0	87	0.00
46	Methyl methacrylate	0.470	0.453	3.6	96	0.00
47	n-amyl Acetate	0.509	0.413	18.9	86	0.01
48 M	t-1,3-Dichloropropene	0.597	0.568	4.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112525\
 Data File : VN088353.D
 Acq On : 25 Nov 2025 11:03
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 26 00:16:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111325W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 14 00:49:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.621	0.637	-2.6	99	0.00
50 M	1,1,2-Trichloroethane	0.343	0.359	-4.7	102	0.00
51	Ethyl methacrylate	0.564	0.539	4.4	98	0.00
52	1,3-Dichloropropane	0.625	0.619	1.0	96	0.00
53 M	Dibromochloromethane	0.394	0.402	-2.0	101	0.00
54 M	1,2-Dibromoethane	0.356	0.359	-0.8	100	0.00
55 M	2-Chloroethyl vinyl ether	0.325	0.315	3.1	98	0.00
56 M	Bromoform	0.249	0.243	2.4	99	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
58 M	4-Methyl-2-Pentanone	0.679	0.716	-5.4	102	0.00
59 M	2-Hexanone	0.451	0.437	3.1	98	0.00
60 S	4-Bromofluorobenzene	0.507	0.480	5.3	95	0.00
61 M	Tetrachloroethene	0.309	0.382	-23.6	114	0.00
62 M	Toluene	1.722	1.798	-4.4	99	0.00
63 S	Toluene-d8	1.420	1.450	-2.1	98	0.00
64 M	Chlorobenzene	1.066	1.064	0.2	98	0.00
65	1,1,1,2-Tetrachloroethane	0.365	0.362	0.8	96	0.00
66 M	Ethyl Benzene	1.897	1.868	1.5	95	0.00
67 M	m/p-Xylenes	0.710	0.713	-0.4	97	0.00
68 M	o-Xylene	0.669	0.664	0.7	98	0.00
69 M	Styrene	1.137	1.095	3.7	96	0.00
70	Isopropylbenzene	1.774	1.713	3.4	93	0.00
71 M	1,1,2,2-Tetrachloroethane	0.575	0.540	6.1	92	0.00
72	1,2,3-Trichloropropane	0.557	0.516	7.4	99	0.00
73	Bromobenzene	0.400	0.401	-0.3	98	0.00
74	n-propylbenzene	2.195	2.157	1.7	94	0.00
75	2-Chlorotoluene	1.335	1.251	6.3	92	0.00
76	1,3,5-Trimethylbenzene	1.489	1.427	4.2	93	0.00
77	t-1,4-Dichloro-2-butene	0.191	0.171	10.5	93	0.00
78	4-Chlorotoluene	1.353	1.310	3.2	94	0.00
79	tert-butylbenzene	1.275	1.178	7.6	91	0.00
80	1,2,4-Trimethylbenzene	1.481	1.399	5.5	93	0.00
81	sec-Butylbenzene	1.859	1.741	6.3	89	0.00
82	p-Isopropyltoluene	1.536	1.423	7.4	88	0.00
83 M	1,3-Dichlorobenzene	0.778	0.775	0.4	95	0.00
84 M	1,4-Dichlorobenzene	0.788	0.764	3.0	95	0.00
85	n-Butylbenzene	1.464	1.333	8.9	85	0.00
86 T	Hexachloroethane	0.294	0.251	14.6	84	0.00
87 M	1,2-Dichlorobenzene	0.734	0.709	3.4	93	0.00
88	1,2-Dibromo-3-Chloropropane	0.135	0.130	3.7	99	0.00
89	1,2,4-Trichlorobenzene	0.412	0.383	7.0	93	0.00
90	Hexachlorobutadiene	0.160	0.148	7.5	88	0.00
91 M	Naphthalene	1.491	1.346	9.7	91	0.00
92	1,2,3-Trichlorobenzene	0.412	0.383	7.0	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0