

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARDM01
 Lab Code: ACE SDG NO.: Q3700
 Lab File ID: VN088353.D Date Analyzed: 11/25/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	57986	7.80	299947	9.08	269125	11.85
UPPER LIMIT	115972	8.3	599894	9.582	538250	12.347
LOWER LIMIT	28993	7.3	149974	8.582	134563	11.347
EPA SAMPLE NO.						
EFF-WW	50615	7.79	245873	9.08	214867	11.85
EFF-WWMS	50762	7.81	258609	9.08	233753	11.84
VN1125WBL01	55920	7.80	279977	9.08	244619	11.85
VN1125WBS01	57156	7.79	284593	9.08	251170	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.