

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARDM01
 Lab Code: ACE SDG NO.: Q3700
 Lab File ID: VN088362.D Date Analyzed: 11/26/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:06
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	55632	7.80	271784	9.08	250226	11.84
UPPER LIMIT	111264	8.3	543568	9.582	500452	12.341
LOWER LIMIT	27816	7.3	135892	8.582	125113	11.341
EPA SAMPLE NO.						
EFF-WWMSD	54953	7.80	283065	9.08	251234	11.84
VN1126WBL01	48647	7.80	246777	9.08	220989	11.85
VN1126WBS01	57477	7.79	302288	9.08	272302	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.