Water Ammonia Preparation Sheet



SOP ID: MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A Start Digest Date: 11/25/2025 Time : 09:10 Temp : 150 °C

Pippete ID: WC

Balance ID: N/A

Hood ID: HOOD#2 Digestion tube ID: M5595 Block Thermometer ID: WC CYANIDE

Block ID: WC-DIST-BLOCK-1 Filter paper ID: N/A Prep Technician Signature:

Weigh By: N/A pH Meter ID: N/A Supervisor Signature: 12

Standared Name	MLS USED	STD REF. # FROM LOG	
LCSW	1.0ML	WP115589	
MS/MSD SPIKE SOL.	1.0ML	WP115588	
PBW	50.0ML	W3112	
RL CHECK	0.1ML	WP115588	
N/A	N/A	N/A	

Chemical Used	ML/SAMPLE USED	Lot Number WP113886		
BORATE BUFFER	2.5ML			
NAOH 6N	0.5-2.0ML	WP113887		
H2SO4 0.04N	5.0ML	WP115336		
pH strip-Ammonia	N/A	W3133		
KI-starch paper	N/A	W3155		
V/A	N/A N/A			
N/A	N/A	N/A		
V/A	N/A	N/A		
N/A	N/A	N/A		
N/A	N/A	N/A		

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT WP114104.Due to bad matrix and client history 1ML was taken as an initial volume for Q3701-01 and Q3701-05.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
175/2025 1075	RHLWCT	RMCWG
	Preparation Group	Analysis Group



Water Ammonia Preparation Sheet

PB170732

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	рH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep
PB170732BL	PBW732	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB170732BS	LCS732	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3701-01DUP	EFFLUENTDUP	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3701-01MS	EFFLUENTMS	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3701-01MSD	EFFLUENTMSD	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3701-01 	EFFLUENT	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
(3701-05	INFLUENT	1	50	<2	N/A	Negative		AFTER ADDING 6N NAOH PH IS 9.5	N/A