

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

RunNo.: LB138011

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.483	0.50	97	90-110	11/21/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.517	0.5	103	90-110	11/21/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.508	0.5	102	90-110	11/21/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

RunNo.: LB138013

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.494	0.50	99	90-110	11/21/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.526	0.50	105	90-110	11/21/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.517	0.50	103	90-110	11/21/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

RunNo.: LB138039

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	11/25/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.93	1	93	90-110	11/25/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.98	1	98	90-110	11/25/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.93	1	93	90-110	11/25/2025