

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

Run No.: LB138011

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138011BS							
Orthophosphate as P	mg/L	0.5	0.47		95	1	90-110	11/21/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

Run No.: LB138012

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138012BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	11/21/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

Run No.: LB138018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138018BS							
TSS	mg/L	550	591		107	1	90-110	11/21/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

Run No.: LB138035

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138035BS							
Oil and Grease	mg/L	20.0	19.1		96	1	78-114	11/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

Run No.: LB138013

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170686BS							
Phosphorus, Total	mg/L	0.50	0.47		95	1	90-110	11/21/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3701

Project: Pre Treatment Plant 2025

Run No.: LB138039

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170732BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	11/25/2025