

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3702</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/21/2025 09:08</u>
Lab File ID:	<u>VX048641.D</u>	Init. Calib. Date(s):	<u>11/07/2025 11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35 16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.405	0.421		3.95	20
Chloromethane	0.537	0.452	0.1	-15.83	20
Vinyl Chloride	0.608	0.583		-4.11	20
Ethyl Acetate	0.713	0.609		-14.59	20
Bromomethane	0.384	0.352		-8.33	20
Chloroethane	0.398	0.339		-14.82	20
Trichlorofluoromethane	0.963	0.904		-6.13	20
1,1,2-Trichlorotrifluoroethane	0.557	0.592		6.28	20
Tert butyl alcohol	0.132	0.138		4.55	20
1,1-Dichloroethene	0.582	0.538		-7.56	20
Acrolein	0.019	0.017		-10.53	20
Acrylonitrile	0.385	0.400		3.9	20
Acetone	0.330	0.276		-16.36	20
Carbon Disulfide	1.638	1.351		-17.52	20
Methyl tert-butyl Ether	2.055	2.103		2.34	20
Methyl Acetate	0.906	0.885		-2.32	20
Methylene Chloride	0.678	0.647		-4.57	20
trans-1,2-Dichloroethene	0.617	0.585		-5.19	20
Vinyl Acetate	1.822	1.853		1.7	20
1,1-Dichloroethane	1.142	1.105	0.1	-3.24	20
Cyclohexane	0.936	0.861		-8.01	20
2-Butanone	0.488	0.441		-9.63	20
Carbon Tetrachloride	0.577	0.562		-2.6	20
2,2-Dichloropropane	0.973	0.970		-0.31	20
cis-1,2-Dichloroethene	0.754	0.673		-10.74	20
Bromochloromethane	0.561	0.528		-5.88	20
Chloroform	1.179	1.147		-2.71	20
1,1,1-Trichloroethane	1.029	1.026		-0.29	20
Methylcyclohexane	0.569	0.561		-1.41	20
1,1-Dichloropropene	0.507	0.463		-8.68	20
Benzene	1.571	1.414		-9.99	20
1,2-Dichloroethane	0.538	0.541		0.56	20
Trichloroethene	0.375	0.365		-2.67	20
1,2-Dichloropropane	0.361	0.363		0.55	20
Dibromomethane	0.263	0.278		5.7	20
Bromodichloromethane	0.526	0.582		10.65	20
4-Methyl-2-Pentanone	0.571	0.632		10.68	20
Toluene	0.827	0.898		8.59	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3702</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/21/2025 09:08</u>
Lab File ID:	<u>VX048641.D</u>	Init. Calib. Date(s):	<u>11/07/2025 11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35 16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18 (mm)</u>

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
t-1,3-Dichloropropene	0.541	0.590		9.06	20
cis-1,3-Dichloropropene	0.561	0.614		9.45	20
1,1,2-Trichloroethane	0.330	0.373		13.03	20
1,3-Dichloropropane	0.575	0.622		8.17	20
2-Chloroethyl Vinyl ether	0.296	0.335		13.18	20
2-Hexanone	0.425	0.462		8.71	20
Dibromochloromethane	0.406	0.463		14.04	20
1,2-Dibromoethane	0.358	0.393		9.78	20
Tetrachloroethene	0.363	0.343		-5.51	20
Chlorobenzene	1.187	1.132	0.3	-4.63	20
1,1,1,2-Tetrachloroethane	0.398	0.414		4.02	20
Hexachloroethane	0.606	0.596		-1.65	20
Ethyl Benzene	1.993	1.929		-3.21	20
m/p-Xylenes	0.751	0.741		-1.33	20
o-Xylene	0.734	0.724		-1.36	20
Styrene	1.232	1.242		0.81	20
Bromoform	0.367	0.366	0.1	-0.27	20
Isopropylbenzene	3.629	3.586		-1.18	20
1,1,2,2-Tetrachloroethane	1.280	1.199	0.3	-6.33	20
1,2,3-Trichloropropane	1.128	1.151		2.04	20
Bromobenzene	0.937	0.886		-5.44	20
n-propylbenzene	4.268	4.207		-1.43	20
2-Chlorotoluene	2.579	2.493		-3.34	20
1,3,5-Trimethylbenzene	2.959	2.914		-1.52	20
4-Chlorotoluene	3.025	2.937		-2.91	20
tert-Butylbenzene	3.060	3.007		-1.73	20
1,2,4-Trimethylbenzene	2.985	2.936		-1.64	20
sec-Butylbenzene	3.782	3.778		-0.11	20
p-Isopropyltoluene	3.180	3.186		0.19	20
1,3-Dichlorobenzene	1.756	1.641		-6.55	20
1,4-Dichlorobenzene	1.814	1.676		-7.61	20
n-Butylbenzene	2.983	2.977		-0.2	20
1,2-Dichlorobenzene	1.686	1.598		-5.22	20
1,2-Dibromo-3-Chloropropane	0.297	0.283		-4.71	20
1,2,4-Trichlorobenzene	1.120	1.041		-7.05	20
Hexachlorobutadiene	0.468	0.475		1.5	20
Naphthalene	3.603	3.793		5.27	20
1,2,3-Trichlorobenzene	1.042	1.015		-2.59	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3702</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/21/2025</u> <u>09:08</u>
Lab File ID:	<u>VX048641.D</u>	Init. Calib. Date(s):	<u>11/07/2025</u> <u>11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35</u> <u>16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.697	0.708		1.58	20
Dibromofluoromethane	0.378	0.372		-1.59	20
Toluene-d8	1.180	1.238		4.91	20
4-Bromofluorobenzene	0.440	0.470		6.82	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.