

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3702	OrderDate:	11/21/2025 11:05:00 AM
Client:	Chemtech Consulting Group	Project:	Weekly Storage Blanks
Contact:	QA Officer	Location:	D41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3702-01	STORAGE-BLANK-SAMPLE-MANAGEMENT	Water			11/14/25			11/21/25
			VOC- Chemtech Full	8260-Low			11/21/25	
Q3702-02	STORAGE-BLANK-WATER-REF-5	Water			11/14/25			11/21/25
			VOC- Chemtech Full	8260-Low			11/21/25	
Q3702-03	STORAGE-BLANK-WATER-REF-4	Water			11/14/25			11/21/25
			VOC- Chemtech Full	8260-Low			11/21/25	
Q3702-04	STORAGE-BLANK-SAMPLE-MANAGEMENT	Water			11/14/25			11/21/25
			VOC- Chemtech Full	8260-Low			11/21/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3702

Client: Chemtech Consulting Group

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:



QC SUMMARY

Surrogate Summary

SDG No.: **Q3702**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3702-01	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	62.2	124		74	125
		Dibromofluoromethane	50	44.9	90		75	124
		Toluene-d8	50	49.0	98		86	113
		4-Bromofluorobenzene	50	58.3	117		77	121
Q3702-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	50.3	101		74	125
		Dibromofluoromethane	50	42.7	85		75	124
		Toluene-d8	50	49.0	98		86	113
		4-Bromofluorobenzene	50	55.2	110		77	121
Q3702-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	54.7	109		74	125
		Dibromofluoromethane	50	43.3	87		75	124
		Toluene-d8	50	47.0	94		86	113
		4-Bromofluorobenzene	50	54.9	110		77	121
Q3702-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	51.7	103		74	125
		Dibromofluoromethane	50	42.5	85		75	124
		Toluene-d8	50	47.8	96		86	113
		4-Bromofluorobenzene	50	54.8	110		77	121
VX1121WBL01	VX1121WBL01	1,2-Dichloroethane-d4	50	54.3	109		74	125
		Dibromofluoromethane	50	46.2	92		75	124
		Toluene-d8	50	52.2	104		86	113
		4-Bromofluorobenzene	50	59.0	118		77	121
VX1121WBS01	VX1121WBS01	1,2-Dichloroethane-d4	50	51.8	104		74	125
		Dibromofluoromethane	50	48.5	97		75	124
		Toluene-d8	50	53.5	107		86	113
		4-Bromofluorobenzene	50	52.5	105		77	121
VX1121WBSD0	VX1121WBSD01	1,2-Dichloroethane-d4	50	45.6	91		74	125
		Dibromofluoromethane	50	46.0	92		75	124
		Toluene-d8	50	49.9	100		86	113
		4-Bromofluorobenzene	50	50.8	102		77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3702</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Chemtech Consulting Group</u>	Datafile :	<u>VX048644.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBS01	Dichlorodifluoromethane	20	20.5	ug/L	103			69	116	
	Chloromethane	20	17.0	ug/L	85			65	116	
	Vinyl chloride	20	18.0	ug/L	90			65	117	
	Ethyl Acetate	20	18.6	ug/L	93			75	115	
	Bromomethane	20	19.8	ug/L	99			58	125	
	Chloroethane	20	17.9	ug/L	90			56	128	
	Trichlorofluoromethane	20	18.8	ug/L	94			73	115	
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/L	102			80	112	
	Tert butyl alcohol	100	98.6	ug/L	99			48	142	
	1,1-Dichloroethene	20	18.1	ug/L	91			74	110	
	Acrolein	100	82.1	ug/L	82			18	155	
	Acrylonitrile	100	98.8	ug/L	99			73	113	
	Acetone	100	88.0	ug/L	88			60	125	
	Carbon disulfide	20	15.5	ug/L	78			64	112	
	Methyl tert-butyl Ether	20	19.7	ug/L	99			78	114	
	Methyl Acetate	20	19.9	ug/L	100			57	150	
	Methylene Chloride	20	18.4	ug/L	92			72	114	
	trans-1,2-Dichloroethene	20	18.3	ug/L	92			75	108	
	Vinyl Acetate	100	96.1	ug/L	96			76	115	
	1,1-Dichloroethane	20	19.4	ug/L	97			78	112	
	Cyclohexane	20	19.3	ug/L	97			75	110	
	2-Butanone	100	97.6	ug/L	98			65	122	
	Carbon Tetrachloride	20	18.7	ug/L	94			77	113	
	2,2-Dichloropropane	20	19.5	ug/L	98			77	109	
	cis-1,2-Dichloroethene	20	18.8	ug/L	94			77	110	
	Bromochloromethane	20	16.9	ug/L	85			70	124	
	Chloroform	20	20.3	ug/L	102			79	113	
	1,1,1-Trichloroethane	20	19.9	ug/L	100			80	108	
	Methylcyclohexane	20	19.1	ug/L	96			72	115	
	1,1-Dichloropropene	20	18.2	ug/L	91			79	110	
	Benzene	20	18.5	ug/L	93			82	109	
	1,2-Dichloroethane	20	19.5	ug/L	98			80	115	
	Trichloroethene	20	19.9	ug/L	100			77	113	
	1,2-Dichloropropane	20	20.6	ug/L	103			83	111	
	Dibromomethane	20	20.6	ug/L	103			82	110	
	Bromodichloromethane	20	21.8	ug/L	109			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	21.7	ug/L	109			82	110	
	t-1,3-Dichloropropene	20	20.8	ug/L	104			79	110	
	cis-1,3-Dichloropropene	20	21.6	ug/L	108			82	110	
	1,1,2-Trichloroethane	20	22.7	ug/L	114		*	83	112	
	1,3-Dichloropropane	20	21.9	ug/L	110			83	111	
	2-Chloroethyl vinyl ether	100	91.4	ug/L	91			75	124	
	2-Hexanone	100	110	ug/L	110			73	117	
	Dibromochloromethane	20	22.6	ug/L	113		*	82	110	
	1,2-Dibromoethane	20	21.7	ug/L	109			81	110	
	Tetrachloroethene	20	18.7	ug/L	94			67	123	
	Chlorobenzene	20	19.0	ug/L	95			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3702</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Chemtech Consulting Group</u>	Datafile :	<u>VX048644.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBS01	1,1,1,2-Tetrachloroethane	20	20.6	ug/L	103			84	111	
	Hexachloroethane	20	19.3	ug/L	97			80	113	
	Ethyl Benzene	20	19.2	ug/L	96			83	109	
	m/p-Xylenes	40	39.0	ug/L	98			82	110	
	o-Xylene	20	19.2	ug/L	96			83	109	
	Styrene	20	19.7	ug/L	99			80	111	
	Bromoform	20	19.1	ug/L	96			79	109	
	Isopropylbenzene	20	20.0	ug/L	100			83	112	
	1,1,2,2-Tetrachloroethane	20	19.4	ug/L	97			76	118	
	1,2,3-Trichloropropane	20	18.8	ug/L	94			75	112	
	Bromobenzene	20	19.8	ug/L	99			77	116	
	N-propylbenzene	20	19.7	ug/L	99			83	112	
	2-Chlorotoluene	20	19.5	ug/L	98			83	110	
	1,3,5-Trimethylbenzene	20	19.7	ug/L	99			85	112	
	4-Chlorotoluene	20	19.6	ug/L	98			82	110	
	tert-Butylbenzene	20	19.7	ug/L	99			83	112	
	1,2,4-Trimethylbenzene	20	20.0	ug/L	100			85	111	
	Sec-butylbenzene	20	19.9	ug/L	100			81	114	
	p-Isopropyltoluene	20	19.9	ug/L	100			78	116	
	1,3-Dichlorobenzene	20	19.0	ug/L	95			82	108	
	1,4-Dichlorobenzene	20	18.7	ug/L	94			82	107	
	n-Butylbenzene	20	19.4	ug/L	97			75	115	
	1,2-Dichlorobenzene	20	18.7	ug/L	94			82	109	
	1,2-Dibromo-3-Chloropropane	20	18.6	ug/L	93			68	112	
	1,2,4-Trichlorobenzene	20	18.4	ug/L	92			75	113	
	Hexachlorobutadiene	20	18.3	ug/L	92			81	121	
	Naphthalene	20	18.7	ug/L	94			78	119	
	1,2,3-Trichlorobenzene	20	18.8	ug/L	94			76	114	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3702</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Chemtech Consulting Group</u>	Datafile :	<u>VX048645.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBSD01	Dichlorodifluoromethane	20	19.4	ug/L	97	6		69	116	19
	Chloromethane	20	16.1	ug/L	81	5		65	116	21
	Vinyl chloride	20	16.8	ug/L	84	7		65	117	19
	Ethyl Acetate	20	15.7	ug/L	79	16		75	115	27
	Bromomethane	20	18.8	ug/L	94	5		58	125	30
	Chloroethane	20	16.8	ug/L	84	7		56	128	20
	Trichlorofluoromethane	20	17.4	ug/L	87	8		73	115	16
	1,1,2-Trichlorotrifluoroethane	20	19.1	ug/L	96	6		80	112	20
	Tert butyl alcohol	100	91.7	ug/L	92	7		48	142	30
	1,1-Dichloroethene	20	17.3	ug/L	86	6		74	110	20
	Acrolein	100	83.9	ug/L	84	2		18	155	30
	Acrylonitrile	100	89.0	ug/L	89	11		73	113	29
	Acetone	100	85.4	ug/L	85	3		60	125	27
	Carbon disulfide	20	14.0	ug/L	70	11		64	112	17
	Methyl tert-butyl Ether	20	18.2	ug/L	91	8		78	114	20
	Methyl Acetate	20	18.0	ug/L	90	11		57	150	20
	Methylene Chloride	20	17.1	ug/L	86	7		72	114	20
	trans-1,2-Dichloroethene	20	16.9	ug/L	85	8		75	108	16
	Vinyl Acetate	100	95.4	ug/L	95	1		76	115	26
	1,1-Dichloroethane	20	17.5	ug/L	88	10		78	112	20
	Cyclohexane	20	16.6	ug/L	83	16		75	110	20
	2-Butanone	100	87.1	ug/L	87	12		65	122	26
	Carbon Tetrachloride	20	16.8	ug/L	84	11		77	113	15
	2,2-Dichloropropane	20	18.3	ug/L	92	6		77	109	20
	cis-1,2-Dichloroethene	20	18.2	ug/L	91	3		77	110	20
	Bromochloromethane	20	14.3	ug/L	72	17		70	124	20
	Chloroform	20	18.0	ug/L	90	13		79	113	20
	1,1,1-Trichloroethane	20	17.9	ug/L	90	11		80	108	20
	Methylcyclohexane	20	18.4	ug/L	92	4		72	115	20
	1,1-Dichloropropene	20	15.9	ug/L	79	14		79	110	16
	Benzene	20	16.9	ug/L	85	9		82	109	15
	1,2-Dichloroethane	20	17.1	ug/L	86	13		80	115	20
	Trichloroethene	20	18.7	ug/L	94	6		77	113	15
	1,2-Dichloropropane	20	20.1	ug/L	101	2		83	111	21
	Dibromomethane	20	20.9	ug/L	104	1		82	110	20
	Bromodichloromethane	20	20.1	ug/L	101	8		83	110	21
	4-Methyl-2-Pentanone	100	99.6	ug/L	100	10		74	118	25
	Toluene	20	20.6	ug/L	103	6		82	110	16
	t-1,3-Dichloropropene	20	19.7	ug/L	99	5		79	110	20
	cis-1,3-Dichloropropene	20	20.1	ug/L	101	7		82	110	16
	1,1,2-Trichloroethane	20	22.2	ug/L	111	3		83	112	20
	1,3-Dichloropropane	20	20.5	ug/L	103	7		83	111	20
	2-Chloroethyl vinyl ether	100	85.3	ug/L	85	7		75	124	20
	2-Hexanone	100	97.9	ug/L	98	12		73	117	25
	Dibromochloromethane	20	22.2	ug/L	111	2	*	82	110	20
	1,2-Dibromoethane	20	21.2	ug/L	106	3		81	110	20
	Tetrachloroethene	20	19.4	ug/L	97	3		67	123	15
	Chlorobenzene	20	18.9	ug/L	95	0		82	109	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3702</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Chemtech Consulting Group</u>	Datafile :	<u>VX048645.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1121WBSD01	1,1,1,2-Tetrachloroethane	20	21.1	ug/L	106	3		84	111	20
	Hexachloroethane	20	19.4	ug/L	97	0		80	113	20
	Ethyl Benzene	20	18.6	ug/L	93	3		83	109	16
	m/p-Xylenes	40	38.9	ug/L	97	1		82	110	20
	o-Xylene	20	19.0	ug/L	95	1		83	109	20
	Styrene	20	19.5	ug/L	98	1		80	111	17
	Bromoform	20	20.9	ug/L	104	8		79	109	20
	Isopropylbenzene	20	19.2	ug/L	96	4		83	112	22
	1,1,2,2-Tetrachloroethane	20	19.0	ug/L	95	2		76	118	20
	1,2,3-Trichloropropane	20	18.1	ug/L	91	3		75	112	20
	Bromobenzene	20	19.2	ug/L	96	3		77	116	20
	N-propylbenzene	20	19.0	ug/L	95	4		83	112	20
	2-Chlorotoluene	20	18.6	ug/L	93	5		83	110	20
	1,3,5-Trimethylbenzene	20	19.3	ug/L	97	2		85	112	20
	4-Chlorotoluene	20	18.9	ug/L	95	3		82	110	20
	tert-Butylbenzene	20	19.7	ug/L	99	0		83	112	20
	1,2,4-Trimethylbenzene	20	19.3	ug/L	97	3		85	111	20
	Sec-butylbenzene	20	19.4	ug/L	97	3		81	114	20
	p-Isopropyltoluene	20	19.3	ug/L	97	3		78	116	20
	1,3-Dichlorobenzene	20	19.2	ug/L	96	1		82	108	20
	1,4-Dichlorobenzene	20	19.0	ug/L	95	1		82	107	20
	n-Butylbenzene	20	18.7	ug/L	94	3		75	115	20
	1,2-Dichlorobenzene	20	19.2	ug/L	96	2		82	109	20
	1,2-Dibromo-3-Chloropropane	20	19.7	ug/L	99	6		68	112	26
	1,2,4-Trichlorobenzene	20	19.3	ug/L	97	5		75	113	21
	Hexachlorobutadiene	20	18.7	ug/L	94	2		81	121	20
	Naphthalene	20	19.5	ug/L	98	4		78	119	20
	1,2,3-Trichlorobenzene	20	20.4	ug/L	102	8		76	114	22

VOLATILE METHOD BLANK SUMMARY

Client ID

VX1121WBL01

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q3702

Lab File ID: VX048643.D

Lab Sample ID: VX1121WBL01

Date Analyzed: 11/21/2025

Time Analyzed: 09:58

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1121WBS01	VX1121WBS01	VX048644.D	11/21/2025
VX1121WBSD01	VX1121WBSD01	VX048645.D	11/21/2025
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3702-01	VX048651.D	11/21/2025
STORAGE-BLANK-WATER-REF-5	Q3702-02	VX048652.D	11/21/2025
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3702-04	VX048654.D	11/21/2025
STORAGE-BLANK-WATER-REF-4	Q3702-03	VX048655.D	11/21/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q3702

Lab File ID: VX048515.D

BFB Injection Date: 11/07/2025

Instrument ID: MSVOA_X

BFB Injection Time: 13:06

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.2
75	30.0 - 60.0% of mass 95	56
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	77.2
175	5.0 - 9.0% of mass 174	5.7 (7.3) 1
176	95.0 - 101.0% of mass 174	75.5 (97.9) 1
177	5.0 - 9.0% of mass 176	4.7 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX048516.D	11/07/2025	13:35
VSTDIC005	VSTDIC005	VX048517.D	11/07/2025	13:56
VSTDIC020	VSTDIC020	VX048518.D	11/07/2025	14:16
VSTDIC100	VSTDIC100	VX048520.D	11/07/2025	14:58
VSTDIC150	VSTDIC150	VX048521.D	11/07/2025	15:18
VSTDIC050	VSTDIC050	VX048524.D	11/07/2025	16:29



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q3702

Lab File ID: VX048640.D

BFB Injection Date: 11/21/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:00

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.8
75	30.0 - 60.0% of mass 95	49.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.4 (0.6) 1
174	50.0 - 100.0% of mass 95	77.4
175	5.0 - 9.0% of mass 174	6.2 (8) 1
176	95.0 - 101.0% of mass 174	74.6 (96.4) 1
177	5.0 - 9.0% of mass 176	4.7 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048641.D	11/21/2025	09:08
VX1121WBL01	VX1121WBL01	VX048643.D	11/21/2025	09:58
VX1121WBS01	VX1121WBS01	VX048644.D	11/21/2025	10:26
VX1121WBSD01	VX1121WBSD01	VX048645.D	11/21/2025	10:55
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3702-01	VX048651.D	11/21/2025	13:04
STORAGE-BLANK-WATER-REF-5	Q3702-02	VX048652.D	11/21/2025	13:25
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3702-04	VX048654.D	11/21/2025	14:07
STORAGE-BLANK-WATER-REF-4	Q3702-03	VX048655.D	11/21/2025	14:44

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3702
 Lab File ID: VX048641.D Date Analyzed: 11/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:08
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	165802	5.53	262765	6.74	235899	10.04
UPPER LIMIT	331604	6.031	525530	7.239	471798	10.537
LOWER LIMIT	82901	5.031	131383	6.239	117950	9.537
EPA SAMPLE NO.						
STORAGE-BLANK-SAMPLE-MANAGEMENT	160862	5.53	334385	6.75	357734	10.04
STORAGE-BLANK-WATER-REF-5	170410	5.54	333480	6.75	361276	10.04
STORAGE-BLANK-WATER-REF-4	160298	5.53	322084	6.74	341432	10.04
STORAGE-BLANK-SAMPLE-MANAGEMENT	157000	5.54	308197	6.75	328677	10.04
VX1121WBL01	203389	5.54	383341	6.75	433179	10.04
VX1121WBS01	173428	5.53	281921	6.74	253067	10.04
VX1121WBSD01	169398	5.53	275788	6.74	240043	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3702
 Lab File ID: VX048641.D Date Analyzed: 11/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:08
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	125375	12				
UPPER LIMIT	250750	12.5				
LOWER LIMIT	62687.5	11.5				
EPA SAMPLE NO.						
STORAGE-BLANK-SAMPLE-MANAGEMENT	192057	12.01				
STORAGE-BLANK-WATER-REF-5	189089	12.01				
STORAGE-BLANK-WATER-REF-4	184333	12.01				
STORAGE-BLANK-SAMPLE-MANAGEMENT	175625	12.01				
VX1121WBL01	227916	12.01				
VX1121WBS01	129952	12.01				
VX1121WBSD01	124653	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-SOIL-1
 Lab Sample ID: Q3702-01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:04	VX112125
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 13:04	VX112125
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	11/21/25 13:04	VX112125
141-78-6	Ethyl Acetate	0.31	U	1	0.31	1.00	ug/L	11/21/25 13:04	VX112125
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	11/21/25 13:04	VX112125
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/21/25 13:04	VX112125
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/21/25 13:04	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:04	VX112125
75-65-0	Tert butyl alcohol	5.50	U	1	5.50	25.0	ug/L	11/21/25 13:04	VX112125
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:04	VX112125
107-02-8	Acrolein	7.10	U	1	7.10	25.0	ug/L	11/21/25 13:04	VX112125
107-13-1	Acrylonitrile	0.83	U	1	0.83	5.00	ug/L	11/21/25 13:04	VX112125
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	11/21/25 13:04	VX112125
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/21/25 13:04	VX112125
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:04	VX112125
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	11/21/25 13:04	VX112125
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/21/25 13:04	VX112125
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:04	VX112125
108-05-4	Vinyl Acetate	0.66	U	1	0.66	5.00	ug/L	11/21/25 13:04	VX112125
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:04	VX112125
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/21/25 13:04	VX112125
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/21/25 13:04	VX112125
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:04	VX112125
594-20-7	2,2-Dichloropropane	0.21	U	1	0.21	1.00	ug/L	11/21/25 13:04	VX112125
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:04	VX112125
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:04	VX112125
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:04	VX112125
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:04	VX112125
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:04	VX112125
563-58-6	1,1-Dichloropropene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:04	VX112125
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:04	VX112125
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:04	VX112125
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/21/25 13:04	VX112125
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:04	VX112125
74-95-3	Dibromomethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:04	VX112125
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:04	VX112125
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	11/21/25 13:04	VX112125
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:04	VX112125
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	11/21/25 13:04	VX112125
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:04	VX112125
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	11/21/25 13:04	VX112125
142-28-9	1,3-Dichloropropane	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:04	VX112125
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	1	0.30	5.00	ug/L	11/21/25 13:04	VX112125

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-SOIL-1
 Lab Sample ID: Q3702-01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	11/21/25 13:04	VX112125
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	11/21/25 13:04	VX112125
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:04	VX112125
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:04	VX112125
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 13:04	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:04	VX112125
67-72-1	Hexachloroethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 13:04	VX112125
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:04	VX112125
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	11/21/25 13:04	VX112125
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	11/21/25 13:04	VX112125
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:04	VX112125
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:04	VX112125
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 13:04	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/21/25 13:04	VX112125
96-18-4	1,2,3-Trichloropropane	0.35	U	1	0.35	1.00	ug/L	11/21/25 13:04	VX112125
108-86-1	Bromobenzene	0.24	U	1	0.24	1.00	ug/L	11/21/25 13:04	VX112125
103-65-1	n-propylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:04	VX112125
95-49-8	2-Chlorotoluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:04	VX112125
108-67-8	1,3,5-Trimethylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:04	VX112125
106-43-4	4-Chlorotoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:04	VX112125
98-06-6	tert-Butylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:04	VX112125
95-63-6	1,2,4-Trimethylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:04	VX112125
135-98-8	sec-Butylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:04	VX112125
99-87-6	p-Isopropyltoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:04	VX112125
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:04	VX112125
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:04	VX112125
104-51-8	n-Butylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:04	VX112125
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:04	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/21/25 13:04	VX112125
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:04	VX112125
87-68-3	Hexachlorobutadiene	0.30	U	1	0.30	1.00	ug/L	11/21/25 13:04	VX112125
91-20-3	Naphthalene	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:04	VX112125
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:04	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	62.2		74 - 125	124%	SPK: 50
1868-53-7	Dibromofluoromethane	44.9		75 - 124	90%	SPK: 50
2037-26-5	Toluene-d8	49.0		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	58.3		77 - 121	117%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	161000
540-36-3	1,4-Difluorobenzene	334000
3114-55-4	Chlorobenzene-d5	358000
3855-82-1	1,4-Dichlorobenzene-d4	192000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Chemtech Consulting Group
Project: Weekly Storage Blanks
Client Sample ID: STORAGE-BLANK-SOIL-]
Lab Sample ID: Q3702-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/14/25
Date Received: 11/21/25
SDG No.: Q3702
Matrix: Water
% Solid: 0
Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048651.D
 Acq On : 21 Nov 2025 13:04
 Operator : JC/MD
 Sample : Q3702-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Nov 21 22:12:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	160862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	334385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	357734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	192057	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	139449	62.213	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	124.420%#
35) Dibromofluoromethane	5.367	113	113700	44.926	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.860%
50) Toluene-d8	8.635	98	386416	48.955	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.920%
62) 4-Bromofluorobenzene	11.061	95	171452	58.286	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.580%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048651.D
Acq On : 21 Nov 2025 13:04
Operator : JC/MD
Sample : Q3702-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Nov 21 22:12:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.531 min Scan# 71
Delta R.T. -0.007 min
Lab File: VX048651.D
Acq: 21 Nov 2025 13:04




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT


Tgt Ion:168 Resp: 160862
Ion Ratio Lower Upper
168 100
99 59.0 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 62.213 ug/l
RT: 5.934 min Scan# 795
Delta R.T. -0.012 min
Lab File: VX048651.D
Acq: 21 Nov 2025 13:04

Tgt Ion: 65 Resp: 139449
Ion Ratio Lower Upper
65 100
67 50.7 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.745 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX048651.D

Acq: 21 Nov 2025 13:04

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANAGEMENT





Tgt Ion:114 Resp: 334385

Ion Ratio Lower Upper

114 100

63 20.5 0.0 41.2

88 16.1 0.0 32.2



#35

Dibromofluoromethane

Concen: 44.926 ug/l

RT: 5.367 min Scan# 702

Delta R.T. -0.006 min

Lab File: VX048651.D

Acq: 21 Nov 2025 13:04




Tgt Ion:113 Resp: 113700


Ion Ratio Lower Upper

113 100

111 102.8 81.9 122.9

192 18.9 15.8 23.6





#50

Toluene-d8

Concen: 48.955 ug/l

RT: 8.635 min Scan# 1238

Delta R.T. -0.000 min

Lab File: VX048651.D

Acq: 21 Nov 2025 13:04

Instrument :

MSVOA_X

ClientSampleId :





STORAGE-BLANK-SAMPLE-MANAGEMENT

Tgt Ion: 98 Resp: 386416

Ion Ratio Lower Upper

98 100

100 65.1 53.6 80.4



#62

4-Bromofluorobenzene

Concen: 58.286 ug/l

RT: 11.061 min Scan# 1636

Delta R.T. -0.000 min

Lab File: VX048651.D

Acq: 21 Nov 2025 13:04




Tgt Ion: 95 Resp: 171452


Ion Ratio Lower Upper

95 100

174 78.4 0.0 161.8

176 74.5 0.0 153.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.037 min Scan# 1468

Delta R.T. -0.000 min

Lab File: VX048651.D

Acq: 21 Nov 2025 13:04

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANAGEMENT





Tgt Ion:117 Resp: 357734

Ion Ratio Lower Upper

117 100

82 59.1 44.4 66.6

119 33.7 25.1 37.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX048651.D

Acq: 21 Nov 2025 13:04




Tgt Ion:152 Resp: 192057

Ion Ratio Lower Upper

152 100

115 60.7 39.2 117.6

150 155.8 0.0 347.0



Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-WATE
 Lab Sample ID: Q3702-02
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:25	VX112125
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 13:25	VX112125
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	11/21/25 13:25	VX112125
141-78-6	Ethyl Acetate	0.31	U	1	0.31	1.00	ug/L	11/21/25 13:25	VX112125
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	11/21/25 13:25	VX112125
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/21/25 13:25	VX112125
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/21/25 13:25	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:25	VX112125
75-65-0	Tert butyl alcohol	5.50	U	1	5.50	25.0	ug/L	11/21/25 13:25	VX112125
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:25	VX112125
107-02-8	Acrolein	7.10	U	1	7.10	25.0	ug/L	11/21/25 13:25	VX112125
107-13-1	Acrylonitrile	0.83	U	1	0.83	5.00	ug/L	11/21/25 13:25	VX112125
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	11/21/25 13:25	VX112125
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/21/25 13:25	VX112125
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:25	VX112125
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	11/21/25 13:25	VX112125
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/21/25 13:25	VX112125
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:25	VX112125
108-05-4	Vinyl Acetate	0.66	U	1	0.66	5.00	ug/L	11/21/25 13:25	VX112125
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:25	VX112125
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/21/25 13:25	VX112125
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/21/25 13:25	VX112125
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:25	VX112125
594-20-7	2,2-Dichloropropane	0.21	U	1	0.21	1.00	ug/L	11/21/25 13:25	VX112125
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:25	VX112125
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:25	VX112125
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:25	VX112125
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:25	VX112125
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:25	VX112125
563-58-6	1,1-Dichloropropene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:25	VX112125
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:25	VX112125
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:25	VX112125
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/21/25 13:25	VX112125
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:25	VX112125
74-95-3	Dibromomethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 13:25	VX112125
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 13:25	VX112125
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	11/21/25 13:25	VX112125
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:25	VX112125
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	11/21/25 13:25	VX112125
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:25	VX112125
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	11/21/25 13:25	VX112125
142-28-9	1,3-Dichloropropane	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:25	VX112125
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	1	0.30	5.00	ug/L	11/21/25 13:25	VX112125

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-WATE
 Lab Sample ID: Q3702-02
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	11/21/25 13:25	VX112125
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	11/21/25 13:25	VX112125
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:25	VX112125
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 13:25	VX112125
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 13:25	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:25	VX112125
67-72-1	Hexachloroethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 13:25	VX112125
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:25	VX112125
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	11/21/25 13:25	VX112125
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	11/21/25 13:25	VX112125
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:25	VX112125
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:25	VX112125
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 13:25	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/21/25 13:25	VX112125
96-18-4	1,2,3-Trichloropropane	0.35	U	1	0.35	1.00	ug/L	11/21/25 13:25	VX112125
108-86-1	Bromobenzene	0.24	U	1	0.24	1.00	ug/L	11/21/25 13:25	VX112125
103-65-1	n-propylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:25	VX112125
95-49-8	2-Chlorotoluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:25	VX112125
108-67-8	1,3,5-Trimethylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:25	VX112125
106-43-4	4-Chlorotoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:25	VX112125
98-06-6	tert-Butylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:25	VX112125
95-63-6	1,2,4-Trimethylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 13:25	VX112125
135-98-8	sec-Butylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:25	VX112125
99-87-6	p-Isopropyltoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 13:25	VX112125
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:25	VX112125
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	11/21/25 13:25	VX112125
104-51-8	n-Butylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 13:25	VX112125
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 13:25	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/21/25 13:25	VX112125
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:25	VX112125
87-68-3	Hexachlorobutadiene	0.30	U	1	0.30	1.00	ug/L	11/21/25 13:25	VX112125
91-20-3	Naphthalene	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:25	VX112125
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 13:25	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	50.3		74 - 125	101%	SPK: 50
1868-53-7	Dibromofluoromethane	42.7		75 - 124	85%	SPK: 50
2037-26-5	Toluene-d8	49.0		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.2		77 - 121	110%	SPK: 50

INTERNAL STANDARDS**Area Count**

363-72-4	Pentafluorobenzene	170000
540-36-3	1,4-Difluorobenzene	333000
3114-55-4	Chlorobenzene-d5	361000
3855-82-1	1,4-Dichlorobenzene-d4	189000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	11/14/25
Project:	Weekly Storage Blanks	Date Received:	11/21/25
Client Sample ID:	STORAGE-BLANK-WATE	SDG No.:	Q3702
Lab Sample ID:	Q3702-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 mL	Test:	VOC- Chemtech Full
	Level: LOW		
	Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048652.D
 Acq On : 21 Nov 2025 13:25
 Operator : JC/MD
 Sample : Q3702-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 21 22:13:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	170410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	333480	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	361276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	189089	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	119507	50.329	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.660%
35) Dibromofluoromethane	5.373	113	107865	42.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.480%
50) Toluene-d8	8.634	98	385595	48.984	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.061	95	162029	55.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.460%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048652.D
Acq On : 21 Nov 2025 13:25
Operator : JC/MD
Sample : Q3702-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 21 22:13:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.537 min Scan# 710
Delta R.T. -0.001 min
Lab File: VX048652.D
Acq: 21 Nov 2025 13:25




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion:168 Resp: 170410
Ion Ratio Lower Upper
168 100
99 55.5 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 50.329 ug/l
RT: 5.940 min Scan# 796
Delta R.T. -0.006 min
Lab File: VX048652.D
Acq: 21 Nov 2025 13:25

Tgt Ion: 65 Resp: 119507
Ion Ratio Lower Upper
65 100
67 52.5 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.745 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX048652.D

Acq: 21 Nov 2025 13:25

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-5





Tgt Ion:114 Resp: 333480

Ion Ratio Lower Upper

114 100

63 18.9 0.0 41.2

88 15.5 0.0 32.2



#35

Dibromofluoromethane

Concen: 42.736 ug/l

RT: 5.373 min Scan# 703

Delta R.T. -0.000 min

Lab File: VX048652.D

Acq: 21 Nov 2025 13:25




Tgt Ion:113 Resp: 107865


Ion Ratio Lower Upper

113 100

111 101.2 81.9 122.9

192 20.5 15.8 23.6





#50

Toluene-d8

Concen: 48.984 ug/l

RT: 8.634 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048652.D

Acq: 21 Nov 2025 13:25

Instrument :

MSVOA_X

ClientSampleId :





STORAGE-BLANK-WATER-REF-5

Tgt Ion: 98 Resp: 385595

Ion Ratio Lower Upper

98 100

100 65.0 53.6 80.4



#62

4-Bromofluorobenzene

Concen: 55.232 ug/l

RT: 11.061 min Scan# 1636

Delta R.T. -0.000 min

Lab File: VX048652.D

Acq: 21 Nov 2025 13:25




Tgt Ion: 95 Resp: 162029


Ion Ratio Lower Upper

95 100

174 81.3 0.0 161.8

176 77.8 0.0 153.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.037 min Scan# 1468

Delta R.T. -0.000 min

Lab File: VX048652.D

Acq: 21 Nov 2025 13:25

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-5





Tgt Ion:117 Resp: 361276

Ion Ratio Lower Upper

117 100

82 57.2 44.4 66.6

119 32.3 25.1 37.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX048652.D

Acq: 21 Nov 2025 13:25




Tgt Ion:152 Resp: 189089

Ion Ratio Lower Upper

152 100

115 58.2 39.2 117.6

150 155.4 0.0 347.0



Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-WATE
 Lab Sample ID: Q3702-03
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:44	VX112125
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 14:44	VX112125
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	11/21/25 14:44	VX112125
141-78-6	Ethyl Acetate	0.31	U	1	0.31	1.00	ug/L	11/21/25 14:44	VX112125
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	11/21/25 14:44	VX112125
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/21/25 14:44	VX112125
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/21/25 14:44	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:44	VX112125
75-65-0	Tert butyl alcohol	5.50	U	1	5.50	25.0	ug/L	11/21/25 14:44	VX112125
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:44	VX112125
107-02-8	Acrolein	7.10	U	1	7.10	25.0	ug/L	11/21/25 14:44	VX112125
107-13-1	Acrylonitrile	0.83	U	1	0.83	5.00	ug/L	11/21/25 14:44	VX112125
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	11/21/25 14:44	VX112125
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/21/25 14:44	VX112125
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:44	VX112125
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	11/21/25 14:44	VX112125
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/21/25 14:44	VX112125
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:44	VX112125
108-05-4	Vinyl Acetate	0.66	U	1	0.66	5.00	ug/L	11/21/25 14:44	VX112125
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:44	VX112125
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/21/25 14:44	VX112125
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/21/25 14:44	VX112125
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:44	VX112125
594-20-7	2,2-Dichloropropane	0.21	U	1	0.21	1.00	ug/L	11/21/25 14:44	VX112125
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:44	VX112125
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:44	VX112125
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:44	VX112125
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:44	VX112125
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:44	VX112125
563-58-6	1,1-Dichloropropene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:44	VX112125
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:44	VX112125
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:44	VX112125
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/21/25 14:44	VX112125
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:44	VX112125
74-95-3	Dibromomethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:44	VX112125
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:44	VX112125
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	11/21/25 14:44	VX112125
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:44	VX112125
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	11/21/25 14:44	VX112125
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:44	VX112125
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	11/21/25 14:44	VX112125
142-28-9	1,3-Dichloropropane	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:44	VX112125
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	1	0.30	5.00	ug/L	11/21/25 14:44	VX112125

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-WATE
 Lab Sample ID: Q3702-03
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	11/21/25 14:44	VX112125
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	11/21/25 14:44	VX112125
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:44	VX112125
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:44	VX112125
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 14:44	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:44	VX112125
67-72-1	Hexachloroethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 14:44	VX112125
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:44	VX112125
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	11/21/25 14:44	VX112125
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	11/21/25 14:44	VX112125
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:44	VX112125
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:44	VX112125
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 14:44	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/21/25 14:44	VX112125
96-18-4	1,2,3-Trichloropropane	0.35	U	1	0.35	1.00	ug/L	11/21/25 14:44	VX112125
108-86-1	Bromobenzene	0.24	U	1	0.24	1.00	ug/L	11/21/25 14:44	VX112125
103-65-1	n-propylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:44	VX112125
95-49-8	2-Chlorotoluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:44	VX112125
108-67-8	1,3,5-Trimethylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:44	VX112125
106-43-4	4-Chlorotoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:44	VX112125
98-06-6	tert-Butylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:44	VX112125
95-63-6	1,2,4-Trimethylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:44	VX112125
135-98-8	sec-Butylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:44	VX112125
99-87-6	p-Isopropyltoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:44	VX112125
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:44	VX112125
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:44	VX112125
104-51-8	n-Butylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:44	VX112125
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:44	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/21/25 14:44	VX112125
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:44	VX112125
87-68-3	Hexachlorobutadiene	0.30	U	1	0.30	1.00	ug/L	11/21/25 14:44	VX112125
91-20-3	Naphthalene	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:44	VX112125
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:44	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	54.7		74 - 125	109%	SPK: 50
1868-53-7	Dibromofluoromethane	43.3		75 - 124	87%	SPK: 50
2037-26-5	Toluene-d8	47.0		86 - 113	94%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.9		77 - 121	110%	SPK: 50

INTERNAL STANDARDS**Area Count**

363-72-4	Pentafluorobenzene	160000
540-36-3	1,4-Difluorobenzene	322000
3114-55-4	Chlorobenzene-d5	341000
3855-82-1	1,4-Dichlorobenzene-d4	184000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	11/14/25
Project:	Weekly Storage Blanks	Date Received:	11/21/25
Client Sample ID:	STORAGE-BLANK-WATE	SDG No.:	Q3702
Lab Sample ID:	Q3702-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 mL	Test:	VOC- Chemtech Full
	Level: LOW		
	Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048655.D
 Acq On : 21 Nov 2025 14:44
 Operator : JC/MD
 Sample : Q3702-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 21 22:14:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	160298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	322084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	341432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	184333	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	122205	54.711	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.420%
35) Dibromofluoromethane	5.367	113	105455	43.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.520%
50) Toluene-d8	8.628	98	357166	46.978	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.960%
62) 4-Bromofluorobenzene	11.061	95	155597	54.916	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.840%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048655.D
Acq On : 21 Nov 2025 14:44
Operator : JC/MD
Sample : Q3702-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 21 22:14:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.531 min Scan# 71
Delta R.T. -0.007 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 160298
Ion Ratio Lower Upper
168 100
99 58.2 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 54.711 ug/l
RT: 5.934 min Scan# 795
Delta R.T. -0.012 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Tgt Ion: 65 Resp: 122205
Ion Ratio Lower Upper
65 100
67 51.0 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.739 min Scan# 91

Delta R.T. -0.006 min

Lab File: VX048655.D

Acq: 21 Nov 2025 14:44

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-4





Tgt Ion:114 Resp: 322084

Ion Ratio Lower Upper

114 100

63 20.2 0.0 41.2

88 15.6 0.0 32.2



#35

Dibromofluoromethane

Concen: 43.259 ug/l

RT: 5.367 min Scan# 702

Delta R.T. -0.006 min

Lab File: VX048655.D

Acq: 21 Nov 2025 14:44




Tgt Ion:113 Resp: 105455


Ion Ratio Lower Upper

113 100

111 102.4 81.9 122.9

192 19.7 15.8 23.6









#50
Toluene-d8
Concen: 46.978 ug/l
RT: 8.628 min Scan# 1237
Delta R.T. -0.006 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 357166
Ion Ratio Lower Upper
98 100
100 66.3 53.6 80.4



#62
4-Bromofluorobenzene
Concen: 54.916 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. -0.000 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Tgt Ion: 95 Resp: 155597
Ion Ratio Lower Upper
95 100
174 79.9 0.0 161.8
176 77.4 0.0 153.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.037 min Scan# 1468

Delta R.T. -0.000 min

Lab File: VX048655.D

Acq: 21 Nov 2025 14:44

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-4





Tgt Ion:117 Resp: 341432

Ion Ratio Lower Upper

117 100

82 56.1 44.4 66.6

119 32.4 25.1 37.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX048655.D

Acq: 21 Nov 2025 14:44




Tgt Ion:152 Resp: 184333

Ion Ratio Lower Upper

152 100

115 58.9 39.2 117.6

150 157.7 0.0 347.0



Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: STORAGE-BLANK-SAMP
 Lab Sample ID: Q3702-04
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/14/25
 Date Received: 11/21/25
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:07	VX112125
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 14:07	VX112125
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	11/21/25 14:07	VX112125
141-78-6	Ethyl Acetate	0.31	U	1	0.31	1.00	ug/L	11/21/25 14:07	VX112125
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	11/21/25 14:07	VX112125
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/21/25 14:07	VX112125
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/21/25 14:07	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:07	VX112125
75-65-0	Tert butyl alcohol	5.50	U	1	5.50	25.0	ug/L	11/21/25 14:07	VX112125
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:07	VX112125
107-02-8	Acrolein	7.10	U	1	7.10	25.0	ug/L	11/21/25 14:07	VX112125
107-13-1	Acrylonitrile	0.83	U	1	0.83	5.00	ug/L	11/21/25 14:07	VX112125
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	11/21/25 14:07	VX112125
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/21/25 14:07	VX112125
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:07	VX112125
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	11/21/25 14:07	VX112125
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/21/25 14:07	VX112125
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:07	VX112125
108-05-4	Vinyl Acetate	0.66	U	1	0.66	5.00	ug/L	11/21/25 14:07	VX112125
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:07	VX112125
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/21/25 14:07	VX112125
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/21/25 14:07	VX112125
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:07	VX112125
594-20-7	2,2-Dichloropropane	0.21	U	1	0.21	1.00	ug/L	11/21/25 14:07	VX112125
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:07	VX112125
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:07	VX112125
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:07	VX112125
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:07	VX112125
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:07	VX112125
563-58-6	1,1-Dichloropropene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:07	VX112125
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:07	VX112125
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:07	VX112125
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/21/25 14:07	VX112125
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:07	VX112125
74-95-3	Dibromomethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 14:07	VX112125
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 14:07	VX112125
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	11/21/25 14:07	VX112125
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:07	VX112125
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	11/21/25 14:07	VX112125
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:07	VX112125
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	11/21/25 14:07	VX112125
142-28-9	1,3-Dichloropropane	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:07	VX112125
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	1	0.30	5.00	ug/L	11/21/25 14:07	VX112125



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Chemtech Consulting Group
Project: Weekly Storage Blanks
Client Sample ID: STORAGE-BLANK-SAMP
Lab Sample ID: Q3702-04
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/14/25
Date Received: 11/21/25
SDG No.: Q3702
Matrix: Water
% Solid: 0
Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	11/21/25 14:07	VX112125
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	11/21/25 14:07	VX112125
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:07	VX112125
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 14:07	VX112125
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 14:07	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:07	VX112125
67-72-1	Hexachloroethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 14:07	VX112125
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:07	VX112125
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	11/21/25 14:07	VX112125
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	11/21/25 14:07	VX112125
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:07	VX112125
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:07	VX112125
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 14:07	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/21/25 14:07	VX112125
96-18-4	1,2,3-Trichloropropane	0.35	U	1	0.35	1.00	ug/L	11/21/25 14:07	VX112125
108-86-1	Bromobenzene	0.24	U	1	0.24	1.00	ug/L	11/21/25 14:07	VX112125
103-65-1	n-propylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:07	VX112125
95-49-8	2-Chlorotoluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:07	VX112125
108-67-8	1,3,5-Trimethylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:07	VX112125
106-43-4	4-Chlorotoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:07	VX112125
98-06-6	tert-Butylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:07	VX112125
95-63-6	1,2,4-Trimethylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 14:07	VX112125
135-98-8	sec-Butylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:07	VX112125
99-87-6	p-Isopropyltoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 14:07	VX112125
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:07	VX112125
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	11/21/25 14:07	VX112125
104-51-8	n-Butylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 14:07	VX112125
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 14:07	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/21/25 14:07	VX112125
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:07	VX112125
87-68-3	Hexachlorobutadiene	0.30	U	1	0.30	1.00	ug/L	11/21/25 14:07	VX112125
91-20-3	Naphthalene	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:07	VX112125
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 14:07	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	51.7		74 - 125	103%	SPK: 50
1868-53-7	Dibromofluoromethane	42.5		75 - 124	85%	SPK: 50
2037-26-5	Toluene-d8	47.8		86 - 113	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.8		77 - 121	110%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	157000
540-36-3	1,4-Difluorobenzene	308000
3114-55-4	Chlorobenzene-d5	329000
3855-82-1	1,4-Dichlorobenzene-d4	176000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Chemtech Consulting Group
Project: Weekly Storage Blanks
Client Sample ID: STORAGE-BLANK-SAMP
Lab Sample ID: Q3702-04
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/14/25
Date Received: 11/21/25
SDG No.: Q3702
Matrix: Water
% Solid: 0
Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048654.D
 Acq On : 21 Nov 2025 14:07
 Operator : JC/MD
 Sample : Q3702-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Nov 21 22:14:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	157000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	308197	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	328677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	175625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	113120	51.708	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.420%	
35) Dibromofluoromethane	5.373	113	99184	42.520	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 85.040%	
50) Toluene-d8	8.635	98	347960	47.829	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.660%	
62) 4-Bromofluorobenzene	11.067	95	148668	54.835	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.680%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048654.D
Acq On : 21 Nov 2025 14:07
Operator : JC/MD
Sample : Q3702-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Nov 21 22:14:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. 0.006 min
Lab File: VX048654.D
Acq: 21 Nov 2025 14:07




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT


Tgt Ion:168 Resp: 157000
Ion Ratio Lower Upper
168 100
99 59.8 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 51.708 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.000 min
Lab File: VX048654.D
Acq: 21 Nov 2025 14:07

Tgt Ion: 65 Resp: 113120
Ion Ratio Lower Upper
65 100
67 51.4 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.745 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX048654.D

Acq: 21 Nov 2025 14:07

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANAGEMENT





Tgt Ion:114 Resp: 308197

Ion Ratio Lower Upper

114 100

63 21.3 0.0 41.2

88 15.3 0.0 32.2



#35

Dibromofluoromethane

Concen: 42.520 ug/l

RT: 5.373 min Scan# 703

Delta R.T. -0.000 min

Lab File: VX048654.D

Acq: 21 Nov 2025 14:07




Tgt Ion:113 Resp: 99184


Ion Ratio Lower Upper

113 100

111 103.5 81.9 122.9

192 18.9 15.8 23.6





#50

Toluene-d8

Concen: 47.829 ug/l

RT: 8.635 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048654.D

Acq: 21 Nov 2025 14:07

Instrument :

MSVOA_X

ClientSampleId :





STORAGE-BLANK-SAMPLE-MANAGEMENT

Tgt Ion: 98 Resp: 347960

Ion Ratio Lower Upper

98 100

100 65.6 53.6 80.4



#62

4-Bromofluorobenzene

Concen: 54.835 ug/l

RT: 11.067 min Scan# 1637

Delta R.T. 0.006 min

Lab File: VX048654.D

Acq: 21 Nov 2025 14:07




Tgt Ion: 95 Resp: 148668


Ion Ratio Lower Upper

95 100

174 81.4 0.0 161.8

176 78.5 0.0 153.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.037 min Scan# 1468

Delta R.T. -0.000 min

Lab File: VX048654.D

Acq: 21 Nov 2025 14:07

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANAGEMENT





Tgt Ion:117 Resp: 328677

Ion Ratio Lower Upper

117 100

82 58.3 44.4 66.6

119 33.5 25.1 37.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX048654.D

Acq: 21 Nov 2025 14:07




Tgt Ion:152 Resp: 175625

Ion Ratio Lower Upper

152 100

115 58.5 39.2 117.6

150 156.4 0.0 347.0





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q3702
 Instrument ID: MSVOA_X Calibration Date(s): 11/07/2025 11/07/2025
 Heated Purge: (Y/N) N Calibration Time(s): 13:35 16:29
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX048516.D		RRF005 = VX048517.D		RRF020 = VX048518.D			
		RRF100 = VX048520.D		RRF150 = VX048521.D		RRF050 = VX048524.D			
COMPOUND	RRF001	RRF005	RRF020	RRF100	RRF150	RRF050	RRF	% RSD	
Dichlorodifluoromethane	0.365	0.364	0.352	0.452	0.474	0.424	0.405	12.8	
Chloromethane	0.603	0.457	0.425	0.581	0.565	0.592	0.537	14.2	
Vinyl Chloride	0.666	0.579	0.608	0.601	0.597	0.600	0.608	4.9	
Ethyl Acetate	0.748	0.901	0.676	0.648	0.639	0.667	0.713	14	
Bromomethane		0.468	0.417	0.323	0.287	0.426	0.384	19.8	
Chloroethane	0.465	0.419	0.397	0.370	0.357	0.383	0.398	9.8	
Trichlorofluoromethane	0.982	1.030	1.019	0.919	0.939	0.886	0.963	6	
1,1,2-Trichlorotrifluoroethane	0.552	0.556	0.616	0.546	0.558	0.515	0.557	5.9	
Tert butyl alcohol		0.129	0.133	0.132	0.127	0.138	0.132	3.2	
1,1-Dichloroethene	0.669	0.584	0.591	0.555	0.548	0.545	0.582	8	
Acrolein		0.023	0.015	0.017	0.020	0.020	0.019	16.4	
Acrylonitrile	0.405	0.383	0.395	0.370	0.356	0.403	0.385	5.1	
Acetone	0.383	0.318	0.318	0.317	0.313	0.330	0.330	8.1	
Carbon Disulfide	2.045	1.739	1.609	1.468	1.477	1.489	1.638	13.7	
Methyl tert-butyl Ether	2.163	2.158	2.049	1.984	1.861	2.117	2.055	5.7	
Methyl Acetate	1.195	0.876	0.857	0.818	0.823	0.869	0.906	15.8	
Methylene Chloride	0.845	0.689	0.663	0.624	0.595	0.653	0.678	13	
trans-1,2-Dichloroethene	0.700	0.650	0.612	0.585	0.557	0.598	0.617	8.3	
Vinyl Acetate	1.912	1.901	1.838	1.747	1.710	1.827	1.822	4.4	
1,1-Dichloroethane	1.255	1.247	1.124	1.083	1.041	1.102	1.142	7.8	
Cyclohexane		1.001	1.004	0.896	0.910	0.867	0.936	6.7	
2-Butanone	0.474	0.502	0.511	0.470	0.465	0.506	0.488	4.2	
Carbon Tetrachloride	0.640	0.766	0.546	0.520	0.504	0.486	0.577	18.6	
2,2-Dichloropropane	1.079	1.044	0.975	0.934	0.925	0.883	0.973	7.7	
cis-1,2-Dichloroethene	0.888	0.787	0.737	0.703	0.684	0.725	0.754	9.9	
Bromochloromethane	0.640	0.630	0.533	0.522	0.497	0.543	0.561	10.6	
Chloroform	1.357	1.277	1.162	1.105	1.052	1.120	1.179	9.8	
1,1,1-Trichloroethane	1.160	1.087	1.026	0.979	0.947	0.973	1.029	7.9	
Methylcyclohexane	0.585	0.607	0.558	0.561	0.592	0.510	0.569	6	
1,1-Dichloropropene	0.548	0.664	0.499	0.453	0.449	0.427	0.507	17.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q3702
 Instrument ID: MSVOA_X Calibration Date(s): 11/07/2025 11/07/2025
 Heated Purge: (Y/N) N Calibration Time(s): 13:35 16:29
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX048516.D		RRF005 = VX048517.D		RRF020 = VX048518.D			
		RRF100 = VX048520.D		RRF150 = VX048521.D		RRF050 = VX048524.D			
COMPOUND	RRF001	RRF005	RRF020	RRF100	RRF150	RRF050	RRF	% RSD	
Benzene	1.745	2.038	1.468	1.396	1.369	1.408	1.571	17	
1,2-Dichloroethane	0.540	0.721	0.525	0.489	0.473	0.481	0.538	17.3	
Trichloroethene	0.423	0.390	0.356	0.366	0.363	0.355	0.375	7.1	
1,2-Dichloropropane	0.417	0.381	0.319	0.352	0.340	0.358	0.361	9.5	
Dibromomethane	0.278	0.303	0.217	0.264	0.251	0.268	0.263	10.8	
Bromodichloromethane	0.517	0.598	0.446	0.539	0.510	0.543	0.526	9.5	
4-Methyl-2-Pentanone	0.505	0.667	0.514	0.573	0.546	0.620	0.571	11	
Toluene	0.709	0.955	0.710	0.868	0.843	0.877	0.827	11.9	
t-1,3-Dichloropropene	0.513	0.609	0.452	0.561	0.542	0.572	0.541	10	
cis-1,3-Dichloropropene	0.472	0.633	0.481	0.595	0.578	0.606	0.561	12.1	
1,1,2-Trichloroethane	0.289	0.373	0.286	0.345	0.329	0.361	0.330	11.1	
1,3-Dichloropropane	0.529	0.674	0.498	0.581	0.551	0.618	0.575	11.1	
2-Chloroethyl Vinyl ether	0.229	0.301	0.251	0.327	0.311	0.354	0.296	15.9	
2-Hexanone	0.360	0.487	0.390	0.429	0.416	0.464	0.425	11	
Dibromochloromethane	0.347	0.468	0.351	0.428	0.404	0.436	0.406	12	
1,2-Dibromoethane	0.309	0.417	0.308	0.374	0.358	0.381	0.358	11.9	
Tetrachloroethene	0.444	0.364	0.352	0.344	0.348	0.327	0.363	11.4	
Chlorobenzene	1.406	1.245	1.153	1.112	1.105	1.101	1.187	10.1	
1,1,1,2-Tetrachloroethane	0.396	0.420	0.389	0.398	0.391	0.392	0.398	2.9	
Hexachloroethane	0.679	0.611	0.600	0.597	0.596	0.551	0.606	6.8	
Ethyl Benzene	2.253	2.045	1.976	1.899	1.898	1.887	1.993	7.1	
m/p-Xylenes	0.764	0.793	0.758	0.736	0.733	0.720	0.751	3.5	
o-Xylene	0.795	0.746	0.721	0.714	0.726	0.703	0.734	4.5	
Styrene	1.228	1.256	1.227	1.226	1.244	1.210	1.232	1.3	
Bromoform	0.427	0.350	0.345	0.359	0.373	0.348	0.367	8.4	
Isopropylbenzene	3.857	3.727	3.683	3.551	3.492	3.466	3.629	4.2	
1,1,1,2,2-Tetrachloroethane	1.544	1.284	1.230	1.207	1.185	1.229	1.280	10.4	
1,2,3-Trichloropropane	1.270	1.113	1.134	1.128	1.112	1.010	1.128	7.4	
Bromobenzene	1.054	0.961	0.913	0.909	0.877	0.909	0.937	6.7	
n-propylbenzene	4.682	4.356	4.371	4.097	4.073	4.027	4.268	5.9	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q3702
 Instrument ID: MSVOA_X Calibration Date(s): 11/07/2025 11/07/2025
 Heated Purge: (Y/N) N Calibration Time(s): 13:35 16:29
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX048516.D		RRF005 = VX048517.D		RRF020 = VX048518.D			
		RRF100 = VX048520.D		RRF150 = VX048521.D		RRF050 = VX048524.D			
COMPOUND	RRF001	RRF005	RRF020	RRF100	RRF150	RRF050	RRF	% RSD	
2-Chlorotoluene	2.995	2.601	2.565	2.493	2.382	2.437	2.579	8.5	
1,3,5-Trimethylbenzene	3.172	2.990	3.003	2.913	2.855	2.819	2.959	4.3	
4-Chlorotoluene	3.437	3.105	2.985	2.916	2.842	2.866	3.025	7.4	
tert-Butylbenzene	3.231	3.058	3.080	3.036	3.010	2.945	3.060	3.1	
1,2,4-Trimethylbenzene	3.224	2.963	2.991	2.951	2.886	2.897	2.985	4.1	
sec-Butylbenzene	4.137	3.883	3.841	3.643	3.693	3.496	3.782	5.9	
p-Isopropyltoluene	3.445	3.138	3.226	3.123	3.159	2.988	3.180	4.8	
1,3-Dichlorobenzene	2.076	1.785	1.686	1.680	1.668	1.644	1.756	9.3	
1,4-Dichlorobenzene	2.329	1.860	1.676	1.671	1.672	1.678	1.814	14.5	
n-Butylbenzene	3.281	3.038	3.007	2.883	2.972	2.718	2.983	6.2	
1,2-Dichlorobenzene	1.944	1.695	1.611	1.627	1.616	1.624	1.686	7.7	
1,2-Dibromo-3-Chloropropane	0.304	0.308	0.294	0.298	0.285	0.297	0.297	2.7	
1,2,4-Trichlorobenzene	1.353	1.032	1.057	1.097	1.117	1.066	1.120	10.5	
Hexachlorobutadiene	0.611	0.467	0.450	0.421	0.455	0.406	0.468	15.7	
Naphthalene	3.675	3.299	3.529	3.732	3.682	3.700	3.603	4.6	
1,2,3-Trichlorobenzene	1.170	0.973	0.994	1.049	1.051	1.016	1.042	6.7	
1,2-Dichloroethane-d4		0.775	0.723	0.677	0.623	0.685	0.697	8.1	
Dibromofluoromethane		0.484	0.374	0.356	0.327	0.351	0.378	16.2	
Toluene-d8		1.341	1.021	1.210	1.128	1.202	1.180	10	
4-Bromofluorobenzene		0.523	0.365	0.445	0.423	0.443	0.440	12.9	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X110725W.M
 Title : SW846 8260
 Last Update : Sat Nov 08 05:08:11 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX048516.D 5 =VX048517.D 20 =VX048518.D 50 =VX048524.D 100 =VX048520.D 150 =VX048521.D

Compound		1	5	20	50	100	150	Avg	%RSD

1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.365	0.364	0.352	0.424	0.452	0.474	0.405	12.78
3) P	Chloromethane	0.603	0.457	0.425	0.592	0.581	0.565	0.537	14.19
4) C	Vinyl Chloride	0.666	0.579	0.608	0.600	0.601	0.597	0.608	4.86#
5) T	Bromomethane		0.468	0.417	0.426	0.323	0.287	0.384	19.75
6) T	Chloroethane	0.465	0.419	0.397	0.383	0.370	0.357	0.398	9.77
7) T	Trichlorofluor...	0.982	1.030	1.019	0.886	0.919	0.939	0.963	5.98
8) T	Diethyl Ether	0.397	0.420	0.379	0.393	0.376	0.353	0.386	5.87
9) T	1,1,2-Trichlor...	0.552	0.556	0.616	0.515	0.546	0.558	0.557	5.91
10) T	Methyl Iodide		0.833	0.822	0.824	0.836	0.808	0.825	1.34
11) T	Tert butyl alc...		0.129	0.133	0.138	0.132	0.127	0.132	3.25
12) CM	1,1-Dichloroet...	0.669	0.584	0.591	0.545	0.555	0.548	0.582	8.02#
13) T	Acrolein		0.023	0.015	0.020	0.017	0.020	0.019	16.42
14) T	Allyl chloride	1.243	1.129	1.065	1.048	1.010	1.008	1.084	8.27
15) T	Acrylonitrile	0.405	0.383	0.395	0.403	0.370	0.356	0.385	5.05
16) T	Acetone	0.383	0.318	0.318	0.330	0.317	0.313	0.330	8.07
17) T	Carbon Disulfide	2.045	1.739	1.609	1.489	1.468	1.477	1.638	13.75
18) T	Methyl Acetate	1.195	0.876	0.857	0.869	0.818	0.823	0.906	15.80
19) T	Methyl tert-bu...	2.163	2.158	2.049	2.117	1.984	1.861	2.055	5.70
20) T	Methylene Chlo...	0.845	0.689	0.663	0.653	0.624	0.595	0.678	13.00
21) T	trans-1,2-Dich...	0.700	0.650	0.612	0.598	0.585	0.557	0.617	8.29
22) T	Diisopropyl ether	2.220	2.180	2.105	2.064	1.977	1.939	2.081	5.30
23) T	Vinyl Acetate	1.912	1.901	1.838	1.827	1.747	1.710	1.822	4.43
24) P	1,1-Dichloroet...	1.255	1.247	1.124	1.102	1.083	1.041	1.142	7.77
25) T	2-Butanone	0.474	0.502	0.511	0.506	0.470	0.465	0.488	4.20
26) T	2,2-Dichloropr...	1.079	1.044	0.975	0.883	0.934	0.925	0.973	7.71
27) T	cis-1,2-Dichlo...	0.888	0.787	0.737	0.725	0.703	0.684	0.754	9.90
28) T	Bromochloromet...	0.640	0.630	0.533	0.543	0.522	0.497	0.561	10.61
29) T	Tetrahydrofuran	0.406	0.357	0.350	0.349	0.321	0.306	0.348	9.91
30) C	Chloroform	1.357	1.277	1.162	1.120	1.105	1.052	1.179	9.79#
31) T	Cyclohexane		1.001	1.004	0.867	0.896	0.910	0.936	6.73
32) T	1,1,1-Trichlor...	1.160	1.087	1.026	0.973	0.979	0.947	1.029	7.90
33) S	1,2-Dichloroet...		0.775	0.723	0.685	0.677	0.623	0.697	8.08
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.484	0.374	0.351	0.356	0.327	0.378	16.22
36) T	1,1-Dichloropr...	0.548	0.664	0.499	0.427	0.453	0.449	0.507	17.46
37) T	Ethyl Acetate	0.748	0.901	0.676	0.667	0.648	0.639	0.713	13.96
38) T	Carbon Tetrach...	0.640	0.766	0.546	0.486	0.520	0.504	0.577	18.63
39) T	Methylcyclohexane	0.585	0.607	0.558	0.510	0.561	0.592	0.569	6.00
40) TM	Benzene	1.745	2.038	1.468	1.408	1.396	1.369	1.571	16.99
41) T	Methacrylonitrile	0.315	0.406	0.310	0.295	0.276	0.280	0.314	15.30
42) TM	1,2-Dichloroet...	0.540	0.721	0.525	0.481	0.489	0.473	0.538	17.33
43) T	Isopropyl Acetate	1.009	1.298	0.958	0.929	0.908	0.914	1.003	14.92
44) TM	Trichloroethene	0.423	0.390	0.356	0.355	0.366	0.363	0.375	7.10
45) C	1,2-Dichloropr...	0.417	0.381	0.319	0.358	0.352	0.340	0.361	9.45#
46) T	Dibromomethane	0.278	0.303	0.217	0.268	0.264	0.251	0.263	10.84
47) T	Bromodichlorom...	0.517	0.598	0.446	0.543	0.539	0.510	0.526	9.49
48) T	Methyl methacr...	0.467	0.476	0.385	0.462	0.448	0.435	0.446	7.40
49) T	1,4-Dioxane	0.006	0.007	0.006	0.007	0.008	0.007	0.007	12.02
50) S	Toluene-d8		1.341	1.021	1.202	1.210	1.128	1.180	9.96
51) T	4-Methyl-2-Pen...	0.505	0.667	0.514	0.620	0.573	0.546	0.571	11.02
52) CM	Toluene	0.709	0.955	0.710	0.877	0.868	0.843	0.827	11.87#
53) T	t-1,3-Dichloro...	0.513	0.609	0.452	0.572	0.561	0.542	0.541	10.02
54) T	cis-1,3-Dichlo...	0.472	0.633	0.481	0.606	0.595	0.578	0.561	12.05
55) T	1,1,2-Trichlor...	0.289	0.373	0.286	0.361	0.345	0.329	0.330	11.11
56) T	Ethyl methacry...	0.446	0.594	0.474	0.618	0.594	0.573	0.550	13.05

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X110725W.M

57)	T	1,3-Dichloropr...	0.529	0.674	0.498	0.618	0.581	0.551	0.575	11.11
58)	T	2-Chloroethyl ...	0.229	0.301	0.251	0.354	0.327	0.311	0.296	15.90
59)	T	2-Hexanone	0.360	0.487	0.390	0.464	0.429	0.416	0.425	10.96
60)	T	Dibromochlorom...	0.347	0.468	0.351	0.436	0.428	0.404	0.406	11.97
61)	T	1,2-Dibromoethane	0.309	0.417	0.308	0.381	0.374	0.358	0.358	11.93
62)	S	4-Bromofluorob...	0.523	0.365	0.443	0.445	0.423	0.440		12.88
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.444	0.364	0.352	0.327	0.344	0.348	0.363	11.38
65)	PM	Chlorobenzene	1.406	1.245	1.153	1.101	1.112	1.105	1.187	10.12
66)	T	1,1,1,2-Tetrac...	0.396	0.420	0.389	0.392	0.398	0.391	0.398	2.88
67)	C	Ethyl Benzene	2.253	2.045	1.976	1.887	1.899	1.898	1.993	7.08#
68)	T	m/p-Xylenes	0.764	0.793	0.758	0.720	0.736	0.733	0.751	3.49
69)	T	o-Xylene	0.795	0.746	0.721	0.703	0.714	0.726	0.734	4.50
70)	T	Styrene	1.228	1.256	1.227	1.210	1.226	1.244	1.232	1.30
71)	P	Bromoform	0.427	0.350	0.345	0.348	0.359	0.373	0.367	8.43
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.857	3.727	3.683	3.466	3.551	3.492	3.629	4.20
74)	T	N-amyl acetate	1.998	1.826	1.851	1.793	1.739	1.727	1.822	5.41
75)	P	1,1,2,2-Tetrac...	1.544	1.284	1.230	1.229	1.207	1.185	1.280	10.45
76)	T	1,2,3-Trichlor...	1.270	1.113	1.134	1.470	1.128	1.112	1.205	11.88
77)	T	Bromobenzene	1.054	0.961	0.913	0.909	0.909	0.877	0.937	6.74
78)	T	n-propylbenzene	4.682	4.356	4.371	4.027	4.097	4.073	4.268	5.89
79)	T	2-Chlorotoluene	2.995	2.601	2.565	2.437	2.493	2.382	2.579	8.50
80)	T	1,3,5-Trimethy...	3.172	2.990	3.003	2.819	2.913	2.855	2.959	4.30
81)	T	trans-1,4-Dich...	0.376	0.398	0.433	0.444	0.454	0.421		7.75
82)	T	4-Chlorotoluene	3.437	3.105	2.985	2.866	2.916	2.842	3.025	7.37
83)	T	tert-Butylbenzene	3.231	3.058	3.080	2.945	3.036	3.010	3.060	3.13
84)	T	1,2,4-Trimethy...	3.224	2.963	2.991	2.897	2.951	2.886	2.985	4.14
85)	T	sec-Butylbenzene	4.137	3.883	3.841	3.496	3.643	3.693	3.782	5.90
86)	T	p-Isopropyltol...	3.445	3.138	3.226	2.988	3.123	3.159	3.180	4.76
87)	T	1,3-Dichlorobe...	2.076	1.785	1.686	1.644	1.680	1.668	1.756	9.34
88)	T	1,4-Dichlorobe...	2.329	1.860	1.676	1.678	1.671	1.672	1.814	14.49
89)	T	n-Butylbenzene	3.281	3.038	3.007	2.718	2.883	2.972	2.983	6.23
90)	T	Hexachloroethane	0.679	0.611	0.600	0.551	0.597	0.596	0.606	6.82
91)	T	1,2-Dichlorobe...	1.944	1.695	1.611	1.624	1.627	1.616	1.686	7.71
92)	T	1,2-Dibromo-3-...	0.304	0.308	0.294	0.297	0.298	0.285	0.297	2.66
93)	T	1,2,4-Trichlor...	1.353	1.032	1.057	1.066	1.097	1.117	1.120	10.53
94)	T	Hexachlorobuta...	0.611	0.467	0.450	0.406	0.421	0.455	0.468	15.66
95)	T	Naphthalene	3.675	3.299	3.529	3.700	3.732	3.682	3.603	4.56
96)	T	1,2,3-Trichlor...	1.170	0.973	0.994	1.016	1.049	1.051	1.042	6.70

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048516.D
 Acq On : 07 Nov 2025 13:35
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/10/2025
 Supervised By : Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:38:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	195603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.738	114	335919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	223304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	110981	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	78 - 117	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	92 - 112	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	83 - 123	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.172	85	1427	0.900	ug/l	95
3) Chloromethane	1.300	50	2360	1.123	ug/l #	87
4) Vinyl Chloride	1.380	62	2604	1.094	ug/l	99
6) Chloroethane	1.697	64	1819	1.167	ug/l #	77
7) Trichlorofluoromethane	1.892	101	3842	1.020	ug/l	98
8) Diethyl Ether	2.130	74	1554	1.028	ug/l	85
9) 1,1,2-Trichlorotrifluo...	2.331	101	2160	0.991	ug/l #	90
12) 1,1-Dichloroethene	2.319	96	2617	1.149	ug/l #	86
14) Allyl chloride	2.666	41	4862	1.147	ug/l	95
15) Acrylonitrile	3.056	53	7915	5.250	ug/l	97
16) Acetone	2.373	43	7490	5.805	ug/l	98
17) Carbon Disulfide	2.514	76	8000	1.249	ug/l #	93
18) Methyl Acetate	2.703	43	4673	1.318	ug/l	93
19) Methyl tert-butyl Ether	3.105	73	8461	1.052	ug/l #	89
20) Methylene Chloride	2.788	84	3307	1.247	ug/l #	87
21) trans-1,2-Dichloroethene	3.087	96	2739	1.135	ug/l	97
22) Diisopropyl ether	3.757	45	8684	1.067	ug/l	86
23) Vinyl Acetate	3.721	43	37398	5.246	ug/l	97
24) 1,1-Dichloroethane	3.605	63	4909	1.099	ug/l	95
25) 2-Butanone	4.544	43	9269	4.857	ug/l	95
26) 2,2-Dichloropropane	4.464	77	4220	1.108	ug/l	92
27) cis-1,2-Dichloroethene	4.477	96	3475	1.178	ug/l	96
28) Bromochloromethane	4.885	49	2504	1.141	ug/l #	99
29) Tetrahydrofuran	4.995	42	7932	5.835	ug/l	92
30) Chloroform	5.080	83	5308	1.151	ug/l	84
32) 1,1,1-Trichloroethane	5.367	97	4539	1.128	ug/l #	49
36) 1,1-Dichloropropene	5.678	75	3685	1.082	ug/l	99
37) Ethyl Acetate	4.714	43	5027	1.049	ug/l #	88
38) Carbon Tetrachloride	5.659	117	4298	1.109	ug/l	89
39) Methylcyclohexane	7.360	83	3932	1.029	ug/l #	88
40) Benzene	6.025	78	11724	1.111	ug/l	98
41) Methacrylonitrile	4.910	41	2117	1.005	ug/l #	74
42) 1,2-Dichloroethane	6.068	62	3627	1.003	ug/l	94
43) Isopropyl Acetate	6.330	43	6781	1.007	ug/l	94
44) Trichloroethene	7.110	130	2845	1.128	ug/l	76
45) 1,2-Dichloropropane	7.409	63	2802	1.155	ug/l	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048516.D
 Acq On : 07 Nov 2025 13:35
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:38:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.561	93	1870	1.057	ug/l	97
47) Bromodichloromethane	7.805	83	3471	0.983	ug/l #	73
48) Methyl methacrylate	7.690	41	3138	1.048	ug/l	90
49) 1,4-Dioxane	7.653	88	811	17.773	ug/l #	86
51) 4-Methyl-2-Pentanone	8.555	43	16970	4.426	ug/l	90
52) Toluene	8.702	92	4764	0.857	ug/l	98
53) t-1,3-Dichloropropene	8.964	75	3445	0.947	ug/l #	86
54) cis-1,3-Dichloropropene	8.354	75	3173	0.842	ug/l #	78
55) 1,1,2-Trichloroethane	9.140	97	1939	0.874	ug/l #	87
56) Ethyl methacrylate	9.104	69	2995	0.811	ug/l #	89
57) 1,3-Dichloropropane	9.293	76	3551	0.919	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.226	63	7704m	3.915	ug/l	
59) 2-Hexanone	9.415	43	12104	4.243	ug/l	91
60) Dibromochloromethane	9.500	129	2328	0.854	ug/l	92
61) 1,2-Dibromoethane	9.598	107	2078	0.864	ug/l	87
64) Tetrachloroethene	9.250	164	1983	1.223	ug/l #	83
65) Chlorobenzene	10.067	112	6278	1.184	ug/l	95
66) 1,1,1,2-Tetrachloroethane	10.140	131	1769	0.996	ug/l #	55
67) Ethyl Benzene	10.177	91	10060	1.130	ug/l	96
68) m/p-Xylenes	10.287	106	6820	2.035	ug/l	96
69) o-Xylene	10.622	106	3552	1.083	ug/l	98
70) Styrene	10.640	104	5483	0.997	ug/l	94
71) Bromoform	10.787	173	1905	1.163	ug/l #	89
73) Isopropylbenzene	10.945	105	8562	1.063	ug/l	99
74) N-amyl acetate	10.829	43	4435	1.096	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.195	83	3428	1.207	ug/l	97
76) 1,2,3-Trichloropropane	11.219	75	2820m	1.192	ug/l	
77) Bromobenzene	11.183	156	2339	1.124	ug/l	98
78) n-propylbenzene	11.286	91	10393	1.097	ug/l	99
79) 2-Chlorotoluene	11.347	91	6648	1.161	ug/l	98
80) 1,3,5-Trimethylbenzene	11.433	105	7041	1.072	ug/l	94
82) 4-Chlorotoluene	11.439	91	7628	1.136	ug/l	97
83) tert-Butylbenzene	11.695	119	7172	1.056	ug/l	100
84) 1,2,4-Trimethylbenzene	11.731	105	7156	1.080	ug/l	97
85) sec-Butylbenzene	11.872	105	9183	1.094	ug/l	96
86) p-Isopropyltoluene	11.994	119	7646	1.083	ug/l	92
87) 1,3-Dichlorobenzene	11.951	146	4609	1.182	ug/l	96
88) 1,4-Dichlorobenzene	12.018	146	5170m	1.314	ug/l	
89) n-Butylbenzene	12.317	91	7282	1.100	ug/l	94
90) Hexachloroethane	12.518	117	1507	1.121	ug/l	95
91) 1,2-Dichlorobenzene	12.323	146	4315	1.153	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.932	75	674	1.021	ug/l	96
93) 1,2,4-Trichlorobenzene	13.567	180	3004	1.208	ug/l	96
94) Hexachlorobutadiene	13.707	225	1356	1.304	ug/l	97
95) Naphthalene	13.762	128	8156	1.020	ug/l	98
96) 1,2,3-Trichlorobenzene	13.944	180	2598	1.123	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048517.D
 Acq On : 07 Nov 2025 13:56
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/10/2025
 Supervised By : Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:38:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	193270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	250338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	225597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	118579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	14969	5.558	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	11.120%#		
35) Dibromofluoromethane	5.373	113	12118	6.396	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	12.800%#		
50) Toluene-d8	8.635	98	33563	5.680	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	11.360%#		
62) 4-Bromofluorobenzene	11.061	95	13095	5.946	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	11.900%#		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	7029	4.488	ug/l	98
3) Chloromethane	1.301	50	8841	4.256	ug/l	89
4) Vinyl Chloride	1.380	62	11197	4.761	ug/l	89
5) Bromomethane	1.611	94	9050	6.088	ug/l	100
6) Chloroethane	1.691	64	8089	5.252	ug/l	90
7) Trichlorofluoromethane	1.892	101	19916	5.353	ug/l	97
8) Diethyl Ether	2.136	74	8115	5.434	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.331	101	10746	4.989	ug/l	98
10) Methyl Iodide	2.453	142	16103	5.052	ug/l	94
11) Tert butyl alcohol	2.947	59	12490	24.476	ug/l	99
12) 1,1-Dichloroethene	2.325	96	11295	5.020	ug/l	90
13) Acrolein	2.246	56	2253	30.799	ug/l	94
14) Allyl chloride	2.666	41	21817	5.208	ug/l	95
15) Acrylonitrile	3.056	53	37042	24.865	ug/l	94
16) Acetone	2.374	43	30713	24.093	ug/l	98
17) Carbon Disulfide	2.514	76	33603	5.308	ug/l	99
18) Methyl Acetate	2.697	43	16930	4.833	ug/l	98
19) Methyl tert-butyl Ether	3.105	73	41711	5.250	ug/l	98
20) Methylene Chloride	2.788	84	13313	5.080	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	12560	5.267	ug/l	87
22) Diisopropyl ether	3.751	45	42130	5.238	ug/l	92
23) Vinyl Acetate	3.715	43	183665	26.073	ug/l	95
24) 1,1-Dichloroethane	3.605	63	24095	5.459	ug/l	98
25) 2-Butanone	4.544	43	48544	25.742	ug/l	90
26) 2,2-Dichloropropane	4.471	77	20186	5.365	ug/l	100
27) cis-1,2-Dichloroethene	4.477	96	15219	5.221	ug/l	95
28) Bromochloromethane	4.885	49	12176	5.616	ug/l	96
29) Tetrahydrofuran	5.001	42	34508	25.691	ug/l	98
30) Chloroform	5.080	83	24671	5.415	ug/l	98
31) Cyclohexane	5.464	56	19337	5.347	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	21014	5.285	ug/l	98
36) 1,1-Dichloropropene	5.678	75	16623	6.552	ug/l	97
37) Ethyl Acetate	4.702	43	22548	6.314	ug/l	98
38) Carbon Tetrachloride	5.666	117	19186	6.643	ug/l #	97
39) Methylcyclohexane	7.367	83	15183	5.330	ug/l	95
40) Benzene	6.025	78	51007	6.486	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048517.D
 Acq On : 07 Nov 2025 13:56
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 11/10/2025

Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:38:58 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M

Quant Title : SW846 8260

QLast Update : Sat Nov 08 04:36:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	10171	6.476	ug/l	99
42) 1,2-Dichloroethane	6.080	62	18051	6.699	ug/l	98
43) Isopropyl Acetate	6.324	43	32505	6.475	ug/l	96
44) Trichloroethene	7.117	130	9754	5.189	ug/l	97
45) 1,2-Dichloropropane	7.415	63	9530	5.270	ug/l	97
46) Dibromomethane	7.568	93	7573	5.742	ug/l	96
47) Bromodichloromethane	7.812	83	14970	5.690	ug/l #	97
48) Methyl methacrylate	7.684	41	11928	5.345	ug/l	93
49) 1,4-Dioxane	7.647	88	3476	102.220	ug/l #	93
51) 4-Methyl-2-Pentanone	8.555	43	83493	29.217	ug/l	95
52) Toluene	8.708	92	23897	5.772	ug/l	95
53) t-1,3-Dichloropropene	8.964	75	15253	5.626	ug/l	100
54) cis-1,3-Dichloropropene	8.354	75	15836	5.641	ug/l	92
55) 1,1,2-Trichloroethane	9.135	97	9348	5.651	ug/l	96
56) Ethyl methacrylate	9.104	69	14858	5.398	ug/l	89
57) 1,3-Dichloropropane	9.293	76	16872	5.860	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.226	63	37670m	25.686	ug/l	
59) 2-Hexanone	9.415	43	60936	28.666	ug/l	94
60) Dibromochloromethane	9.506	129	11712	5.766	ug/l	98
61) 1,2-Dibromoethane	9.598	107	10441	5.827	ug/l	100
64) Tetrachloroethene	9.256	164	8201	5.005	ug/l	98
65) Chlorobenzene	10.061	112	28081	5.244	ug/l	95
66) 1,1,1,2-Tetrachloroethane	10.147	131	9473	5.280	ug/l	94
67) Ethyl Benzene	10.177	91	46125	5.130	ug/l	98
68) m/p-Xylenes	10.287	106	35759	10.559	ug/l	100
69) o-Xylene	10.628	106	16824	5.078	ug/l	99
70) Styrene	10.640	104	28334	5.098	ug/l	97
71) Bromoform	10.787	173	7895	4.770	ug/l #	99
73) Isopropylbenzene	10.945	105	44193	5.134	ug/l	99
74) N-amyl acetate	10.829	43	21656	5.011	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.195	83	15227	5.017	ug/l	100
76) 1,2,3-Trichloropropane	11.226	75	13198m	5.220	ug/l	
77) Bromobenzene	11.183	156	11396	5.127	ug/l	99
78) n-propylbenzene	11.287	91	51649	5.103	ug/l	100
79) 2-Chlorotoluene	11.348	91	30839	5.043	ug/l	98
80) 1,3,5-Trimethylbenzene	11.433	105	35455	5.053	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	4463	4.469	ug/l #	80
82) 4-Chlorotoluene	11.439	91	36813	5.132	ug/l	99
83) tert-Butylbenzene	11.695	119	36264	4.997	ug/l	97
84) 1,2,4-Trimethylbenzene	11.732	105	35130	4.962	ug/l	100
85) sec-Butylbenzene	11.872	105	46046	5.133	ug/l	100
86) p-Isopropyltoluene	11.994	119	37206	4.934	ug/l	97
87) 1,3-Dichlorobenzene	11.951	146	21165	5.081	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	22051	5.247	ug/l	95
89) n-Butylbenzene	12.317	91	36025	5.092	ug/l	98
90) Hexachloroethane	12.518	117	7251	5.047	ug/l	97
91) 1,2-Dichlorobenzene	12.317	146	20105	5.028	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.927	75	3647	5.170	ug/l	90
93) 1,2,4-Trichlorobenzene	13.573	180	12241	4.606	ug/l	99
94) Hexachlorobutadiene	13.707	225	5537	4.984	ug/l	99
95) Naphthalene	13.756	128	39122	4.579	ug/l	98
96) 1,2,3-Trichlorobenzene	13.945	180	11538	4.668	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048517.D
Acq On : 07 Nov 2025 13:56
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:38:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048517.D
Acq On : 07 Nov 2025 13:56
Operator : JC/MD
Sample : VSTDIC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:38:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048518.D
 Acq On : 07 Nov 2025 14:16
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: Nov 08 04:39:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 11/10/2025
 Supervised By : Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	203076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	331319	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	232027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	120318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	58757	20.764	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	41.520%#		
35) Dibromofluoromethane	5.373	113	49531	19.752	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	39.500%#		
50) Toluene-d8	8.635	98	135275	17.297	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	34.600%#		
62) 4-Bromofluorobenzene	11.061	95	48386	16.601	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	33.200%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.173	85	28628	17.397	ug/l	98
3) Chloromethane	1.301	50	34522	15.817	ug/l	99
4) Vinyl Chloride	1.380	62	49358	19.974	ug/l	97
5) Bromomethane	1.618	94	33868	21.683	ug/l	98
6) Chloroethane	1.697	64	32257	19.932	ug/l	98
7) Trichlorofluoromethane	1.898	101	82782	21.175	ug/l	95
8) Diethyl Ether	2.136	74	30818	19.639	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.337	101	50073	22.124	ug/l	97
10) Methyl Iodide	2.459	142	66788	19.940	ug/l	100
11) Tert butyl alcohol	2.941	59	53988	100.687	ug/l	98
12) 1,1-Dichloroethene	2.325	96	48003	20.306	ug/l	95
13) Acrolein	2.239	56	6094m	79.282	ug/l	
14) Allyl chloride	2.666	41	86510	19.655	ug/l	97
15) Acrylonitrile	3.056	53	160471	102.519	ug/l	99
16) Acetone	2.374	43	129334	96.557	ug/l	97
17) Carbon Disulfide	2.514	76	130735	19.652	ug/l	99
18) Methyl Acetate	2.697	43	69637	18.918	ug/l	99
19) Methyl tert-butyl Ether	3.105	73	166418	19.935	ug/l	100
20) Methylene Chloride	2.788	84	53829	19.547	ug/l	94
21) trans-1,2-Dichloroethene	3.093	96	49731	19.848	ug/l	98
22) Diisopropyl ether	3.751	45	170973	20.230	ug/l	98
23) Vinyl Acetate	3.715	43	746382	100.838	ug/l	98
24) 1,1-Dichloroethane	3.605	63	91293	19.686	ug/l	98
25) 2-Butanone	4.538	43	207425	104.681	ug/l	98
26) 2,2-Dichloropropane	4.471	77	79199	20.032	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	59896	19.557	ug/l	98
28) Bromochloromethane	4.885	49	43296	19.005	ug/l	99
29) Tetrahydrofuran	4.989	42	142200	100.757	ug/l	98
30) Chloroform	5.087	83	94424	19.723	ug/l	98
31) Cyclohexane	5.464	56	81570	21.467	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	83345	19.948	ug/l	98
36) 1,1-Dichloropropene	5.684	75	66129	19.694	ug/l	97
37) Ethyl Acetate	4.696	43	89614	18.959	ug/l	100
38) Carbon Tetrachloride	5.666	117	72325	18.922	ug/l	98
39) Methylcyclohexane	7.367	83	73943	19.612	ug/l	99
40) Benzene	6.025	78	194589	18.695	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048518.D
 Acq On : 07 Nov 2025 14:16
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:39:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	41068	19.757	ug/l	97
42) 1,2-Dichloroethane	6.074	62	69540	19.499	ug/l	99
43) Isopropyl Acetate	6.324	43	126911	19.101	ug/l	99
44) Trichloroethene	7.117	130	47204	18.973	ug/l	99
45) 1,2-Dichloropropane	7.415	63	42265	17.660	ug/l	98
46) Dibromomethane	7.568	93	28753	16.472	ug/l	96
47) Bromodichloromethane	7.806	83	59084	16.967	ug/l	99
48) Methyl methacrylate	7.678	41	51074	17.292	ug/l	91
49) 1,4-Dioxane	7.641	88	14684	326.274	ug/l #	94
51) 4-Methyl-2-Pentanone	8.555	43	340555	90.045	ug/l	97
52) Toluene	8.702	92	94154	17.182	ug/l	98
53) t-1,3-Dichloropropene	8.964	75	59921	16.700	ug/l	99
54) cis-1,3-Dichloropropene	8.354	75	63747	17.157	ug/l	93
55) 1,1,2-Trichloroethane	9.135	97	37862	17.293	ug/l	96
56) Ethyl methacrylate	9.098	69	62768	17.232	ug/l #	90
57) 1,3-Dichloropropane	9.293	76	65949	17.306	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.226	63	166309m	85.683	ug/l	
59) 2-Hexanone	9.415	43	258663	91.942	ug/l	96
60) Dibromochloromethane	9.506	129	46524	17.307	ug/l	97
61) 1,2-Dibromoethane	9.592	107	40837	17.221	ug/l	99
64) Tetrachloroethene	9.256	164	32666	19.383	ug/l	98
65) Chlorobenzene	10.061	112	106968	19.423	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.147	131	36079	19.552	ug/l	98
67) Ethyl Benzene	10.177	91	183401	19.832	ug/l	97
68) m/p-Xylenes	10.287	106	140701	40.397	ug/l	96
69) o-Xylene	10.628	106	66957	19.649	ug/l	98
70) Styrene	10.640	104	113854	19.917	ug/l	98
71) Bromoform	10.781	173	32046	18.824	ug/l #	98
73) Isopropylbenzene	10.945	105	177238	20.294	ug/l	98
74) N-amyl acetate	10.823	43	89078	20.313	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.195	83	59190	19.218	ug/l	99
76) 1,2,3-Trichloropropane	11.220	75	54573m	21.273	ug/l	
77) Bromobenzene	11.183	156	43931	19.480	ug/l	98
78) n-propylbenzene	11.287	91	210347	20.483	ug/l	98
79) 2-Chlorotoluene	11.348	91	123455	19.895	ug/l	98
80) 1,3,5-Trimethylbenzene	11.433	105	144509	20.297	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	19161	18.908	ug/l	90
82) 4-Chlorotoluene	11.439	91	143646	19.735	ug/l	99
83) tert-Butylbenzene	11.695	119	148245	20.131	ug/l	97
84) 1,2,4-Trimethylbenzene	11.732	105	143930	20.036	ug/l	100
85) sec-Butylbenzene	11.872	105	184842	20.309	ug/l	99
86) p-Isopropyltoluene	11.994	119	155276	20.293	ug/l	98
87) 1,3-Dichlorobenzene	11.951	146	81125	19.193	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	80643	18.910	ug/l	97
89) n-Butylbenzene	12.311	91	144697	20.158	ug/l	98
90) Hexachloroethane	12.518	117	28866	19.800	ug/l	100
91) 1,2-Dichlorobenzene	12.317	146	77511	19.103	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.920	75	14148	19.767	ug/l	95
93) 1,2,4-Trichlorobenzene	13.567	180	50883	18.871	ug/l	99
94) Hexachlorobutadiene	13.707	225	21648	19.206	ug/l	100
95) Naphthalene	13.756	128	169835	19.591	ug/l	100
96) 1,2,3-Trichlorobenzene	13.939	180	47818	19.066	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048518.D
Acq On : 07 Nov 2025 14:16
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:39:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048518.D
Acq On : 07 Nov 2025 14:16
Operator : JC/MD
Sample : VSTDIC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Quant Time: Nov 08 04:39:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048520.D
 Acq On : 07 Nov 2025 14:58
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/10/2025
 Supervised By : Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	179485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	297765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	257038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	132080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	243155	97.224	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 194.440%#			
35) Dibromofluoromethane	5.373	113	212256	94.182	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 188.360%#			
50) Toluene-d8	8.635	98	720420	102.495	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 204.980%#			
62) 4-Bromofluorobenzene	11.061	95	264846	101.109	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 202.220%#			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.173	85	162366	111.639	ug/l	99
3) Chloromethane	1.301	50	208707	108.195	ug/l	99
4) Vinyl Chloride	1.380	62	215712	98.767	ug/l	99
5) Bromomethane	1.605	94	115797m	83.881	ug/l	
6) Chloroethane	1.685	64	132914	92.926	ug/l	100
7) Trichlorofluoromethane	1.892	101	329740	95.432	ug/l	99
8) Diethyl Ether	2.136	74	134965	97.310	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	196084	98.026	ug/l	99
10) Methyl Iodide	2.453	142	300124	101.382	ug/l	99
11) Tert butyl alcohol	2.959	59	237309	500.751	ug/l	98
12) 1,1-Dichloroethene	2.325	96	199260	95.367	ug/l	96
13) Acrolein	2.239	56	31030	456.759	ug/l	97
14) Allyl chloride	2.666	41	362419	93.164	ug/l	99
15) Acrylonitrile	3.056	53	663857	479.859	ug/l	99
16) Acetone	2.374	43	568356	480.092	ug/l	99
17) Carbon Disulfide	2.514	76	527024	89.637	ug/l	100
18) Methyl Acetate	2.697	43	293788	90.303	ug/l	99
19) Methyl tert-butyl Ether	3.105	73	712323	96.543	ug/l	99
20) Methylene Chloride	2.788	84	223836	91.966	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	209909	94.789	ug/l	98
22) Diisopropyl ether	3.751	45	709722	95.012	ug/l	97
23) Vinyl Acetate	3.715	43	3136295	479.414	ug/l	100
24) 1,1-Dichloroethane	3.605	63	388811	94.862	ug/l	99
25) 2-Butanone	4.538	43	843690	481.750	ug/l	99
26) 2,2-Dichloropropane	4.465	77	335236	95.939	ug/l	100
27) cis-1,2-Dichloroethene	4.477	96	252407	93.248	ug/l	98
28) Bromochloromethane	4.885	49	187288	93.016	ug/l	99
29) Tetrahydrofuran	4.989	42	575336	461.240	ug/l	99
30) Chloroform	5.086	83	396642	93.738	ug/l	99
31) Cyclohexane	5.458	56	321505	95.734	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	351320	95.136	ug/l	99
36) 1,1-Dichloropropene	5.678	75	269881	89.431	ug/l	99
37) Ethyl Acetate	4.696	43	386137	90.900	ug/l	99
38) Carbon Tetrachloride	5.666	117	309414	90.070	ug/l	99
39) Methylcyclohexane	7.367	83	334291	98.656	ug/l	96
40) Benzene	6.025	78	831305	88.867	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048520.D
 Acq On : 07 Nov 2025 14:58
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	164274	87.935	ug/l	95
42) 1,2-Dichloroethane	6.068	62	291402	90.918	ug/l	99
43) Isopropyl Acetate	6.324	43	540612	90.535	ug/l	100
44) Trichloroethene	7.110	130	218005	97.499	ug/l	98
45) 1,2-Dichloropropane	7.415	63	209425	97.368	ug/l	100
46) Dibromomethane	7.568	93	156924	100.030	ug/l	98
47) Bromodichloromethane	7.805	83	321173	102.623	ug/l	97
48) Methyl methacrylate	7.677	41	266560	100.420	ug/l	100
49) 1,4-Dioxane	7.647	88	89615	2215.601	ug/l	97
51) 4-Methyl-2-Pentanone	8.561	43	1705402	501.730	ug/l	99
52) Toluene	8.702	92	517017	104.979	ug/l	99
53) t-1,3-Dichloropropene	8.964	75	334066	103.593	ug/l	99
54) cis-1,3-Dichloropropene	8.354	75	354217	106.080	ug/l	98
55) 1,1,2-Trichloroethane	9.134	97	205483	104.430	ug/l	98
56) Ethyl methacrylate	9.104	69	353946	108.118	ug/l	98
57) 1,3-Dichloropropane	9.293	76	345885	100.991	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.226	63	972237	557.345	ug/l	100
59) 2-Hexanone	9.415	43	1278151	505.516	ug/l	100
60) Dibromochloromethane	9.506	129	255107	105.596	ug/l	98
61) 1,2-Dibromoethane	9.592	107	222500	104.402	ug/l	100
64) Tetrachloroethene	9.256	164	177093	94.858	ug/l	96
65) Chlorobenzene	10.061	112	571492	93.674	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.147	131	204743	100.159	ug/l	99
67) Ethyl Benzene	10.177	91	976399	95.307	ug/l	98
68) m/p-Xylenes	10.287	106	756955	196.183	ug/l	99
69) o-Xylene	10.628	106	367131	97.252	ug/l	100
70) Styrene	10.640	104	630372	99.544	ug/l	100
71) Bromoform	10.781	173	184392	97.773	ug/l	99
73) Isopropylbenzene	10.945	105	937990	97.838	ug/l	99
74) N-amyl acetate	10.829	43	459285	95.409	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.195	83	318823	94.300	ug/l	100
76) 1,2,3-Trichloropropane	11.219	75	298071m	105.844	ug/l	
77) Bromobenzene	11.183	156	240181	97.018	ug/l	99
78) n-propylbenzene	11.287	91	1082239	96.002	ug/l	99
79) 2-Chlorotoluene	11.348	91	658434	96.659	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	769561	98.464	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	117309	105.450	ug/l	98
82) 4-Chlorotoluene	11.439	91	770164	96.385	ug/l	100
83) tert-Butylbenzene	11.695	119	801942	99.204	ug/l	99
84) 1,2,4-Trimethylbenzene	11.732	105	779410	98.837	ug/l	99
85) sec-Butylbenzene	11.872	105	962376	96.323	ug/l	100
86) p-Isopropyltoluene	11.994	119	825080	98.227	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	443784	95.645	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	441447	94.299	ug/l	99
89) n-Butylbenzene	12.317	91	761466	96.635	ug/l	99
90) Hexachloroethane	12.518	117	157787	98.594	ug/l	100
91) 1,2-Dichlorobenzene	12.317	146	429792	96.491	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.920	75	78635	100.084	ug/l	98
93) 1,2,4-Trichlorobenzene	13.567	180	289894	97.941	ug/l	99
94) Hexachlorobutadiene	13.707	225	111275	89.932	ug/l	98
95) Naphthalene	13.756	128	985717	103.580	ug/l	100
96) 1,2,3-Trichlorobenzene	13.939	180	277082	100.642	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048520.D
Acq On : 07 Nov 2025 14:58
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048520.D
Acq On : 07 Nov 2025 14:58
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Quant Time: Nov 08 04:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048521.D
 Acq On : 07 Nov 2025 15:18
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Nov 08 04:42:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	178567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	292795	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	246117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131303	50.000	ug/l	# 0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.940	65	333785	134.148	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 268.300%#	
35) Dibromofluoromethane	5.373	113	287214	129.606	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 259.220%#	
50) Toluene-d8	8.635	98	990842	143.361	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 286.720%#	
62) 4-Bromofluorobenzene	11.067	95	371661	144.295	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 288.600%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.173	85	253735	175.359	ug/l	99
3) Chloromethane	1.301	50	302480	157.614	ug/l	99
4) Vinyl Chloride	1.380	62	320003	147.272	ug/l	99
5) Bromomethane	1.605	94	153950	112.091	ug/l	96
6) Chloroethane	1.685	64	191395	134.501	ug/l	99
7) Trichlorofluoromethane	1.892	101	503070	146.345	ug/l	100
8) Diethyl Ether	2.136	74	188944	136.929	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	298670	150.078	ug/l	98
10) Methyl Iodide	2.453	142	432790	146.948	ug/l	99
11) Tert butyl alcohol	2.977	59	340733	722.684	ug/l	98
12) 1,1-Dichloroethene	2.319	96	293656	141.269	ug/l	97
13) Acrolein	2.239	56	52843	781.843	ug/l	96
14) Allyl chloride	2.666	41	540073	139.545	ug/l	100
15) Acrylonitrile	3.056	53	953673	692.892	ug/l	99
16) Acetone	2.373	43	838719	712.110	ug/l	100
17) Carbon Disulfide	2.514	76	791417	135.297	ug/l	99
18) Methyl Acetate	2.697	43	440781	136.181	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	997177	135.845	ug/l	99
20) Methylene Chloride	2.788	84	318695	131.614	ug/l	98
21) trans-1,2-Dichloroethene	3.093	96	298171	135.338	ug/l	98
22) Diisopropyl ether	3.751	45	1038977	139.806	ug/l	97
23) Vinyl Acetate	3.715	43	4581534	703.934	ug/l	99
24) 1,1-Dichloroethane	3.605	63	557501	136.718	ug/l	99
25) 2-Butanone	4.544	43	1244508	714.272	ug/l	97
26) 2,2-Dichloropropane	4.471	77	495685	142.586	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	366206	135.984	ug/l	99
28) Bromochloromethane	4.891	49	266434	133.003	ug/l	100
29) Tetrahydrofuran	4.989	42	818751	659.757	ug/l	98
30) Chloroform	5.080	83	563304	133.810	ug/l	98
31) Cyclohexane	5.458	56	487596	145.937	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	507525	138.143	ug/l	99
36) 1,1-Dichloropropene	5.678	75	394330	132.888	ug/l	100
37) Ethyl Acetate	4.696	43	561632	134.458	ug/l	99
38) Carbon Tetrachloride	5.666	117	442446	130.982	ug/l	97
39) Methylcyclohexane	7.367	83	520380	156.182	ug/l	96
40) Benzene	6.025	78	1202907	130.774	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048521.D
 Acq On : 07 Nov 2025 15:18
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:42:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	245609	133.706	ug/l	98
42) 1,2-Dichloroethane	6.074	62	415612	131.873	ug/l	99
43) Isopropyl Acetate	6.324	43	802693	136.707	ug/l	100
44) Trichloroethene	7.110	130	318538	144.879	ug/l	97
45) 1,2-Dichloropropane	7.415	63	298821	141.289	ug/l	98
46) Dibromomethane	7.568	93	220703	143.074	ug/l	99
47) Bromodichloromethane	7.805	83	447995	145.576	ug/l	100
48) Methyl methacrylate	7.677	41	382424	146.515	ug/l	99
49) 1,4-Dioxane	7.653	88	128773	3237.768	ug/l	98
51) 4-Methyl-2-Pentanone	8.561	43	2398543	717.630	ug/l	100
52) Toluene	8.702	92	740196	152.846	ug/l	99
53) t-1,3-Dichloropropene	8.964	75	475896	150.079	ug/l	99
54) cis-1,3-Dichloropropene	8.354	75	507432	154.544	ug/l	99
55) 1,1,2-Trichloroethane	9.134	97	288895	149.313	ug/l	97
56) Ethyl methacrylate	9.104	69	503008	156.260	ug/l	99
57) 1,3-Dichloropropane	9.293	76	484188	143.772	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.232	63	1366958	796.924	ug/l	100
59) 2-Hexanone	9.415	43	1828924	735.629	ug/l	100
60) Dibromochloromethane	9.506	129	354928	149.408	ug/l	99
61) 1,2-Dibromoethane	9.598	107	314737	150.188	ug/l	100
64) Tetrachloroethene	9.256	164	256641	143.566	ug/l	99
65) Chlorobenzene	10.061	112	815516	139.603	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.146	131	288555	147.422	ug/l	100
67) Ethyl Benzene	10.177	91	1401386	142.860	ug/l	98
68) m/p-Xylenes	10.287	106	1082548	293.018	ug/l	99
69) o-Xylene	10.628	106	536318	148.373	ug/l	100
70) Styrene	10.640	104	918795	151.528	ug/l	100
71) Bromoform	10.781	173	275305	152.457	ug/l	99
73) Isopropylbenzene	10.945	105	1375604	144.333	ug/l	99
74) N-amyl acetate	10.829	43	680374	142.173	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.195	83	466659	138.843	ug/l	100
76) 1,2,3-Trichloropropane	11.219	75	437992m	156.450	ug/l	
77) Bromobenzene	11.183	156	345419	140.354	ug/l	98
78) n-propylbenzene	11.287	91	1604351	143.159	ug/l	100
79) 2-Chlorotoluene	11.348	91	938129	138.534	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	1124584	144.740	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	178669	161.557	ug/l	99
82) 4-Chlorotoluene	11.439	91	1119611	140.947	ug/l	99
83) tert-Butylbenzene	11.695	119	1185835	147.561	ug/l	99
84) 1,2,4-Trimethylbenzene	11.732	105	1136998	145.036	ug/l	98
85) sec-Butylbenzene	11.872	105	1454720	146.463	ug/l	100
86) p-Isopropyltoluene	11.994	119	1244220	149.002	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	657153	142.468	ug/l	100
88) 1,4-Dichlorobenzene	12.024	146	658598	141.517	ug/l	100
89) n-Butylbenzene	12.317	91	1170811	149.462	ug/l	99
90) Hexachloroethane	12.518	117	234767	147.563	ug/l	97
91) 1,2-Dichlorobenzene	12.317	146	636660	143.779	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.920	75	112194	143.641	ug/l	95
93) 1,2,4-Trichlorobenzene	13.567	180	439995	149.532	ug/l	99
94) Hexachlorobutadiene	13.707	225	179249	145.725	ug/l	98
95) Naphthalene	13.756	128	1450179	153.288	ug/l	100
96) 1,2,3-Trichlorobenzene	13.938	180	414012	151.267	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048521.D
Acq On : 07 Nov 2025 15:18
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:42:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048521.D
Acq On : 07 Nov 2025 15:18
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:42:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048524.D
 Acq On : 07 Nov 2025 16:29
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Nov 08 04:43:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	180041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	307413	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	271611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	137081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	123374	49.178	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.360%		
35) Dibromofluoromethane	5.373	113	107888	46.370	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.740%		
50) Toluene-d8	8.635	98	369557	50.927	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.860%		
62) 4-Bromofluorobenzene	11.061	95	136244	50.381	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.760%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	76351	52.335	ug/l	100
3) Chloromethane	1.301	50	106667	55.126	ug/l	100
4) Vinyl Chloride	1.380	62	107960	49.279	ug/l	100
5) Bromomethane	1.612	94	76656	55.356	ug/l	100
6) Chloroethane	1.691	64	68875	48.005	ug/l	100
7) Trichlorofluoromethane	1.892	101	159505	46.021	ug/l	100
8) Diethyl Ether	2.136	74	70765	50.864	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	92738	46.218	ug/l	100
10) Methyl Iodide	2.453	142	148356	49.960	ug/l	100
11) Tert butyl alcohol	2.941	59	124670	262.256	ug/l	100
12) 1,1-Dichloroethene	2.325	96	98074	46.794	ug/l	100
13) Acrolein	2.240	56	18158	266.459	ug/l	100
14) Allyl chloride	2.666	41	188637	48.341	ug/l	100
15) Acrylonitrile	3.056	53	363119	261.664	ug/l	100
16) Acetone	2.374	43	296864	249.987	ug/l	100
17) Carbon Disulfide	2.514	76	268053	45.450	ug/l	100
18) Methyl Acetate	2.697	43	156429	47.934	ug/l	100
19) Methyl tert-butyl Ether	3.105	73	381141	51.497	ug/l	100
20) Methylene Chloride	2.788	84	117536	48.142	ug/l	100
21) trans-1,2-Dichloroethene	3.093	96	107632	48.453	ug/l	100
22) Diisopropyl ether	3.751	45	371667	49.602	ug/l	100
23) Vinyl Acetate	3.715	43	1644477	250.599	ug/l	100
24) 1,1-Dichloroethane	3.605	63	198326	48.238	ug/l	100
25) 2-Butanone	4.538	43	455146	259.087	ug/l	100
26) 2,2-Dichloropropane	4.465	77	159000	45.363	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	130445	48.042	ug/l	100
28) Bromochloromethane	4.885	49	97819	48.431	ug/l	100
29) Tetrahydrofuran	4.989	42	313988	250.943	ug/l	100
30) Chloroform	5.087	83	201702	47.521	ug/l	100
31) Cyclohexane	5.458	56	156130	46.347	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	175119	47.275	ug/l	100
36) 1,1-Dichloropropene	5.678	75	131207	42.114	ug/l	100
37) Ethyl Acetate	4.696	43	205009	46.746	ug/l	100
38) Carbon Tetrachloride	5.666	117	149370	42.117	ug/l	100
39) Methylcyclohexane	7.367	83	156904	44.852	ug/l	100
40) Benzene	6.025	78	432976	44.833	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048524.D
 Acq On : 07 Nov 2025 16:29
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Nov 08 04:43:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	90811	47.085	ug/l	100
42) 1,2-Dichloroethane	6.074	62	147877	44.690	ug/l	100
43) Isopropyl Acetate	6.324	43	285622	46.331	ug/l	100
44) Trichloroethene	7.111	130	109058	47.244	ug/l	100
45) 1,2-Dichloropropane	7.415	63	110204	49.629	ug/l	100
46) Dibromomethane	7.568	93	82377	50.863	ug/l	100
47) Bromodichloromethane	7.806	83	167028	51.695	ug/l	100
48) Methyl methacrylate	7.678	41	142167	51.877	ug/l	100
49) 1,4-Dioxane	7.647	88	45365	1086.384	ug/l	100
51) 4-Methyl-2-Pentanone	8.555	43	952270	271.365	ug/l	100
52) Toluene	8.702	92	269581	53.020	ug/l	100
53) t-1,3-Dichloropropene	8.964	75	175854	52.820	ug/l	100
54) cis-1,3-Dichloropropene	8.354	75	186249	54.027	ug/l	100
55) 1,1,2-Trichloroethane	9.135	97	110910	54.597	ug/l	100
56) Ethyl methacrylate	9.098	69	190086	56.242	ug/l	100
57) 1,3-Dichloropropane	9.293	76	190119	53.768	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.226	63	544454	302.318	ug/l	100
59) 2-Hexanone	9.415	43	713506	273.339	ug/l	100
60) Dibromochloromethane	9.507	129	134082	53.758	ug/l	100
61) 1,2-Dibromoethane	9.592	107	117046	53.197	ug/l	100
64) Tetrachloroethene	9.257	164	88926	45.076	ug/l	100
65) Chlorobenzene	10.061	112	299156	46.404	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.147	131	106481	49.295	ug/l	100
67) Ethyl Benzene	10.177	91	512428	47.335	ug/l	100
68) m/p-Xylenes	10.287	106	391085	95.921	ug/l	100
69) o-Xylene	10.622	106	190935	47.864	ug/l	100
70) Styrene	10.634	104	328665	49.116	ug/l	100
71) Bromoform	10.781	173	94463	47.401	ug/l	100
73) Isopropylbenzene	10.945	105	475097	47.748	ug/l	100
74) N-amyl acetate	10.823	43	245758	49.190	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.195	83	168512	48.023	ug/l	100
76) 1,2,3-Trichloropropane	11.220	75	138398m	47.352	ug/l	
77) Bromobenzene	11.177	156	124641	48.510	ug/l	100
78) n-propylbenzene	11.287	91	551985	47.178	ug/l	100
79) 2-Chlorotoluene	11.348	91	334083	47.255	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	386446	47.641	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.000	75	59424	51.468	ug/l	100
82) 4-Chlorotoluene	11.433	91	392809	47.366	ug/l	100
83) tert-Butylbenzene	11.695	119	403731	48.121	ug/l	100
84) 1,2,4-Trimethylbenzene	11.732	105	397157	48.526	ug/l	100
85) sec-Butylbenzene	11.872	105	479253	46.218	ug/l	100
86) p-Isopropyltoluene	11.988	119	409592	46.983	ug/l	100
87) 1,3-Dichlorobenzene	11.951	146	225309	46.787	ug/l	100
88) 1,4-Dichlorobenzene	12.024	146	230056	47.350	ug/l	100
89) n-Butylbenzene	12.311	91	372536	45.552	ug/l	100
90) Hexachloroethane	12.518	117	75599	45.515	ug/l	100
91) 1,2-Dichlorobenzene	12.317	146	222592	48.150	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.921	75	40699	49.910	ug/l	100
93) 1,2,4-Trichlorobenzene	13.567	180	146070	47.549	ug/l	100
94) Hexachlorobutadiene	13.707	225	55716	43.386	ug/l	100
95) Naphthalene	13.756	128	507140	51.347	ug/l	100
96) 1,2,3-Trichlorobenzene	13.939	180	139324	48.759	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048524.D
Acq On : 07 Nov 2025 16:29
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:43:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048525.D
 Acq On : 07 Nov 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX110725

Quant Time: Nov 08 05:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:01:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 11/10/2025
 Supervised By : Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	175429	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	300694	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	261891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	133606	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	125725	51.433	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.860%			
35) Dibromofluoromethane	5.373	113	104074	45.730	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.460%			
50) Toluene-d8	8.635	98	360437	50.780	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.560%			
62) 4-Bromofluorobenzene	11.061	95	134490	50.843	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.680%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.173	85	79523	55.942	ug/l	99
3) Chloromethane	1.301	50	100103	53.094	ug/l	100
4) Vinyl Chloride	1.380	62	105708	49.519	ug/l	96
5) Bromomethane	1.611	94	74133	54.997	ug/l	97
6) Chloroethane	1.691	64	66991	47.919	ug/l	100
7) Trichlorofluoromethane	1.892	101	163901	48.532	ug/l	98
8) Diethyl Ether	2.136	74	68302	50.384	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.337	101	97010	49.618	ug/l	99
10) Methyl Iodide	2.453	142	148327	51.263	ug/l	99
11) Tert butyl alcohol	2.941	59	119049	257.016	ug/l	98
12) 1,1-Dichloroethene	2.325	96	96593	47.299	ug/l	95
13) Acrolein	2.239	56	17055	254.482	ug/l	98
14) Allyl chloride	2.666	41	180497	47.471	ug/l	99
15) Acrylonitrile	3.056	53	339497	251.074	ug/l	99
16) Acetone	2.374	43	289425	250.130	ug/l	98
17) Carbon Disulfide	2.514	76	263284	45.815	ug/l	100
18) Methyl Acetate	2.697	43	151597	47.674	ug/l	99
19) Methyl tert-butyl Ether	3.105	73	366166	50.775	ug/l	100
20) Methylene Chloride	2.788	84	112574	47.322	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	104272	48.175	ug/l	97
22) Diisopropyl ether	3.751	45	365126	50.011	ug/l	98
23) Vinyl Acetate	3.715	43	1623158	253.844	ug/l	99
24) 1,1-Dichloroethane	3.605	63	193671	48.344	ug/l	99
25) 2-Butanone	4.538	43	453513	264.945	ug/l	97
26) 2,2-Dichloropropane	4.465	77	162290	47.519	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	129253	48.854	ug/l	98
28) Bromochloromethane	4.885	49	97361	49.472	ug/l	98
29) Tetrahydrofuran	4.989	42	307563	251.928	ug/l	99
30) Chloroform	5.080	83	203402	49.181	ug/l	99
31) Cyclohexane	5.464	56	166346	50.678	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	177144	49.079	ug/l	100
36) 1,1-Dichloropropene	5.678	75	136129	44.670	ug/l	98
37) Ethyl Acetate	4.696	43	206366	48.107	ug/l	98
38) Carbon Tetrachloride	5.666	117	153372	44.212	ug/l	97
39) Methylcyclohexane	7.367	83	168487	49.240	ug/l	98
40) Benzene	6.025	78	428574	45.369	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048525.D
 Acq On : 07 Nov 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX110725

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/10/2025
 Supervised By : Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 05:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	96023	50.900	ug/l	95
42) 1,2-Dichloroethane	6.074	62	149489	46.187	ug/l	99
43) Isopropyl Acetate	6.324	43	292708	48.542	ug/l	99
44) Trichloroethene	7.111	130	108399	48.007	ug/l	95
45) 1,2-Dichloropropane	7.415	63	109605	50.462	ug/l	97
46) Dibromomethane	7.568	93	81296	51.317	ug/l	97
47) Bromodichloromethane	7.806	83	164733	52.124	ug/l	99
48) Methyl methacrylate	7.677	41	141804	52.901	ug/l	99
49) 1,4-Dioxane	7.647	88	43542	1066.027	ug/l	97
51) 4-Methyl-2-Pentanone	8.555	43	943143	274.770	ug/l	99
52) Toluene	8.702	92	263658	53.014	ug/l	99
53) t-1,3-Dichloropropene	8.964	75	172732	53.042	ug/l	99
54) cis-1,3-Dichloropropene	8.354	75	183768	54.498	ug/l	99
55) 1,1,2-Trichloroethane	9.135	97	109093	54.903	ug/l	98
56) Ethyl methacrylate	9.098	69	186912	56.539	ug/l	98
57) 1,3-Dichloropropane	9.293	76	182103	52.652	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.226	63	533039	299.907	ug/l	100
59) 2-Hexanone	9.415	43	708416	277.453	ug/l	99
60) Dibromochloromethane	9.506	129	130424	53.460	ug/l	100
61) 1,2-Dibromoethane	9.592	107	115352	53.598	ug/l	99
64) Tetrachloroethene	9.256	164	91541	48.124	ug/l	96
65) Chlorobenzene	10.061	112	292507	47.057	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.147	131	103725	49.801	ug/l	99
67) Ethyl Benzene	10.177	91	505794	48.456	ug/l	97
68) m/p-Xylenes	10.287	106	391090	99.482	ug/l	98
69) o-Xylene	10.622	106	190122	49.430	ug/l	98
70) Styrene	10.640	104	318495	49.362	ug/l	99
71) Bromoform	10.781	173	92512	48.145	ug/l #	100
73) Isopropylbenzene	10.945	105	479112	49.403	ug/l	99
74) N-amyl acetate	10.823	43	244571	50.225	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.195	83	167085	48.855	ug/l	99
76) 1,2,3-Trichloropropane	11.220	75	133852m	41.585	ug/l	
77) Bromobenzene	11.183	156	120400	48.079	ug/l	98
78) n-propylbenzene	11.287	91	560609	49.162	ug/l	100
79) 2-Chlorotoluene	11.348	91	336034	48.767	ug/l	99
80) 1,3,5-Trimethylbenzene	11.433	105	396355	50.134	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	57267	50.890	ug/l	96
82) 4-Chlorotoluene	11.433	91	396344	49.035	ug/l	100
83) tert-Butylbenzene	11.695	119	406392	49.698	ug/l	98
84) 1,2,4-Trimethylbenzene	11.732	105	399935	50.137	ug/l	99
85) sec-Butylbenzene	11.872	105	498225	49.297	ug/l	99
86) p-Isopropyltoluene	11.988	119	419888	49.417	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	223034	47.519	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	225606	46.536	ug/l	99
89) n-Butylbenzene	12.311	91	390447	48.984	ug/l	100
90) Hexachloroethane	12.518	117	78393	48.425	ug/l	100
91) 1,2-Dichlorobenzene	12.317	146	216201	47.984	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.920	75	40188	50.566	ug/l	99
93) 1,2,4-Trichlorobenzene	13.567	180	148227	49.507	ug/l	99
94) Hexachlorobutadiene	13.707	225	58865	47.031	ug/l	99
95) Naphthalene	13.756	128	506116	52.576	ug/l	100
96) 1,2,3-Trichlorobenzene	13.939	180	140560	50.471	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048525.D
Acq On : 07 Nov 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX110725

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 05:04:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:01:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048525.D
Acq On : 07 Nov 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX110725

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 05:04:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:01:56 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048525.D
 Acq On : 07 Nov 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX110725

Quant Time: Nov 08 05:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:01:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.405	0.453	-11.9	104	0.00
3 P	Chloromethane	0.537	0.571	-6.3	94	0.00
4 C	Vinyl Chloride	0.608	0.603	0.8#	98	0.00
5 T	Bromomethane	0.384	0.423	-10.2	97	0.00
6 T	Chloroethane	0.398	0.382	4.0	97	0.00
7 T	Trichlorofluoromethane	0.963	0.934	3.0	103	0.00
8 T	Diethyl Ether	0.386	0.389	-0.8	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.553	0.7	105	0.00
10 T	Methyl Iodide	0.825	0.846	-2.5	100	0.00
11 T	Tert butyl alcohol	0.132	0.136	-3.0	95	0.00
12 CM	1,1-Dichloroethene	0.582	0.551	5.3#	98	0.00
13 T	Acrolein	0.019	0.019	0.0	94	0.00
14 T	Allyl chloride	1.084	1.029	5.1	96	0.00
15 T	Acrylonitrile	0.385	0.387	-0.5	93	0.00
16 T	Acetone	0.330	0.330	0.0	97	0.00
17 T	Carbon Disulfide	1.638	1.501	8.4	98	0.00
18 T	Methyl Acetate	0.906	0.864	4.6	97	0.00
19 T	Methyl tert-butyl Ether	2.055	2.087	-1.6	96	0.00
20 T	Methylene Chloride	0.678	0.642	5.3	96	0.00
21 T	trans-1,2-Dichloroethene	0.617	0.594	3.7	97	0.00
22 T	Diisopropyl ether	2.081	2.081	0.0	98	0.00
23 T	Vinyl Acetate	1.822	1.851	-1.6	99	0.00
24 P	1,1-Dichloroethane	1.142	1.104	3.3	98	0.00
25 T	2-Butanone	0.488	0.517	-5.9	100	0.00
26 T	2,2-Dichloropropane	0.973	0.925	4.9	102	0.00
27 T	cis-1,2-Dichloroethene	0.754	0.737	2.3	99	0.00
28 T	Bromochloromethane	0.561	0.555	1.1	100	0.00
29 T	Tetrahydrofuran	0.348	0.351	-0.9	98	0.00
30 C	Chloroform	1.179	1.159	1.7#	101	0.00
31 T	Cyclohexane	0.936	0.948	-1.3	107	0.00
32 T	1,1,1-Trichloroethane	1.029	1.010	1.8	101	0.00
33 S	1,2-Dichloroethane-d4	0.697	0.717	-2.9	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.378	0.346	8.5	96	0.00
36 T	1,1-Dichloropropene	0.507	0.453	10.7	104	0.00
37 T	Ethyl Acetate	0.713	0.686	3.8	101	0.00
38 T	Carbon Tetrachloride	0.577	0.510	11.6	103	0.00
39 T	Methylcyclohexane	0.569	0.560	1.6	107	0.00
40 TM	Benzene	1.571	1.425	9.3	99	0.00
41 T	Methacrylonitrile	0.314	0.319	-1.6	106	0.00
42 TM	1,2-Dichloroethane	0.538	0.497	7.6	101	0.00
43 T	Isopropyl Acetate	1.003	0.973	3.0	102	0.00
44 TM	Trichloroethene	0.375	0.360	4.0	99	0.00
45 C	1,2-Dichloropropane	0.361	0.365	-1.1#	99	0.00
46 T	Dibromomethane	0.263	0.270	-2.7	99	0.00
47 T	Bromodichloromethane	0.526	0.548	-4.2	99	0.00
48 T	Methyl methacrylate	0.446	0.472	-5.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048525.D
 Acq On : 07 Nov 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX110725

Quant Time: Nov 08 05:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:01:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	96	0.00
50 S	Toluene-d8	1.180	1.199	-1.6	98	0.00
51 T	4-Methyl-2-Pentanone	0.571	0.627	-9.8	99	0.00
52 CM	Toluene	0.827	0.877	-6.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.541	0.574	-6.1	98	0.00
54 T	cis-1,3-Dichloropropene	0.561	0.611	-8.9	99	0.00
55 T	1,1,2-Trichloroethane	0.330	0.363	-10.0	98	0.00
56 T	Ethyl methacrylate	0.550	0.622	-13.1	98	0.00
57 T	1,3-Dichloropropane	0.575	0.606	-5.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.296	0.355	-19.9	98	0.00
59 T	2-Hexanone	0.425	0.471	-10.8	99	0.00
60 T	Dibromochloromethane	0.406	0.434	-6.9	97	0.00
61 T	1,2-Dibromoethane	0.358	0.384	-7.3	99	0.00
62 S	4-Bromofluorobenzene	0.440	0.447	-1.6	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.363	0.350	3.6	103	0.00
65 PM	Chlorobenzene	1.187	1.117	5.9	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.396	0.5	97	0.00
67 C	Ethyl Benzene	1.993	1.931	3.1#	99	0.00
68 T	m/p-Xylenes	0.751	0.747	0.5	100	0.00
69 T	o-Xylene	0.734	0.726	1.1	100	0.00
70 T	Styrene	1.232	1.216	1.3	97	0.00
71 P	Bromoform	0.367	0.353	3.8	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.629	3.586	1.2	101	0.00
74 T	N-amyl acetate	1.822	1.831	-0.5	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.280	1.251	2.3	99	0.00
76 T	1,2,3-Trichloropropane	1.205	1.002	16.8	66	0.00
77 T	Bromobenzene	0.937	0.901	3.8	97	0.00
78 T	n-propylbenzene	4.268	4.196	1.7	102	0.00
79 T	2-Chlorotoluene	2.579	2.515	2.5	101	0.00
80 T	1,3,5-Trimethylbenzene	2.959	2.967	-0.3	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.421	0.429	-1.9	96	0.00
82 T	4-Chlorotoluene	3.025	2.967	1.9	101	0.00
83 T	tert-Butylbenzene	3.060	3.042	0.6	101	0.00
84 T	1,2,4-Trimethylbenzene	2.985	2.993	-0.3	101	0.00
85 T	sec-Butylbenzene	3.782	3.729	1.4	104	0.00
86 T	p-Isopropyltoluene	3.180	3.143	1.2	103	0.00
87 T	1,3-Dichlorobenzene	1.756	1.669	5.0	99	0.00
88 T	1,4-Dichlorobenzene	1.814	1.689	6.9	98	0.00
89 T	n-Butylbenzene	2.983	2.922	2.0	105	0.00
90 T	Hexachloroethane	0.606	0.587	3.1	104	0.00
91 T	1,2-Dichlorobenzene	1.686	1.618	4.0	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.297	0.301	-1.3	99	0.00
93 T	1,2,4-Trichlorobenzene	1.120	1.109	1.0	101	0.00
94 T	Hexachlorobutadiene	0.468	0.441	5.8	106	0.00
95 T	Naphthalene	3.603	3.788	-5.1	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048525.D
Acq On : 07 Nov 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX110725

Quant Time: Nov 08 05:04:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:01:56 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.042	1.052	-1.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048525.D
 Acq On : 07 Nov 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX110725

Quant Time: Nov 08 05:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:01:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	55.942	-11.9	104	0.00
3 P	Chloromethane	50.000	53.094	-6.2	94	0.00
4 C	Vinyl Chloride	50.000	49.519	1.0#	98	0.00
5 T	Bromomethane	50.000	54.997	-10.0	97	0.00
6 T	Chloroethane	50.000	47.919	4.2	97	0.00
7 T	Trichlorofluoromethane	50.000	48.532	2.9	103	0.00
8 T	Diethyl Ether	50.000	50.384	-0.8	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.618	0.8	105	0.00
10 T	Methyl Iodide	50.000	51.263	-2.5	100	0.00
11 T	Tert butyl alcohol	250.000	257.016	-2.8	95	0.00
12 CM	1,1-Dichloroethene	50.000	47.299	5.4#	98	0.00
13 T	Acrolein	250.000	254.482	-1.8	94	0.00
14 T	Allyl chloride	50.000	47.471	5.1	96	0.00
15 T	Acrylonitrile	250.000	251.074	-0.4	93	0.00
16 T	Acetone	250.000	250.130	-0.1	97	0.00
17 T	Carbon Disulfide	50.000	45.815	8.4	98	0.00
18 T	Methyl Acetate	50.000	47.674	4.7	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.775	-1.5	96	0.00
20 T	Methylene Chloride	50.000	47.322	5.4	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.175	3.7	97	0.00
22 T	Diisopropyl ether	50.000	50.011	-0.0	98	0.00
23 T	Vinyl Acetate	250.000	253.844	-1.5	99	0.00
24 P	1,1-Dichloroethane	50.000	48.344	3.3	98	0.00
25 T	2-Butanone	250.000	264.945	-6.0	100	0.00
26 T	2,2-Dichloropropane	50.000	47.519	5.0	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.854	2.3	99	0.00
28 T	Bromochloromethane	50.000	49.472	1.1	100	0.00
29 T	Tetrahydrofuran	250.000	251.928	-0.8	98	0.00
30 C	Chloroform	50.000	49.181	1.6#	101	0.00
31 T	Cyclohexane	50.000	50.678	-1.4	107	0.00
32 T	1,1,1-Trichloroethane	50.000	49.079	1.8	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.433	-2.9	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	45.730	8.5	96	0.00
36 T	1,1-Dichloropropene	50.000	44.670	10.7	104	0.00
37 T	Ethyl Acetate	50.000	48.107	3.8	101	0.00
38 T	Carbon Tetrachloride	50.000	44.212	11.6	103	0.00
39 T	Methylcyclohexane	50.000	49.240	1.5	107	0.00
40 TM	Benzene	50.000	45.369	9.3	99	0.00
41 T	Methacrylonitrile	50.000	50.900	-1.8	106	0.00
42 TM	1,2-Dichloroethane	50.000	46.187	7.6	101	0.00
43 T	Isopropyl Acetate	50.000	48.542	2.9	102	0.00
44 TM	Trichloroethene	50.000	48.007	4.0	99	0.00
45 C	1,2-Dichloropropane	50.000	50.462	-0.9#	99	0.00
46 T	Dibromomethane	50.000	51.317	-2.6	99	0.00
47 T	Bromodichloromethane	50.000	52.124	-4.2	99	0.00
48 T	Methyl methacrylate	50.000	52.901	-5.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048525.D
 Acq On : 07 Nov 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX110725

Quant Time: Nov 08 05:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:01:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1066.027	-6.6	96	0.00
50 S	Toluene-d8	50.000	50.780	-1.6	98	0.00
51 T	4-Methyl-2-Pentanone	250.000	274.770	-9.9	99	0.00
52 CM	Toluene	50.000	53.014	-6.0#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	53.042	-6.1	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.498	-9.0	99	0.00
55 T	1,1,2-Trichloroethane	50.000	54.903	-9.8	98	0.00
56 T	Ethyl methacrylate	50.000	56.539	-13.1	98	0.00
57 T	1,3-Dichloropropane	50.000	52.652	-5.3	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	299.907	-20.0	98	0.00
59 T	2-Hexanone	250.000	277.453	-11.0	99	0.00
60 T	Dibromochloromethane	50.000	53.460	-6.9	97	0.00
61 T	1,2-Dibromoethane	50.000	53.598	-7.2	99	0.00
62 S	4-Bromofluorobenzene	50.000	50.843	-1.7	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	48.124	3.8	103	0.00
65 PM	Chlorobenzene	50.000	47.057	5.9	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.801	0.4	97	0.00
67 C	Ethyl Benzene	50.000	48.456	3.1#	99	0.00
68 T	m/p-Xylenes	100.000	99.482	0.5	100	0.00
69 T	o-Xylene	50.000	49.430	1.1	100	0.00
70 T	Styrene	50.000	49.362	1.3	97	0.00
71 P	Bromoform	50.000	48.145	3.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	97	0.00
73 T	Isopropylbenzene	50.000	49.403	1.2	101	0.00
74 T	N-amyl acetate	50.000	50.225	-0.5	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.855	2.3	99	0.00
76 T	1,2,3-Trichloropropane	50.000	41.585	16.8	66	0.00
77 T	Bromobenzene	50.000	48.079	3.8	97	0.00
78 T	n-propylbenzene	50.000	49.162	1.7	102	0.00
79 T	2-Chlorotoluene	50.000	48.767	2.5	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.134	-0.3	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.890	-1.8	96	0.00
82 T	4-Chlorotoluene	50.000	49.035	1.9	101	0.00
83 T	tert-Butylbenzene	50.000	49.698	0.6	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.137	-0.3	101	0.00
85 T	sec-Butylbenzene	50.000	49.297	1.4	104	0.00
86 T	p-Isopropyltoluene	50.000	49.417	1.2	103	0.00
87 T	1,3-Dichlorobenzene	50.000	47.519	5.0	99	0.00
88 T	1,4-Dichlorobenzene	50.000	46.536	6.9	98	0.00
89 T	n-Butylbenzene	50.000	48.984	2.0	105	0.00
90 T	Hexachloroethane	50.000	48.425	3.2	104	0.00
91 T	1,2-Dichlorobenzene	50.000	47.984	4.0	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.566	-1.1	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.507	1.0	101	0.00
94 T	Hexachlorobutadiene	50.000	47.031	5.9	106	0.00
95 T	Naphthalene	50.000	52.576	-5.2	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048525.D
Acq On : 07 Nov 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX110725

Quant Time: Nov 08 05:04:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:01:56 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.471	-0.9	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3702</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/21/2025 09:08</u>
Lab File ID:	<u>VX048641.D</u>	Init. Calib. Date(s):	<u>11/07/2025 11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35 16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.405	0.421		3.95	20
Chloromethane	0.537	0.452	0.1	-15.83	20
Vinyl Chloride	0.608	0.583		-4.11	20
Ethyl Acetate	0.713	0.609		-14.59	20
Bromomethane	0.384	0.352		-8.33	20
Chloroethane	0.398	0.339		-14.82	20
Trichlorofluoromethane	0.963	0.904		-6.13	20
1,1,2-Trichlorotrifluoroethane	0.557	0.592		6.28	20
Tert butyl alcohol	0.132	0.138		4.55	20
1,1-Dichloroethene	0.582	0.538		-7.56	20
Acrolein	0.019	0.017		-10.53	20
Acrylonitrile	0.385	0.400		3.9	20
Acetone	0.330	0.276		-16.36	20
Carbon Disulfide	1.638	1.351		-17.52	20
Methyl tert-butyl Ether	2.055	2.103		2.34	20
Methyl Acetate	0.906	0.885		-2.32	20
Methylene Chloride	0.678	0.647		-4.57	20
trans-1,2-Dichloroethene	0.617	0.585		-5.19	20
Vinyl Acetate	1.822	1.853		1.7	20
1,1-Dichloroethane	1.142	1.105	0.1	-3.24	20
Cyclohexane	0.936	0.861		-8.01	20
2-Butanone	0.488	0.441		-9.63	20
Carbon Tetrachloride	0.577	0.562		-2.6	20
2,2-Dichloropropane	0.973	0.970		-0.31	20
cis-1,2-Dichloroethene	0.754	0.673		-10.74	20
Bromochloromethane	0.561	0.528		-5.88	20
Chloroform	1.179	1.147		-2.71	20
1,1,1-Trichloroethane	1.029	1.026		-0.29	20
Methylcyclohexane	0.569	0.561		-1.41	20
1,1-Dichloropropene	0.507	0.463		-8.68	20
Benzene	1.571	1.414		-9.99	20
1,2-Dichloroethane	0.538	0.541		0.56	20
Trichloroethene	0.375	0.365		-2.67	20
1,2-Dichloropropane	0.361	0.363		0.55	20
Dibromomethane	0.263	0.278		5.7	20
Bromodichloromethane	0.526	0.582		10.65	20
4-Methyl-2-Pentanone	0.571	0.632		10.68	20
Toluene	0.827	0.898		8.59	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3702</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/21/2025 09:08</u>
Lab File ID:	<u>VX048641.D</u>	Init. Calib. Date(s):	<u>11/07/2025 11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35 16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
t-1,3-Dichloropropene	0.541	0.590		9.06	20
cis-1,3-Dichloropropene	0.561	0.614		9.45	20
1,1,2-Trichloroethane	0.330	0.373		13.03	20
1,3-Dichloropropane	0.575	0.622		8.17	20
2-Chloroethyl Vinyl ether	0.296	0.335		13.18	20
2-Hexanone	0.425	0.462		8.71	20
Dibromochloromethane	0.406	0.463		14.04	20
1,2-Dibromoethane	0.358	0.393		9.78	20
Tetrachloroethene	0.363	0.343		-5.51	20
Chlorobenzene	1.187	1.132	0.3	-4.63	20
1,1,1,2-Tetrachloroethane	0.398	0.414		4.02	20
Hexachloroethane	0.606	0.596		-1.65	20
Ethyl Benzene	1.993	1.929		-3.21	20
m/p-Xylenes	0.751	0.741		-1.33	20
o-Xylene	0.734	0.724		-1.36	20
Styrene	1.232	1.242		0.81	20
Bromoform	0.367	0.366	0.1	-0.27	20
Isopropylbenzene	3.629	3.586		-1.18	20
1,1,2,2-Tetrachloroethane	1.280	1.199	0.3	-6.33	20
1,2,3-Trichloropropane	1.128	1.151		2.04	20
Bromobenzene	0.937	0.886		-5.44	20
n-propylbenzene	4.268	4.207		-1.43	20
2-Chlorotoluene	2.579	2.493		-3.34	20
1,3,5-Trimethylbenzene	2.959	2.914		-1.52	20
4-Chlorotoluene	3.025	2.937		-2.91	20
tert-Butylbenzene	3.060	3.007		-1.73	20
1,2,4-Trimethylbenzene	2.985	2.936		-1.64	20
sec-Butylbenzene	3.782	3.778		-0.11	20
p-Isopropyltoluene	3.180	3.186		0.19	20
1,3-Dichlorobenzene	1.756	1.641		-6.55	20
1,4-Dichlorobenzene	1.814	1.676		-7.61	20
n-Butylbenzene	2.983	2.977		-0.2	20
1,2-Dichlorobenzene	1.686	1.598		-5.22	20
1,2-Dibromo-3-Chloropropane	0.297	0.283		-4.71	20
1,2,4-Trichlorobenzene	1.120	1.041		-7.05	20
Hexachlorobutadiene	0.468	0.475		1.5	20
Naphthalene	3.603	3.793		5.27	20
1,2,3-Trichlorobenzene	1.042	1.015		-2.59	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
Lab Code: ACE SDG No.: Q3702
Instrument ID: MSVOA_X Calibration Date/Time: 11/21/2025 09:08
Lab File ID: VX048641.D Init. Calib. Date(s): 11/07/2025 11/07/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 13:35 16:29
GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.697	0.708		1.58	20
Dibromofluoromethane	0.378	0.372		-1.59	20
Toluene-d8	1.180	1.238		4.91	20
4-Bromofluorobenzene	0.440	0.470		6.82	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/24/2025
 Supervised By : Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	165802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	262765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	235899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.000	152	125375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.928	65	117461	50.842	ug/l	-0.02
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.680%			
35) Dibromofluoromethane	5.367	113	97622	49.087	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.180%			
50) Toluene-d8	8.628	98	325415	52.464	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.920%			
62) 4-Bromofluorobenzene	11.061	95	123540	53.445	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.900%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.173	85	69797	51.951	ug/l	97
3) Chloromethane	1.301	50	74862	42.012	ug/l	97
4) Vinyl Chloride	1.380	62	96674	47.917	ug/l	97
5) Bromomethane	1.611	94	58357	45.807	ug/l	96
6) Chloroethane	1.691	64	56183	42.522	ug/l	96
7) Trichlorofluoromethane	1.892	101	149931	46.973	ug/l	100
8) Diethyl Ether	2.130	74	54956	42.893	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.331	101	98112	53.096	ug/l	98
10) Methyl Iodide	2.453	142	130114	47.580	ug/l	100
11) Tert butyl alcohol	2.934	59	114258	260.995	ug/l	98
12) 1,1-Dichloroethene	2.319	96	89162	46.195	ug/l	96
13) Acrolein	2.233	56	14465	228.368	ug/l	91
14) Allyl chloride	2.660	41	168018	46.755	ug/l	99
15) Acrylonitrile	3.050	53	331251	259.200	ug/l	99
16) Acetone	2.367	43	228670	209.099	ug/l	99
17) Carbon Disulfide	2.514	76	223980	41.238	ug/l	99
18) Methyl Acetate	2.691	43	146761	48.833	ug/l	99
19) Methyl tert-butyl Ether	3.099	73	348713	51.162	ug/l	99
20) Methylene Chloride	2.782	84	107242	47.698	ug/l	97
21) trans-1,2-Dichloroethene	3.087	96	97026	47.430	ug/l	97
22) Diisopropyl ether	3.745	45	357493	51.808	ug/l	95
23) Vinyl Acetate	3.709	43	1535926	254.149	ug/l	99
24) 1,1-Dichloroethane	3.599	63	183152	48.373	ug/l	99
25) 2-Butanone	4.526	43	365805	226.114	ug/l	97
26) 2,2-Dichloropropane	4.458	77	160786	49.812	ug/l	98
27) cis-1,2-Dichloroethene	4.471	96	111643	44.648	ug/l	94
28) Bromochloromethane	4.879	49	87502	47.044	ug/l	97
29) Tetrahydrofuran	4.977	42	261645	226.760	ug/l	99
30) Chloroform	5.074	83	190100	48.634	ug/l	96
31) Cyclohexane	5.452	56	142771	46.021	ug/l	100
32) 1,1,1-Trichloroethane	5.367	97	170121	49.870	ug/l	99
36) 1,1-Dichloropropene	5.672	75	121610	45.666	ug/l	99
37) Ethyl Acetate	4.684	43	160025	42.689	ug/l	96
38) Carbon Tetrachloride	5.660	117	147713	48.727	ug/l	98
39) Methylcyclohexane	7.360	83	147285	49.257	ug/l	99
40) Benzene	6.019	78	371515	45.005	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/24/2025
 Supervised By : Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.891	41	81982	49.730	ug/l	95
42) 1,2-Dichloroethane	6.062	62	142161	50.263	ug/l	97
43) Isopropyl Acetate	6.318	43	251531	47.734	ug/l	99
44) Trichloroethene	7.104	130	95926	48.616	ug/l	90
45) 1,2-Dichloropropane	7.409	63	95294	50.206	ug/l	95
46) Dibromomethane	7.562	93	72982	52.719	ug/l	97
47) Bromodichloromethane	7.799	83	152885	55.358	ug/l	98
48) Methyl methacrylate	7.671	41	124745	53.254	ug/l	95
49) 1,4-Dioxane	7.641	88	41733	1169.221	ug/l	95
51) 4-Methyl-2-Pentanone	8.555	43	830144	276.759	ug/l	99
52) Toluene	8.702	92	235955	54.292	ug/l	100
53) t-1,3-Dichloropropene	8.958	75	155065	54.490	ug/l	100
54) cis-1,3-Dichloropropene	8.348	75	161344	54.755	ug/l	96
55) 1,1,2-Trichloroethane	9.134	97	97896	56.379	ug/l	98
56) Ethyl methacrylate	9.098	69	163057	56.443	ug/l	96
57) 1,3-Dichloropropane	9.293	76	163380	54.057	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.220	63	440769	283.789	ug/l	99
59) 2-Hexanone	9.409	43	607109	272.098	ug/l	98
60) Dibromochloromethane	9.500	129	121704	57.087	ug/l	98
61) 1,2-Dibromoethane	9.592	107	103217	54.883	ug/l	100
64) Tetrachloroethene	9.256	164	80887	47.209	ug/l	94
65) Chlorobenzene	10.061	112	267037	47.693	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.147	131	97653	52.052	ug/l	99
67) Ethyl Benzene	10.177	91	455038	48.397	ug/l	98
68) m/p-Xylenes	10.281	106	349467	98.689	ug/l	97
69) o-Xylene	10.622	106	170714	49.274	ug/l	100
70) Styrene	10.634	104	292981	50.411	ug/l	99
71) Bromoform	10.781	173	86299	49.860	ug/l #	99
73) Isopropylbenzene	10.945	105	449572	49.401	ug/l	99
74) N-amyl acetate	10.823	43	212984	46.610	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.195	83	150331	46.842	ug/l	99
76) 1,2,3-Trichloropropane	11.219	75	144332m	47.785	ug/l	
77) Bromobenzene	11.177	156	111069	47.264	ug/l	99
78) n-propylbenzene	11.287	91	527451	49.291	ug/l	100
79) 2-Chlorotoluene	11.348	91	312507	48.330	ug/l	99
80) 1,3,5-Trimethylbenzene	11.433	105	365358	49.247	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	51625	48.888	ug/l #	79
82) 4-Chlorotoluene	11.433	91	368235	48.549	ug/l	100
83) tert-Butylbenzene	11.695	119	376969	49.127	ug/l	97
84) 1,2,4-Trimethylbenzene	11.732	105	368114	49.177	ug/l	99
85) sec-Butylbenzene	11.872	105	473697	49.947	ug/l	100
86) p-Isopropyltoluene	11.988	119	399435	50.096	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	205748	46.714	ug/l	98
88) 1,4-Dichlorobenzene	12.024	146	210178	46.200	ug/l	100
89) n-Butylbenzene	12.311	91	373206	49.895	ug/l	100
90) Hexachloroethane	12.518	117	74749	49.205	ug/l	99
91) 1,2-Dichlorobenzene	12.317	146	200374	47.391	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.920	75	35521	47.628	ug/l	98
93) 1,2,4-Trichlorobenzene	13.567	180	130455	46.431	ug/l	100
94) Hexachlorobutadiene	13.707	225	59503	50.662	ug/l	98
95) Naphthalene	13.756	128	475597	52.649	ug/l	99
96) 1,2,3-Trichlorobenzene	13.938	180	127205	48.674	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048641.D
Acq On : 21 Nov 2025 09:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/24/2025
Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations APPROVED

Reviewed By : John Carlone 11/24/2025
 Supervised By : Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.405	0.421	-4.0	91	0.00
3 P	Chloromethane	0.537	0.452	15.8	70	0.00
4 C	Vinyl Chloride	0.608	0.583	4.1#	90	0.00
5 T	Bromomethane	0.384	0.352	8.3	76	0.00
6 T	Chloroethane	0.398	0.339	14.8	82	0.00
7 T	Trichlorofluoromethane	0.963	0.904	6.1	94	0.00
8 T	Diethyl Ether	0.386	0.331	14.2	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.592	-6.3	106	0.00
10 T	Methyl Iodide	0.825	0.785	4.8	88	0.00
11 T	Tert butyl alcohol	0.132	0.138	-4.5	92	0.00
12 CM	1,1-Dichloroethene	0.582	0.538	7.6#	91	0.00
13 T	Acrolein	0.019	0.017	10.5	80	0.00
14 T	Allyl chloride	1.084	1.013	6.5	89	0.00
15 T	Acrylonitrile	0.385	0.400	-3.9	91	0.00
16 T	Acetone	0.330	0.276	16.4	77	0.00
17 T	Carbon Disulfide	1.638	1.351	17.5	84	0.00
18 T	Methyl Acetate	0.906	0.885	2.3	94	0.00
19 T	Methyl tert-butyl Ether	2.055	2.103	-2.3	91	0.00
20 T	Methylene Chloride	0.678	0.647	4.6	91	0.00
21 T	trans-1,2-Dichloroethene	0.617	0.585	5.2	90	0.00
22 T	Diisopropyl ether	2.081	2.156	-3.6	96	0.00
23 T	Vinyl Acetate	1.822	1.853	-1.7	93	0.00
24 P	1,1-Dichloroethane	1.142	1.105	3.2	92	0.00
25 T	2-Butanone	0.488	0.441	9.6	80	-0.01
26 T	2,2-Dichloropropane	0.973	0.970	0.3	101	0.00
27 T	cis-1,2-Dichloroethene	0.754	0.673	10.7	86	-0.01
28 T	Bromochloromethane	0.561	0.528	5.9	89	0.00
29 T	Tetrahydrofuran	0.348	0.316	9.2	83	-0.01
30 C	Chloroform	1.179	1.147	2.7#	94	-0.01
31 T	Cyclohexane	0.936	0.861	8.0	91	0.00
32 T	1,1,1-Trichloroethane	1.029	1.026	0.3	97	0.00
33 S	1,2-Dichloroethane-d4	0.697	0.708	-1.6	95	-0.02
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.378	0.372	1.6	90	0.00
36 T	1,1-Dichloropropene	0.507	0.463	8.7	93	0.00
37 T	Ethyl Acetate	0.713	0.609	14.6	78	-0.01
38 T	Carbon Tetrachloride	0.577	0.562	2.6	99	0.00
39 T	Methylcyclohexane	0.569	0.561	1.4	94	0.00
40 TM	Benzene	1.571	1.414	10.0	86	0.00
41 T	Methacrylonitrile	0.314	0.312	0.6	90	-0.01
42 TM	1,2-Dichloroethane	0.538	0.541	-0.6	96	-0.01
43 T	Isopropyl Acetate	1.003	0.957	4.6	88	0.00
44 TM	Trichloroethene	0.375	0.365	2.7	88	0.00
45 C	1,2-Dichloropropane	0.361	0.363	-0.6#	86	0.00
46 T	Dibromomethane	0.263	0.278	-5.7	89	0.00
47 T	Bromodichloromethane	0.526	0.582	-10.6	92	0.00
48 T	Methyl methacrylate	0.446	0.475	-6.5	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	92	0.00
50 S	Toluene-d8	1.180	1.238	-4.9	88	0.00
51 T	4-Methyl-2-Pentanone	0.571	0.632	-10.7	87	0.00
52 CM	Toluene	0.827	0.898	-8.6#	88	0.00
53 T	t-1,3-Dichloropropene	0.541	0.590	-9.1	88	0.00
54 T	cis-1,3-Dichloropropene	0.561	0.614	-9.4	87	0.00
55 T	1,1,2-Trichloroethane	0.330	0.373	-13.0	88	0.00
56 T	Ethyl methacrylate	0.550	0.621	-12.9	86	0.00
57 T	1,3-Dichloropropane	0.575	0.622	-8.2	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.296	0.335	-13.2	81	0.00
59 T	2-Hexanone	0.425	0.462	-8.7	85	0.00
60 T	Dibromochloromethane	0.406	0.463	-14.0	91	0.00
61 T	1,2-Dibromoethane	0.358	0.393	-9.8	88	0.00
62 S	4-Bromofluorobenzene	0.440	0.470	-6.8	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.363	0.343	5.5	91	0.00
65 PM	Chlorobenzene	1.187	1.132	4.6	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.414	-4.0	92	0.00
67 C	Ethyl Benzene	1.993	1.929	3.2#	89	0.00
68 T	m/p-Xylenes	0.751	0.741	1.3	89	0.00
69 T	o-Xylene	0.734	0.724	1.4	89	0.00
70 T	Styrene	1.232	1.242	-0.8	89	0.00
71 P	Bromoform	0.367	0.366	0.3	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.629	3.586	1.2	95	0.00
74 T	N-amyl acetate	1.822	1.699	6.8	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.280	1.199	6.3	89	0.00
76 T	1,2,3-Trichloropropane	1.205	1.151	4.5	72	0.00
77 T	Bromobenzene	0.937	0.886	5.4	89	0.00
78 T	n-propylbenzene	4.268	4.207	1.4	96	0.00
79 T	2-Chlorotoluene	2.579	2.493	3.3	94	0.00
80 T	1,3,5-Trimethylbenzene	2.959	2.914	1.5	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.421	0.412	2.1	87	0.00
82 T	4-Chlorotoluene	3.025	2.937	2.9	94	0.00
83 T	tert-Butylbenzene	3.060	3.007	1.7	93	0.00
84 T	1,2,4-Trimethylbenzene	2.985	2.936	1.6	93	0.00
85 T	sec-Butylbenzene	3.782	3.778	0.1	99	0.00
86 T	p-Isopropyltoluene	3.180	3.186	-0.2	98	0.00
87 T	1,3-Dichlorobenzene	1.756	1.641	6.5	91	0.00
88 T	1,4-Dichlorobenzene	1.814	1.676	7.6	91	0.00
89 T	n-Butylbenzene	2.983	2.977	0.2	100	0.00
90 T	Hexachloroethane	0.606	0.596	1.7	99	0.00
91 T	1,2-Dichlorobenzene	1.686	1.598	5.2	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.297	0.283	4.7	87	0.00
93 T	1,2,4-Trichlorobenzene	1.120	1.041	7.1	89	0.00
94 T	Hexachlorobutadiene	0.468	0.475	-1.5	107	0.00
95 T	Naphthalene	3.603	3.793	-5.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048641.D
Acq On : 21 Nov 2025 09:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 21 22:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.042	1.015	2.6	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	50.000	51.951	-3.9	91	0.00
3 P	Chloromethane	50.000	42.012	16.0	70	0.00
4 C	Vinyl Chloride	50.000	47.917	4.2#	90	0.00
5 T	Bromomethane	50.000	45.807	8.4	76	0.00
6 T	Chloroethane	50.000	42.522	15.0	82	0.00
7 T	Trichlorofluoromethane	50.000	46.973	6.1	94	0.00
8 T	Diethyl Ether	50.000	42.893	14.2	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.096	-6.2	106	0.00
10 T	Methyl Iodide	50.000	47.580	4.8	88	0.00
11 T	Tert butyl alcohol	250.000	260.995	-4.4	92	0.00
12 CM	1,1-Dichloroethene	50.000	46.195	7.6#	91	0.00
13 T	Acrolein	250.000	228.368	8.7	80	0.00
14 T	Allyl chloride	50.000	46.755	6.5	89	0.00
15 T	Acrylonitrile	250.000	259.200	-3.7	91	0.00
16 T	Acetone	250.000	209.099	16.4	77	0.00
17 T	Carbon Disulfide	50.000	41.238	17.5	84	0.00
18 T	Methyl Acetate	50.000	48.833	2.3	94	0.00
19 T	Methyl tert-butyl Ether	50.000	51.162	-2.3	91	0.00
20 T	Methylene Chloride	50.000	47.698	4.6	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.430	5.1	90	0.00
22 T	Diisopropyl ether	50.000	51.808	-3.6	96	0.00
23 T	Vinyl Acetate	250.000	254.149	-1.7	93	0.00
24 P	1,1-Dichloroethane	50.000	48.373	3.3	92	0.00
25 T	2-Butanone	250.000	226.114	9.6	80	-0.01
26 T	2,2-Dichloropropane	50.000	49.812	0.4	101	0.00
27 T	cis-1,2-Dichloroethene	50.000	44.648	10.7	86	-0.01
28 T	Bromochloromethane	50.000	47.044	5.9	89	0.00
29 T	Tetrahydrofuran	250.000	226.760	9.3	83	-0.01
30 C	Chloroform	50.000	48.634	2.7#	94	-0.01
31 T	Cyclohexane	50.000	46.021	8.0	91	0.00
32 T	1,1,1-Trichloroethane	50.000	49.870	0.3	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.842	-1.7	95	-0.02
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	85	0.00
35 S	Dibromofluoromethane	50.000	49.087	1.8	90	0.00
36 T	1,1-Dichloropropene	50.000	45.666	8.7	93	0.00
37 T	Ethyl Acetate	50.000	42.689	14.6	78	-0.01
38 T	Carbon Tetrachloride	50.000	48.727	2.5	99	0.00
39 T	Methylcyclohexane	50.000	49.257	1.5	94	0.00
40 TM	Benzene	50.000	45.005	10.0	86	0.00
41 T	Methacrylonitrile	50.000	49.730	0.5	90	-0.01
42 TM	1,2-Dichloroethane	50.000	50.263	-0.5	96	-0.01
43 T	Isopropyl Acetate	50.000	47.734	4.5	88	0.00
44 TM	Trichloroethene	50.000	48.616	2.8	88	0.00
45 C	1,2-Dichloropropane	50.000	50.206	-0.4#	86	0.00
46 T	Dibromomethane	50.000	52.719	-5.4	89	0.00
47 T	Bromodichloromethane	50.000	55.358	-10.7	92	0.00
48 T	Methyl methacrylate	50.000	53.254	-6.5	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048641.D
 Acq On : 21 Nov 2025 09:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 22:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1169.221	-16.9	92	0.00
50 S	Toluene-d8	50.000	52.464	-4.9	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	276.759	-10.7	87	0.00
52 CM	Toluene	50.000	54.292	-8.6#	88	0.00
53 T	t-1,3-Dichloropropene	50.000	54.490	-9.0	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.755	-9.5	87	0.00
55 T	1,1,2-Trichloroethane	50.000	56.379	-12.8	88	0.00
56 T	Ethyl methacrylate	50.000	56.443	-12.9	86	0.00
57 T	1,3-Dichloropropane	50.000	54.057	-8.1	86	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	283.789	-13.5	81	0.00
59 T	2-Hexanone	250.000	272.098	-8.8	85	0.00
60 T	Dibromochloromethane	50.000	57.087	-14.2	91	0.00
61 T	1,2-Dibromoethane	50.000	54.883	-9.8	88	0.00
62 S	4-Bromofluorobenzene	50.000	53.445	-6.9	91	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	87	0.00
64 T	Tetrachloroethene	50.000	47.209	5.6	91	0.00
65 PM	Chlorobenzene	50.000	47.693	4.6	89	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.052	-4.1	92	0.00
67 C	Ethyl Benzene	50.000	48.397	3.2#	89	0.00
68 T	m/p-Xylenes	100.000	98.689	1.3	89	0.00
69 T	o-Xylene	50.000	49.274	1.5	89	0.00
70 T	Styrene	50.000	50.411	-0.8	89	0.00
71 P	Bromoform	50.000	49.860	0.3	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	91	0.00
73 T	Isopropylbenzene	50.000	49.401	1.2	95	0.00
74 T	N-ethyl acetate	50.000	46.610	6.8	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.842	6.3	89	0.00
76 T	1,2,3-Trichloropropane	50.000	47.785	4.4	72	0.00
77 T	Bromobenzene	50.000	47.264	5.5	89	0.00
78 T	n-propylbenzene	50.000	49.291	1.4	96	0.00
79 T	2-Chlorotoluene	50.000	48.330	3.3	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.247	1.5	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.888	2.2	87	0.00
82 T	4-Chlorotoluene	50.000	48.549	2.9	94	0.00
83 T	tert-Butylbenzene	50.000	49.127	1.7	93	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.177	1.6	93	0.00
85 T	sec-Butylbenzene	50.000	49.947	0.1	99	0.00
86 T	p-Isopropyltoluene	50.000	50.096	-0.2	98	0.00
87 T	1,3-Dichlorobenzene	50.000	46.714	6.6	91	0.00
88 T	1,4-Dichlorobenzene	50.000	46.200	7.6	91	0.00
89 T	n-Butylbenzene	50.000	49.895	0.2	100	0.00
90 T	Hexachloroethane	50.000	49.205	1.6	99	0.00
91 T	1,2-Dichlorobenzene	50.000	47.391	5.2	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.628	4.7	87	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.431	7.1	89	0.00
94 T	Hexachlorobutadiene	50.000	50.662	-1.3	107	0.00
95 T	Naphthalene	50.000	52.649	-5.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048641.D
Acq On : 21 Nov 2025 09:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 21 22:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.674	2.7	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048515.D
Acq On : 07 Nov 2025 13:06
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Title : SW846 8260
Last Update : Sat Nov 08 05:08:11 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	10458	PASS
75	95	30	60	56.0	27675	PASS
95	95	100	100	100.0	49437	PASS
96	95	5	9	6.8	3386	PASS
173	174	0.00	2	1.2	472	PASS
174	95	50	100	77.2	38160	PASS
175	174	5	9	7.3	2798	PASS
176	174	95	101	97.9	37347	PASS
177	176	5	9	6.2	2308	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048640.D
Acq On : 21 Nov 2025 08:00
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Title : SW846 8260
Last Update : Sat Nov 08 05:08:11 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	4938	PASS
75	95	30	60	49.6	14581	PASS
95	95	100	100	100.0	29421	PASS
96	95	5	9	6.8	1998	PASS
173	174	0.00	2	0.6	132	PASS
174	95	50	100	77.4	22781	PASS
175	174	5	9	8.0	1817	PASS
176	174	95	101	96.4	21955	PASS
177	176	5	9	6.3	1387	PASS

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: VX1121WBL01
 Lab Sample ID: VX1121WBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 09:58	VX112125
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 09:58	VX112125
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	11/21/25 09:58	VX112125
141-78-6	Ethyl Acetate	0.31	U	1	0.31	1.00	ug/L	11/21/25 09:58	VX112125
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	11/21/25 09:58	VX112125
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	11/21/25 09:58	VX112125
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	11/21/25 09:58	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 09:58	VX112125
75-65-0	Tert butyl alcohol	5.50	U	1	5.50	25.0	ug/L	11/21/25 09:58	VX112125
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 09:58	VX112125
107-02-8	Acrolein	7.10	U	1	7.10	25.0	ug/L	11/21/25 09:58	VX112125
107-13-1	Acrylonitrile	0.83	U	1	0.83	5.00	ug/L	11/21/25 09:58	VX112125
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	11/21/25 09:58	VX112125
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	11/21/25 09:58	VX112125
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	11/21/25 09:58	VX112125
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	11/21/25 09:58	VX112125
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	11/21/25 09:58	VX112125
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 09:58	VX112125
108-05-4	Vinyl Acetate	0.66	U	1	0.66	5.00	ug/L	11/21/25 09:58	VX112125
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	11/21/25 09:58	VX112125
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	11/21/25 09:58	VX112125
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	11/21/25 09:58	VX112125
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	11/21/25 09:58	VX112125
594-20-7	2,2-Dichloropropane	0.21	U	1	0.21	1.00	ug/L	11/21/25 09:58	VX112125
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	11/21/25 09:58	VX112125
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 09:58	VX112125
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	11/21/25 09:58	VX112125
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	11/21/25 09:58	VX112125
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	11/21/25 09:58	VX112125
563-58-6	1,1-Dichloropropene	0.15	U	1	0.15	1.00	ug/L	11/21/25 09:58	VX112125
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 09:58	VX112125
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 09:58	VX112125
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	11/21/25 09:58	VX112125
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	11/21/25 09:58	VX112125
74-95-3	Dibromomethane	0.25	U	1	0.25	1.00	ug/L	11/21/25 09:58	VX112125
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	11/21/25 09:58	VX112125
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	11/21/25 09:58	VX112125
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 09:58	VX112125
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	11/21/25 09:58	VX112125
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	11/21/25 09:58	VX112125
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	11/21/25 09:58	VX112125
142-28-9	1,3-Dichloropropane	0.19	U	1	0.19	1.00	ug/L	11/21/25 09:58	VX112125
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	1	0.30	5.00	ug/L	11/21/25 09:58	VX112125

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: VX1121WBL01
 Lab Sample ID: VX1121WBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	11/21/25 09:58	VX112125
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	11/21/25 09:58	VX112125
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	11/21/25 09:58	VX112125
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	11/21/25 09:58	VX112125
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 09:58	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	1	0.19	1.00	ug/L	11/21/25 09:58	VX112125
67-72-1	Hexachloroethane	0.32	U	1	0.32	1.00	ug/L	11/21/25 09:58	VX112125
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 09:58	VX112125
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	11/21/25 09:58	VX112125
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	11/21/25 09:58	VX112125
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	11/21/25 09:58	VX112125
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	11/21/25 09:58	VX112125
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	11/21/25 09:58	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	11/21/25 09:58	VX112125
96-18-4	1,2,3-Trichloropropane	0.35	U	1	0.35	1.00	ug/L	11/21/25 09:58	VX112125
108-86-1	Bromobenzene	0.24	U	1	0.24	1.00	ug/L	11/21/25 09:58	VX112125
103-65-1	n-propylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 09:58	VX112125
95-49-8	2-Chlorotoluene	0.14	U	1	0.14	1.00	ug/L	11/21/25 09:58	VX112125
108-67-8	1,3,5-Trimethylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 09:58	VX112125
106-43-4	4-Chlorotoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 09:58	VX112125
98-06-6	tert-Butylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 09:58	VX112125
95-63-6	1,2,4-Trimethylbenzene	0.14	U	1	0.14	1.00	ug/L	11/21/25 09:58	VX112125
135-98-8	sec-Butylbenzene	0.13	U	1	0.13	1.00	ug/L	11/21/25 09:58	VX112125
99-87-6	p-Isopropyltoluene	0.13	U	1	0.13	1.00	ug/L	11/21/25 09:58	VX112125
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 09:58	VX112125
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	11/21/25 09:58	VX112125
104-51-8	n-Butylbenzene	0.15	U	1	0.15	1.00	ug/L	11/21/25 09:58	VX112125
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	11/21/25 09:58	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	11/21/25 09:58	VX112125
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 09:58	VX112125
87-68-3	Hexachlorobutadiene	0.30	U	1	0.30	1.00	ug/L	11/21/25 09:58	VX112125
91-20-3	Naphthalene	0.20	U	1	0.20	1.00	ug/L	11/21/25 09:58	VX112125
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	11/21/25 09:58	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	54.3		74 - 125	109%	SPK: 50
1868-53-7	Dibromofluoromethane	46.2		75 - 124	92%	SPK: 50
2037-26-5	Toluene-d8	52.2		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	59.0		77 - 121	118%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	203000
540-36-3	1,4-Difluorobenzene	383000
3114-55-4	Chlorobenzene-d5	433000
3855-82-1	1,4-Dichlorobenzene-d4	228000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Chemtech Consulting Group

Project: Weekly Storage Blanks

Client Sample ID: VX1121WBL01

Lab Sample ID: VX1121WBL01

Analytical Method: 8260D

Sample Wt/Vol: 5 mL

Level: LOW

Final Vol: 5000 uL

Date Collected:

Date Received:

SDG No.: Q3702

Matrix: Water

% Solid: 0

Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048643.D
 Acq On : 21 Nov 2025 09:58
 Operator : JC/MD
 Sample : VX1121WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121WBL01

Quant Time: Nov 21 22:08:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	203389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	383341	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	433179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	227916	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	153859	54.289	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.580%
35) Dibromofluoromethane	5.373	113	134165	46.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	8.635	98	472658	52.234	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.460%
62) 4-Bromofluorobenzene	11.061	95	199123	59.048	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048643.D
Acq On : 21 Nov 2025 09:58
Operator : JC/MD
Sample : VX1121WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121WBL01

Quant Time: Nov 21 22:08:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX048643.D
Acq: 21 Nov 2025 09:58




Instrument :
MSVOA_X
ClientSampleId :
VX1121WBL01


Tgt Ion:168 Resp: 203389
Ion Ratio Lower Upper
168 100
99 57.9 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 54.289 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.000 min
Lab File: VX048643.D
Acq: 21 Nov 2025 09:58

Tgt Ion: 65 Resp: 153859
Ion Ratio Lower Upper
65 100
67 51.2 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.745 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX048643.D

Acq: 21 Nov 2025 09:58

Instrument :

MSVOA_X

ClientSampleId :

VX1121WBL01





Tgt Ion:114 Resp: 383341

Ion Ratio Lower Upper

114 100

63 19.9 0.0 41.2

88 15.7 0.0 32.2



#35

Dibromofluoromethane

Concen: 46.242 ug/l

RT: 5.373 min Scan# 703

Delta R.T. -0.000 min

Lab File: VX048643.D

Acq: 21 Nov 2025 09:58




Tgt Ion:113 Resp: 134165


Ion Ratio Lower Upper

113 100

111 101.8 81.9 122.9

192 19.6 15.8 23.6





#50

Toluene-d8

Concen: 52.234 ug/l

RT: 8.635 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048643.D

Acq: 21 Nov 2025 09:58

Instrument :

MSVOA_X

ClientSampleId :





VX1121WBL01

Tgt Ion: 98 Resp: 472658

Ion Ratio Lower Upper

98 100

100 65.5 53.6 80.4



#62

4-Bromofluorobenzene

Concen: 59.048 ug/l

RT: 11.061 min Scan# 1636

Delta R.T. -0.000 min

Lab File: VX048643.D

Acq: 21 Nov 2025 09:58




Tgt Ion: 95 Resp: 199123


Ion Ratio Lower Upper

95 100

174 80.8 0.0 161.8

176 78.0 0.0 153.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX048643.D
Acq: 21 Nov 2025 09:58




Instrument :
MSVOA_X
ClientSampleId :
VX1121WBL01

Tgt Ion:117 Resp: 433179
Ion Ratio Lower Upper
117 100
82 56.9 44.4 66.6
119 31.8 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX048643.D
Acq: 21 Nov 2025 09:58

Tgt Ion:152 Resp: 227916
Ion Ratio Lower Upper
152 100
115 58.6 39.2 117.6
150 155.5 0.0 347.0



Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: VX1121WBS01
 Lab Sample ID: VX1121WBS01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	20.5		1	0.22	1.00	ug/L	11/21/25 10:26	VX112125
74-87-3	Chloromethane	17.0		1	0.32	1.00	ug/L	11/21/25 10:26	VX112125
75-01-4	Vinyl Chloride	18.0		1	0.26	1.00	ug/L	11/21/25 10:26	VX112125
141-78-6	Ethyl Acetate	18.6		1	0.31	1.00	ug/L	11/21/25 10:26	VX112125
74-83-9	Bromomethane	19.8		1	1.40	5.00	ug/L	11/21/25 10:26	VX112125
75-00-3	Chloroethane	17.9		1	0.47	1.00	ug/L	11/21/25 10:26	VX112125
75-69-4	Trichlorofluoromethane	18.8		1	0.33	1.00	ug/L	11/21/25 10:26	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	20.4		1	0.25	1.00	ug/L	11/21/25 10:26	VX112125
75-65-0	Tert butyl alcohol	98.6		1	5.50	25.0	ug/L	11/21/25 10:26	VX112125
75-35-4	1,1-Dichloroethene	18.1		1	0.23	1.00	ug/L	11/21/25 10:26	VX112125
107-02-8	Acrolein	82.1		1	7.10	25.0	ug/L	11/21/25 10:26	VX112125
107-13-1	Acrylonitrile	98.8		1	0.83	5.00	ug/L	11/21/25 10:26	VX112125
67-64-1	Acetone	88.0		1	1.50	5.00	ug/L	11/21/25 10:26	VX112125
75-15-0	Carbon Disulfide	15.5		1	0.21	1.00	ug/L	11/21/25 10:26	VX112125
1634-04-4	Methyl tert-butyl Ether	19.7		1	0.16	1.00	ug/L	11/21/25 10:26	VX112125
79-20-9	Methyl Acetate	19.9		1	0.27	1.00	ug/L	11/21/25 10:26	VX112125
75-09-2	Methylene Chloride	18.4		1	0.28	1.00	ug/L	11/21/25 10:26	VX112125
156-60-5	trans-1,2-Dichloroethene	18.3		1	0.23	1.00	ug/L	11/21/25 10:26	VX112125
108-05-4	Vinyl Acetate	96.1		1	0.66	5.00	ug/L	11/21/25 10:26	VX112125
75-34-3	1,1-Dichloroethane	19.4		1	0.23	1.00	ug/L	11/21/25 10:26	VX112125
110-82-7	Cyclohexane	19.3		1	1.50	5.00	ug/L	11/21/25 10:26	VX112125
78-93-3	2-Butanone	97.6		1	0.98	5.00	ug/L	11/21/25 10:26	VX112125
56-23-5	Carbon Tetrachloride	18.7		1	0.25	1.00	ug/L	11/21/25 10:26	VX112125
594-20-7	2,2-Dichloropropane	19.5		1	0.21	1.00	ug/L	11/21/25 10:26	VX112125
156-59-2	cis-1,2-Dichloroethene	18.8		1	0.19	1.00	ug/L	11/21/25 10:26	VX112125
74-97-5	Bromochloromethane	16.9		1	0.22	1.00	ug/L	11/21/25 10:26	VX112125
67-66-3	Chloroform	20.3		1	0.25	1.00	ug/L	11/21/25 10:26	VX112125
71-55-6	1,1,1-Trichloroethane	19.9		1	0.20	1.00	ug/L	11/21/25 10:26	VX112125
108-87-2	Methylcyclohexane	19.1		1	0.16	1.00	ug/L	11/21/25 10:26	VX112125
563-58-6	1,1-Dichloropropene	18.2		1	0.15	1.00	ug/L	11/21/25 10:26	VX112125
71-43-2	Benzene	18.5		1	0.15	1.00	ug/L	11/21/25 10:26	VX112125
107-06-2	1,2-Dichloroethane	19.5		1	0.22	1.00	ug/L	11/21/25 10:26	VX112125
79-01-6	Trichloroethene	19.9		1	0.090	1.00	ug/L	11/21/25 10:26	VX112125
78-87-5	1,2-Dichloropropane	20.6		1	0.20	1.00	ug/L	11/21/25 10:26	VX112125
74-95-3	Dibromomethane	20.6		1	0.25	1.00	ug/L	11/21/25 10:26	VX112125
75-27-4	Bromodichloromethane	21.8		1	0.22	1.00	ug/L	11/21/25 10:26	VX112125
108-10-1	4-Methyl-2-Pentanone	110		1	0.68	5.00	ug/L	11/21/25 10:26	VX112125
108-88-3	Toluene	21.7		1	0.14	1.00	ug/L	11/21/25 10:26	VX112125
10061-02-6	t-1,3-Dichloropropene	20.8		1	0.17	1.00	ug/L	11/21/25 10:26	VX112125
10061-01-5	cis-1,3-Dichloropropene	21.6		1	0.16	1.00	ug/L	11/21/25 10:26	VX112125
79-00-5	1,1,2-Trichloroethane	22.7		1	0.21	1.00	ug/L	11/21/25 10:26	VX112125
142-28-9	1,3-Dichloropropane	21.9		1	0.19	1.00	ug/L	11/21/25 10:26	VX112125
110-75-8	2-Chloroethyl Vinyl ether	91.4		1	0.30	5.00	ug/L	11/21/25 10:26	VX112125

Report of Analysis

Client: Chemtech Consulting Group
Project: Weekly Storage Blanks
Client Sample ID: VX1121WBS01
Lab Sample ID: VX1121WBS01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3702
Matrix: Water
% Solid: 0
Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	110		1	0.89	5.00	ug/L	11/21/25 10:26	VX112125
124-48-1	Dibromochloromethane	22.6		1	0.18	1.00	ug/L	11/21/25 10:26	VX112125
106-93-4	1,2-Dibromoethane	21.7		1	0.15	1.00	ug/L	11/21/25 10:26	VX112125
127-18-4	Tetrachloroethene	18.7		1	0.23	1.00	ug/L	11/21/25 10:26	VX112125
108-90-7	Chlorobenzene	19.0		1	0.12	1.00	ug/L	11/21/25 10:26	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	20.6		1	0.19	1.00	ug/L	11/21/25 10:26	VX112125
67-72-1	Hexachloroethane	19.3		1	0.32	1.00	ug/L	11/21/25 10:26	VX112125
100-41-4	Ethyl Benzene	19.2		1	0.13	1.00	ug/L	11/21/25 10:26	VX112125
179601-23-1	m/p-Xylenes	39.0		1	0.24	2.00	ug/L	11/21/25 10:26	VX112125
95-47-6	o-Xylene	19.2		1	0.12	1.00	ug/L	11/21/25 10:26	VX112125
100-42-5	Styrene	19.7		1	0.15	1.00	ug/L	11/21/25 10:26	VX112125
75-25-2	Bromoform	19.1		1	0.19	1.00	ug/L	11/21/25 10:26	VX112125
98-82-8	Isopropylbenzene	20.0		1	0.12	1.00	ug/L	11/21/25 10:26	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	19.4		1	0.26	1.00	ug/L	11/21/25 10:26	VX112125
96-18-4	1,2,3-Trichloropropane	18.8		1	0.35	1.00	ug/L	11/21/25 10:26	VX112125
108-86-1	Bromobenzene	19.8		1	0.24	1.00	ug/L	11/21/25 10:26	VX112125
103-65-1	n-propylbenzene	19.7		1	0.13	1.00	ug/L	11/21/25 10:26	VX112125
95-49-8	2-Chlorotoluene	19.5		1	0.14	1.00	ug/L	11/21/25 10:26	VX112125
108-67-8	1,3,5-Trimethylbenzene	19.7		1	0.15	1.00	ug/L	11/21/25 10:26	VX112125
106-43-4	4-Chlorotoluene	19.6		1	0.13	1.00	ug/L	11/21/25 10:26	VX112125
98-06-6	tert-Butylbenzene	19.7		1	0.14	1.00	ug/L	11/21/25 10:26	VX112125
95-63-6	1,2,4-Trimethylbenzene	20.0		1	0.14	1.00	ug/L	11/21/25 10:26	VX112125
135-98-8	sec-Butylbenzene	19.9		1	0.13	1.00	ug/L	11/21/25 10:26	VX112125
99-87-6	p-Isopropyltoluene	19.9		1	0.13	1.00	ug/L	11/21/25 10:26	VX112125
541-73-1	1,3-Dichlorobenzene	19.0		1	0.16	1.00	ug/L	11/21/25 10:26	VX112125
106-46-7	1,4-Dichlorobenzene	18.7		1	0.19	1.00	ug/L	11/21/25 10:26	VX112125
104-51-8	n-Butylbenzene	19.4		1	0.15	1.00	ug/L	11/21/25 10:26	VX112125
95-50-1	1,2-Dichlorobenzene	18.7		1	0.16	1.00	ug/L	11/21/25 10:26	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	18.6		1	0.53	1.00	ug/L	11/21/25 10:26	VX112125
120-82-1	1,2,4-Trichlorobenzene	18.4		1	0.20	1.00	ug/L	11/21/25 10:26	VX112125
87-68-3	Hexachlorobutadiene	18.3		1	0.30	1.00	ug/L	11/21/25 10:26	VX112125
91-20-3	Naphthalene	18.7		1	0.20	1.00	ug/L	11/21/25 10:26	VX112125
87-61-6	1,2,3-Trichlorobenzene	18.8		1	0.20	1.00	ug/L	11/21/25 10:26	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	51.8	74 - 125	104%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5	75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	53.5	86 - 113	107%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.5	77 - 121	105%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	173000
540-36-3	1,4-Difluorobenzene	282000
3114-55-4	Chlorobenzene-d5	253000
3855-82-1	1,4-Dichlorobenzene-d4	130000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Chemtech Consulting Group

Project: Weekly Storage Blanks

Client Sample ID: VX1121WBS01

Lab Sample ID: VX1121WBS01

Analytical Method: 8260D

Sample Wt/Vol: 5 mL

Level: LOW

Final Vol: 5000 uL

Date Collected:

Date Received:

SDG No.: Q3702

Matrix: Water

% Solid: 0

Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048644.D
 Acq On : 21 Nov 2025 10:26
 Operator : JC/MD
 Sample : VX1121WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/24/2025
 Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:08:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	173428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	281921	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	253067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129952	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.934	65	125111	51.772	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.540%	
35) Dibromofluoromethane	5.367	113	103564	48.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.080%	
50) Toluene-d8	8.628	98	356013	53.497	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.000%	
62) 4-Bromofluorobenzene	11.061	95	130107	52.462	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.920%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.173	85	28746	20.455	ug/l	99
3) Chloromethane	1.301	50	31654	16.983	ug/l	98
4) Vinyl Chloride	1.380	62	37908	17.963	ug/l	99
5) Bromomethane	1.618	94	26440	19.841	ug/l	99
6) Chloroethane	1.697	64	24695	17.868	ug/l	93
7) Trichlorofluoromethane	1.892	101	62887	18.836	ug/l	98
8) Diethyl Ether	2.130	74	24685	18.420	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	39481	20.427	ug/l	96
10) Methyl Iodide	2.453	142	46129	16.127	ug/l	98
11) Tert butyl alcohol	2.928	59	45147	98.593	ug/l	99
12) 1,1-Dichloroethene	2.325	96	36605	18.131	ug/l	95
13) Acrolein	2.233	56	5442	82.138	ug/l	94
14) Allyl chloride	2.666	41	70167	18.667	ug/l	97
15) Acrylonitrile	3.050	53	132025	98.765	ug/l	99
16) Acetone	2.367	43	100673	88.009	ug/l	96
17) Carbon Disulfide	2.514	76	88039	15.497	ug/l	100
18) Methyl Acetate	2.697	43	62659	19.932	ug/l	99
19) Methyl tert-butyl Ether	3.105	73	140532	19.712	ug/l	99
20) Methylene Chloride	2.782	84	43357	18.436	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	39123	18.284	ug/l	97
22) Diisopropyl ether	3.745	45	146606	20.312	ug/l	95
23) Vinyl Acetate	3.709	43	607659	96.128	ug/l	98
24) 1,1-Dichloroethane	3.599	63	76659	19.356	ug/l	99
25) 2-Butanone	4.532	43	165204	97.627	ug/l	98
26) 2,2-Dichloropropane	4.465	77	65871	19.510	ug/l	99
27) cis-1,2-Dichloroethene	4.471	96	49230	18.822	ug/l	97
28) Bromochloromethane	4.879	49	32923	16.922	ug/l	94
29) Tetrahydrofuran	4.977	42	115093	95.362	ug/l	95
30) Chloroform	5.074	83	83184	20.345	ug/l	98
31) Cyclohexane	5.452	56	62595	19.290	ug/l	100
32) 1,1,1-Trichloroethane	5.367	97	71157	19.942	ug/l	98
36) 1,1-Dichloropropene	5.678	75	52076	18.226	ug/l	99
37) Ethyl Acetate	4.690	43	74886	18.620	ug/l	98
38) Carbon Tetrachloride	5.666	117	60912	18.728	ug/l	95
39) Methylcyclohexane	7.360	83	61261	19.096	ug/l	95
40) Benzene	6.019	78	163891	18.505	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048644.D
 Acq On : 21 Nov 2025 10:26
 Operator : JC/MD
 Sample : VX1121WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/24/2025
 Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:08:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.897	41	34850	19.704	ug/l	96
42) 1,2-Dichloroethane	6.068	62	59219	19.515	ug/l	98
43) Isopropyl Acetate	6.318	43	107217	18.965	ug/l	99
44) Trichloroethene	7.110	130	42046	19.861	ug/l	98
45) 1,2-Dichloropropane	7.409	63	41920	20.585	ug/l	95
46) Dibromomethane	7.568	93	30655	20.639	ug/l	97
47) Bromodichloromethane	7.805	83	64733	21.846	ug/l	98
48) Methyl methacrylate	7.677	41	52355	20.832	ug/l	96
49) 1,4-Dioxane	7.641	88	15241	397.989	ug/l #	94
51) 4-Methyl-2-Pentanone	8.555	43	362350	112.595	ug/l	97
52) Toluene	8.702	92	101283	21.721	ug/l	98
53) t-1,3-Dichloropropene	8.964	75	63393	20.763	ug/l	99
54) cis-1,3-Dichloropropene	8.348	75	68172	21.563	ug/l	97
55) 1,1,2-Trichloroethane	9.134	97	42292	22.701	ug/l	98
56) Ethyl methacrylate	9.098	69	67930	21.916	ug/l	98
57) 1,3-Dichloropropane	9.293	76	70878	21.858	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.226	63	152309	91.401	ug/l	99
59) 2-Hexanone	9.415	43	263762	110.182	ug/l	98
60) Dibromochloromethane	9.506	129	51799	22.646	ug/l	99
61) 1,2-Dibromoethane	9.592	107	43763	21.689	ug/l	97
64) Tetrachloroethene	9.256	164	34401	18.716	ug/l	95
65) Chlorobenzene	10.061	112	113932	18.968	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.147	131	41512	20.626	ug/l	99
67) Ethyl Benzene	10.177	91	193315	19.166	ug/l	99
68) m/p-Xylenes	10.287	106	148280	39.033	ug/l	98
69) o-Xylene	10.622	106	71433	19.219	ug/l	99
70) Styrene	10.640	104	122558	19.657	ug/l	100
71) Bromoform	10.781	173	35372	19.050	ug/l #	100
73) Isopropylbenzene	10.945	105	188957	20.032	ug/l	98
74) N-amyl acetate	10.823	43	87436	18.461	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.195	83	64648	19.434	ug/l	99
76) 1,2,3-Trichloropropane	11.219	75	58726m	18.758	ug/l	
77) Bromobenzene	11.177	156	48236	19.803	ug/l	97
78) n-propylbenzene	11.287	91	218792	19.726	ug/l	99
79) 2-Chlorotoluene	11.348	91	130854	19.524	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	151773	19.737	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.000	75	19852	18.137	ug/l #	84
82) 4-Chlorotoluene	11.439	91	154443	19.645	ug/l	99
83) tert-Butylbenzene	11.695	119	156409	19.665	ug/l	97
84) 1,2,4-Trimethylbenzene	11.732	105	155265	20.012	ug/l	99
85) sec-Butylbenzene	11.872	105	195955	19.934	ug/l	99
86) p-Isopropyltoluene	11.988	119	164200	19.868	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	86661	18.983	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	88305	18.727	ug/l	98
89) n-Butylbenzene	12.311	91	150223	19.376	ug/l	100
90) Hexachloroethane	12.518	117	30433	19.328	ug/l	99
91) 1,2-Dichlorobenzene	12.317	146	81826	18.671	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.926	75	14410	18.641	ug/l	96
93) 1,2,4-Trichlorobenzene	13.567	180	53507	18.373	ug/l	99
94) Hexachlorobutadiene	13.707	225	22283	18.304	ug/l	98
95) Naphthalene	13.756	128	174740	18.663	ug/l	99
96) 1,2,3-Trichlorobenzene	13.938	180	51058	18.849	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048644.D
Acq On : 21 Nov 2025 10:26
Operator : JC/MD
Sample : VX1121WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/24/2025
Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:08:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048644.D
Acq On : 21 Nov 2025 10:26
Operator : JC/MD
Sample : VX1121WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 11/24/2025
Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:08:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration



Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: VX1121WBSD01
 Lab Sample ID: VX1121WBSD01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	19.4		1	0.22	1.00	ug/L	11/21/25 10:55	VX112125
74-87-3	Chloromethane	16.1		1	0.32	1.00	ug/L	11/21/25 10:55	VX112125
75-01-4	Vinyl Chloride	16.8		1	0.26	1.00	ug/L	11/21/25 10:55	VX112125
141-78-6	Ethyl Acetate	15.7		1	0.31	1.00	ug/L	11/21/25 10:55	VX112125
74-83-9	Bromomethane	18.8		1	1.40	5.00	ug/L	11/21/25 10:55	VX112125
75-00-3	Chloroethane	16.8		1	0.47	1.00	ug/L	11/21/25 10:55	VX112125
75-69-4	Trichlorofluoromethane	17.4		1	0.33	1.00	ug/L	11/21/25 10:55	VX112125
76-13-1	1,1,2-Trichlorotrifluoroethane	19.1		1	0.25	1.00	ug/L	11/21/25 10:55	VX112125
75-65-0	Tert butyl alcohol	91.7		1	5.50	25.0	ug/L	11/21/25 10:55	VX112125
75-35-4	1,1-Dichloroethene	17.3		1	0.23	1.00	ug/L	11/21/25 10:55	VX112125
107-02-8	Acrolein	83.9		1	7.10	25.0	ug/L	11/21/25 10:55	VX112125
107-13-1	Acrylonitrile	89.0		1	0.83	5.00	ug/L	11/21/25 10:55	VX112125
67-64-1	Acetone	85.4		1	1.50	5.00	ug/L	11/21/25 10:55	VX112125
75-15-0	Carbon Disulfide	14.0		1	0.21	1.00	ug/L	11/21/25 10:55	VX112125
1634-04-4	Methyl tert-butyl Ether	18.2		1	0.16	1.00	ug/L	11/21/25 10:55	VX112125
79-20-9	Methyl Acetate	18.0		1	0.27	1.00	ug/L	11/21/25 10:55	VX112125
75-09-2	Methylene Chloride	17.1		1	0.28	1.00	ug/L	11/21/25 10:55	VX112125
156-60-5	trans-1,2-Dichloroethene	16.9		1	0.23	1.00	ug/L	11/21/25 10:55	VX112125
108-05-4	Vinyl Acetate	95.4		1	0.66	5.00	ug/L	11/21/25 10:55	VX112125
75-34-3	1,1-Dichloroethane	17.5		1	0.23	1.00	ug/L	11/21/25 10:55	VX112125
110-82-7	Cyclohexane	16.6		1	1.50	5.00	ug/L	11/21/25 10:55	VX112125
78-93-3	2-Butanone	87.1		1	0.98	5.00	ug/L	11/21/25 10:55	VX112125
56-23-5	Carbon Tetrachloride	16.8		1	0.25	1.00	ug/L	11/21/25 10:55	VX112125
594-20-7	2,2-Dichloropropane	18.3		1	0.21	1.00	ug/L	11/21/25 10:55	VX112125
156-59-2	cis-1,2-Dichloroethene	18.2		1	0.19	1.00	ug/L	11/21/25 10:55	VX112125
74-97-5	Bromochloromethane	14.3		1	0.22	1.00	ug/L	11/21/25 10:55	VX112125
67-66-3	Chloroform	18.0		1	0.25	1.00	ug/L	11/21/25 10:55	VX112125
71-55-6	1,1,1-Trichloroethane	17.9		1	0.20	1.00	ug/L	11/21/25 10:55	VX112125
108-87-2	Methylcyclohexane	18.4		1	0.16	1.00	ug/L	11/21/25 10:55	VX112125
563-58-6	1,1-Dichloropropene	15.9		1	0.15	1.00	ug/L	11/21/25 10:55	VX112125
71-43-2	Benzene	16.9		1	0.15	1.00	ug/L	11/21/25 10:55	VX112125
107-06-2	1,2-Dichloroethane	17.1		1	0.22	1.00	ug/L	11/21/25 10:55	VX112125
79-01-6	Trichloroethene	18.7		1	0.090	1.00	ug/L	11/21/25 10:55	VX112125
78-87-5	1,2-Dichloropropane	20.1		1	0.20	1.00	ug/L	11/21/25 10:55	VX112125
74-95-3	Dibromomethane	20.9		1	0.25	1.00	ug/L	11/21/25 10:55	VX112125
75-27-4	Bromodichloromethane	20.1		1	0.22	1.00	ug/L	11/21/25 10:55	VX112125
108-10-1	4-Methyl-2-Pentanone	99.6		1	0.68	5.00	ug/L	11/21/25 10:55	VX112125
108-88-3	Toluene	20.6		1	0.14	1.00	ug/L	11/21/25 10:55	VX112125
10061-02-6	t-1,3-Dichloropropene	19.7		1	0.17	1.00	ug/L	11/21/25 10:55	VX112125
10061-01-5	cis-1,3-Dichloropropene	20.1		1	0.16	1.00	ug/L	11/21/25 10:55	VX112125
79-00-5	1,1,2-Trichloroethane	22.2		1	0.21	1.00	ug/L	11/21/25 10:55	VX112125
142-28-9	1,3-Dichloropropane	20.5		1	0.19	1.00	ug/L	11/21/25 10:55	VX112125
110-75-8	2-Chloroethyl Vinyl ether	85.3		1	0.30	5.00	ug/L	11/21/25 10:55	VX112125

Report of Analysis

Client: Chemtech Consulting Group
 Project: Weekly Storage Blanks
 Client Sample ID: VX1121WBSD01
 Lab Sample ID: VX1121WBSD01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 mL

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3702
 Matrix: Water
 % Solid: 0
 Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
591-78-6	2-Hexanone	97.9		1	0.89	5.00	ug/L	11/21/25 10:55	VX112125
124-48-1	Dibromochloromethane	22.2		1	0.18	1.00	ug/L	11/21/25 10:55	VX112125
106-93-4	1,2-Dibromoethane	21.2		1	0.15	1.00	ug/L	11/21/25 10:55	VX112125
127-18-4	Tetrachloroethene	19.4		1	0.23	1.00	ug/L	11/21/25 10:55	VX112125
108-90-7	Chlorobenzene	18.9		1	0.12	1.00	ug/L	11/21/25 10:55	VX112125
630-20-6	1,1,1,2-Tetrachloroethane	21.1		1	0.19	1.00	ug/L	11/21/25 10:55	VX112125
67-72-1	Hexachloroethane	19.4		1	0.32	1.00	ug/L	11/21/25 10:55	VX112125
100-41-4	Ethyl Benzene	18.6		1	0.13	1.00	ug/L	11/21/25 10:55	VX112125
179601-23-1	m/p-Xylenes	38.9		1	0.24	2.00	ug/L	11/21/25 10:55	VX112125
95-47-6	o-Xylene	19.0		1	0.12	1.00	ug/L	11/21/25 10:55	VX112125
100-42-5	Styrene	19.5		1	0.15	1.00	ug/L	11/21/25 10:55	VX112125
75-25-2	Bromoform	20.9		1	0.19	1.00	ug/L	11/21/25 10:55	VX112125
98-82-8	Isopropylbenzene	19.2		1	0.12	1.00	ug/L	11/21/25 10:55	VX112125
79-34-5	1,1,2,2-Tetrachloroethane	19.0		1	0.26	1.00	ug/L	11/21/25 10:55	VX112125
96-18-4	1,2,3-Trichloropropane	18.1		1	0.35	1.00	ug/L	11/21/25 10:55	VX112125
108-86-1	Bromobenzene	19.2		1	0.24	1.00	ug/L	11/21/25 10:55	VX112125
103-65-1	n-propylbenzene	19.0		1	0.13	1.00	ug/L	11/21/25 10:55	VX112125
95-49-8	2-Chlorotoluene	18.6		1	0.14	1.00	ug/L	11/21/25 10:55	VX112125
108-67-8	1,3,5-Trimethylbenzene	19.3		1	0.15	1.00	ug/L	11/21/25 10:55	VX112125
106-43-4	4-Chlorotoluene	18.9		1	0.13	1.00	ug/L	11/21/25 10:55	VX112125
98-06-6	tert-Butylbenzene	19.7		1	0.14	1.00	ug/L	11/21/25 10:55	VX112125
95-63-6	1,2,4-Trimethylbenzene	19.3		1	0.14	1.00	ug/L	11/21/25 10:55	VX112125
135-98-8	sec-Butylbenzene	19.4		1	0.13	1.00	ug/L	11/21/25 10:55	VX112125
99-87-6	p-Isopropyltoluene	19.3		1	0.13	1.00	ug/L	11/21/25 10:55	VX112125
541-73-1	1,3-Dichlorobenzene	19.2		1	0.16	1.00	ug/L	11/21/25 10:55	VX112125
106-46-7	1,4-Dichlorobenzene	19.0		1	0.19	1.00	ug/L	11/21/25 10:55	VX112125
104-51-8	n-Butylbenzene	18.7		1	0.15	1.00	ug/L	11/21/25 10:55	VX112125
95-50-1	1,2-Dichlorobenzene	19.2		1	0.16	1.00	ug/L	11/21/25 10:55	VX112125
96-12-8	1,2-Dibromo-3-Chloropropane	19.7		1	0.53	1.00	ug/L	11/21/25 10:55	VX112125
120-82-1	1,2,4-Trichlorobenzene	19.3		1	0.20	1.00	ug/L	11/21/25 10:55	VX112125
87-68-3	Hexachlorobutadiene	18.7		1	0.30	1.00	ug/L	11/21/25 10:55	VX112125
91-20-3	Naphthalene	19.5		1	0.20	1.00	ug/L	11/21/25 10:55	VX112125
87-61-6	1,2,3-Trichlorobenzene	20.4		1	0.20	1.00	ug/L	11/21/25 10:55	VX112125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	45.6			74 - 125	91%	SPK: 50
1868-53-7	Dibromofluoromethane	46.0			75 - 124	92%	SPK: 50
2037-26-5	Toluene-d8	49.9			86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.8			77 - 121	102%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	169000
540-36-3	1,4-Difluorobenzene	276000
3114-55-4	Chlorobenzene-d5	240000
3855-82-1	1,4-Dichlorobenzene-d4	125000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Chemtech Consulting Group

Project: Weekly Storage Blanks

Client Sample ID: VX1121WBSD01

Lab Sample ID: VX1121WBSD01

Analytical Method: 8260D

Sample Wt/Vol: 5 mL

Level: LOW

Final Vol: 5000 uL

Date Collected:

Date Received:

SDG No.: Q3702

Matrix: Water

% Solid: 0

Test: VOC- Chemtech Full

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
------------	-----------	-------	------	----	-----	------------	-------	-----------	---------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048645.D
 Acq On : 21 Nov 2025 10:55
 Operator : JC/MD
 Sample : VX1121WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/24/2025
 Supervised By : Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:09:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	169398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	275788	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	240043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	107541	45.560	ug/l	-0.01
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.120%		
35) Dibromofluoromethane	5.367	113	96045	46.013	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.020%		
50) Toluene-d8	8.629	98	324656	49.870	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.740%		
62) 4-Bromofluorobenzene	11.061	95	123291	50.819	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.640%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	26654	19.418	ug/l	99
3) Chloromethane	1.301	50	29229	16.055	ug/l	98
4) Vinyl Chloride	1.380	62	34659	16.814	ug/l	99
5) Bromomethane	1.618	94	24408	18.752	ug/l	94
6) Chloroethane	1.697	64	22693	16.810	ug/l	98
7) Trichlorofluoromethane	1.892	101	56878	17.442	ug/l	99
8) Diethyl Ether	2.130	74	23464	17.925	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.331	101	36149	19.148	ug/l	96
10) Methyl Iodide	2.453	142	45511	16.289	ug/l	98
11) Tert butyl alcohol	2.928	59	41008	91.684	ug/l	98
12) 1,1-Dichloroethene	2.325	96	34082	17.283	ug/l	93
13) Acrolein	2.233	56	5431	83.922	ug/l	96
14) Allyl chloride	2.666	41	56116	15.284	ug/l	98
15) Acrylonitrile	3.050	53	116203	88.997	ug/l	99
16) Acetone	2.367	43	95465	85.441	ug/l	96
17) Carbon Disulfide	2.514	76	77678	13.998	ug/l	98
18) Methyl Acetate	2.697	43	55401	18.043	ug/l	99
19) Methyl tert-butyl Ether	3.105	73	127057	18.246	ug/l	99
20) Methylene Chloride	2.782	84	39210	17.069	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	35313	16.896	ug/l	93
22) Diisopropyl ether	3.745	45	138475	19.642	ug/l	97
23) Vinyl Acetate	3.709	43	589015	95.395	ug/l	98
24) 1,1-Dichloroethane	3.605	63	67830	17.535	ug/l	99
25) 2-Butanone	4.526	43	143921	87.073	ug/l	100
26) 2,2-Dichloropropane	4.459	77	60481	18.339	ug/l	100
27) cis-1,2-Dichloroethene	4.471	96	46477	18.193	ug/l	100
28) Bromochloromethane	4.879	49	27107	14.264	ug/l	89
29) Tetrahydrofuran	4.983	42	99693	84.567	ug/l	97
30) Chloroform	5.080	83	71869	17.996	ug/l	97
31) Cyclohexane	5.452	56	52750	16.643	ug/l	95
32) 1,1,1-Trichloroethane	5.361	97	62399	17.904	ug/l	99
36) 1,1-Dichloropropene	5.678	75	44408	15.888	ug/l	97
37) Ethyl Acetate	4.690	43	61811	15.710	ug/l	99
38) Carbon Tetrachloride	5.660	117	53440	16.796	ug/l	94
39) Methylcyclohexane	7.360	83	57646	18.368	ug/l	95
40) Benzene	6.019	78	146091	16.862	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048645.D
 Acq On : 21 Nov 2025 10:55
 Operator : JC/MD
 Sample : VX1121WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/24/2025
 Supervised By : Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:09:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.898	41	29470	17.032	ug/l	97
42) 1,2-Dichloroethane	6.068	62	50715	17.084	ug/l	100
43) Isopropyl Acetate	6.318	43	92634	16.749	ug/l	99
44) Trichloroethene	7.111	130	38716	18.695	ug/l	100
45) 1,2-Dichloropropane	7.409	63	40036	20.097	ug/l	97
46) Dibromomethane	7.562	93	30306	20.858	ug/l	99
47) Bromodichloromethane	7.806	83	58327	20.122	ug/l	98
48) Methyl methacrylate	7.671	41	44232	17.991	ug/l	98
49) 1,4-Dioxane	7.641	88	15481	413.246	ug/l	95
51) 4-Methyl-2-Pentanone	8.555	43	313436	99.561	ug/l	100
52) Toluene	8.702	92	93794	20.562	ug/l	98
53) t-1,3-Dichloropropene	8.964	75	58818	19.693	ug/l	96
54) cis-1,3-Dichloropropene	8.348	75	62109	20.082	ug/l	99
55) 1,1,2-Trichloroethane	9.135	97	40486	22.215	ug/l	98
56) Ethyl methacrylate	9.098	69	62232	20.525	ug/l	97
57) 1,3-Dichloropropane	9.293	76	65146	20.537	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.220	63	139017	85.280	ug/l	98
59) 2-Hexanone	9.415	43	229351	97.938	ug/l	98
60) Dibromochloromethane	9.506	129	49759	22.238	ug/l	100
61) 1,2-Dibromoethane	9.592	107	41941	21.248	ug/l	99
64) Tetrachloroethene	9.256	164	33886	19.436	ug/l	96
65) Chlorobenzene	10.061	112	107597	18.885	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.147	131	40216	21.066	ug/l	98
67) Ethyl Benzene	10.177	91	178050	18.610	ug/l	98
68) m/p-Xylenes	10.287	106	140244	38.921	ug/l	98
69) o-Xylene	10.622	106	66879	18.970	ug/l	99
70) Styrene	10.640	104	115287	19.494	ug/l	100
71) Bromoform	10.781	173	36814	20.903	ug/l #	99
73) Isopropylbenzene	10.945	105	173422	19.167	ug/l	100
74) N-amyl acetate	10.823	43	75172	16.546	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.195	83	60542	18.974	ug/l	99
76) 1,2,3-Trichloropropane	11.220	75	54393m	18.112	ug/l	
77) Bromobenzene	11.183	156	44955	19.241	ug/l	95
78) n-propylbenzene	11.287	91	201721	18.960	ug/l	98
79) 2-Chlorotoluene	11.348	91	119560	18.597	ug/l	97
80) 1,3,5-Trimethylbenzene	11.433	105	142282	19.289	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	18855	17.959	ug/l	93
82) 4-Chlorotoluene	11.433	91	142338	18.875	ug/l	98
83) tert-Butylbenzene	11.695	119	150189	19.686	ug/l	100
84) 1,2,4-Trimethylbenzene	11.732	105	143897	19.335	ug/l	98
85) sec-Butylbenzene	11.872	105	182609	19.366	ug/l	99
86) p-Isopropyltoluene	11.988	119	153078	19.310	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	83876	19.154	ug/l	100
88) 1,4-Dichlorobenzene	12.024	146	85954	19.003	ug/l	98
89) n-Butylbenzene	12.317	91	139136	18.709	ug/l	100
90) Hexachloroethane	12.518	117	29320	19.412	ug/l	100
91) 1,2-Dichlorobenzene	12.317	146	80748	19.208	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.927	75	14599	19.688	ug/l	98
93) 1,2,4-Trichlorobenzene	13.567	180	53948	19.312	ug/l	98
94) Hexachlorobutadiene	13.707	225	21878	18.735	ug/l	100
95) Naphthalene	13.756	128	174763	19.458	ug/l	99
96) 1,2,3-Trichlorobenzene	13.939	180	53089	20.432	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048645.D
Acq On : 21 Nov 2025 10:55
Operator : JC/MD
Sample : VX1121WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/24/2025
Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:09:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048645.D
Acq On : 21 Nov 2025 10:55
Operator : JC/MD
Sample : VX1121WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 11/24/2025
Supervised By :Mahesh Dadoda 11/24/2025

Quant Time: Nov 21 22:09:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration



Manual Integration Report

Sequence:	VX110725	Instrument	MSVOA_x
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VX048516.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:19 AM	sam	11/10/2025 10:11:26 AM	Peak Integrated by Software
VSTDICC001	VX048516.D	1,4-Dichlorobenzene	JOHN	11/10/2025 8:27:19 AM	sam	11/10/2025 10:11:26 AM	Peak Integrated by Software
VSTDICC001	VX048516.D	2-Chloroethyl Vinyl ether	JOHN	11/10/2025 8:27:19 AM	sam	11/10/2025 10:11:26 AM	Peak Integrated by Software
VSTDICC005	VX048517.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:24 AM	sam	11/10/2025 10:11:30 AM	Peak Integrated by Software
VSTDICC005	VX048517.D	2-Chloroethyl Vinyl ether	JOHN	11/10/2025 8:27:24 AM	sam	11/10/2025 10:11:30 AM	Peak Integrated by Software
VSTDICC020	VX048518.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:29 AM	sam	11/10/2025 10:11:35 AM	Peak Integrated by Software
VSTDICC020	VX048518.D	2-Chloroethyl Vinyl ether	JOHN	11/10/2025 8:27:29 AM	sam	11/10/2025 10:11:35 AM	Peak Integrated by Software
VSTDICC020	VX048518.D	Acrolein	JOHN	11/10/2025 8:27:29 AM	sam	11/10/2025 10:11:35 AM	Peak Integrated by Software
VSTDICC100	VX048520.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:38 AM	sam	11/10/2025 10:12:15 AM	Peak Integrated by Software
VSTDICC100	VX048520.D	Bromomethane	JOHN	11/10/2025 8:27:38 AM	sam	11/10/2025 10:12:15 AM	Peak Integrated by Software
VSTDICC150	VX048521.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:44 AM	sam	11/10/2025 10:11:39 AM	Peak Integrated by Software
VSTDICCC050	VX048524.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:53 AM	sam	11/10/2025 10:12:22 AM	Peak Integrated by Software
VSTDICV050	VX048525.D	1,2,3-Trichloropropane	JOHN	11/10/2025 8:27:58 AM	sam	11/10/2025 10:11:48 AM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VX110725

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

Vx112125

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX048641.D	1,2,3-Trichloropropane	JOHN	11/24/2025 8:18:56 AM	MMDadoda	11/24/2025 9:50:22 AM	Peak Integrated by Software
VX1121WBS01	VX048644.D	1,2,3-Trichloropropane	JOHN	11/24/2025 8:19:00 AM	MMDadoda	11/24/2025 9:50:25 AM	Peak Integrated by Software
VX1121WBSD0 1	VX048645.D	1,2,3-Trichloropropane	JOHN	11/24/2025 8:19:04 AM	MMDadoda	11/24/2025 9:50:28 AM	Peak Integrated by Software
VSTDCCC050	VX048656.D	1,2,3-Trichloropropane	JOHN	11/24/2025 8:19:09 AM	MMDadoda	11/24/2025 9:50:31 AM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX110725

Review By	John Carlone	Review On	11/10/2025 8:28:24 AM		
Supervise By	Amit Patel	Supervise On	11/10/2025 9:54:56 AM		
SubDirectory	VX110725	HP Acquire Method	MSVOA_X	HP Processing Method	82X110725W.M
STD. NAME		STD REF.#			
Tune/Reschk		VP136104			
Initial Calibration Stds		VP136097,VP136098,VP135099,VP136100,VP136101,VP135102			
CCC					
Internal Standard/PEM					
ICV/I.BLK		VP136103			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX048515.D	07 Nov 2025 13:06	JC/MD	Ok
2	VSTDIC001	VX048516.D	07 Nov 2025 13:35	JC/MD	Ok,M
3	VSTDIC005	VX048517.D	07 Nov 2025 13:56	JC/MD	Ok,M
4	VSTDIC020	VX048518.D	07 Nov 2025 14:16	JC/MD	Ok,M
5	VSTDIC050	VX048519.D	07 Nov 2025 14:37	JC/MD	Not Ok
6	VSTDIC100	VX048520.D	07 Nov 2025 14:58	JC/MD	Ok,M
7	VSTDIC150	VX048521.D	07 Nov 2025 15:18	JC/MD	Ok,M
8	VIBLK	VX048522.D	07 Nov 2025 15:39	JC/MD	Ok
9	VSTDIC005	VX048523.D	07 Nov 2025 16:08	JC/MD	Not Ok
10	VSTDIC050	VX048524.D	07 Nov 2025 16:29	JC/MD	Ok,M
11	VSTDICV050	VX048525.D	07 Nov 2025 16:50	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX112125

Review By	John Carlone	Review On	11/24/2025 8:21:25 AM
Supervise By	Amit Patel	Supervise On	11/24/2025 2:13:04 PM
SubDirectory	VX112125	HP Acquire Method	HP Processing Method 82X110725W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136361		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136362,VP136363		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX048640.D	21 Nov 2025 08:00	JC/MD	Ok
2	VSTDCCC050	VX048641.D	21 Nov 2025 09:08	JC/MD	Ok,M
3	VX1121MBL01	VX048642.D	21 Nov 2025 09:37	JC/MD	Ok
4	VX1121WBL01	VX048643.D	21 Nov 2025 09:58	JC/MD	Ok
5	VX1121WBS01	VX048644.D	21 Nov 2025 10:26	JC/MD	Ok,M
6	VX1121WBSD01	VX048645.D	21 Nov 2025 10:55	JC/MD	Ok,M
7	Q3685-01	VX048646.D	21 Nov 2025 11:15	JC/MD	Ok
8	Q3685-02	VX048647.D	21 Nov 2025 11:36	JC/MD	ReRun
9	PB170646TB	VX048648.D	21 Nov 2025 11:56	JC/MD	Ok
10	Q3678-02	VX048649.D	21 Nov 2025 12:17	JC/MD	Ok
11	IBLK	VX048650.D	21 Nov 2025 12:38	JC/MD	Ok
12	Q3702-01	VX048651.D	21 Nov 2025 13:04	JC/MD	Ok
13	Q3702-02	VX048652.D	21 Nov 2025 13:25	JC/MD	Ok
14	Q3702-03	VX048653.D	21 Nov 2025 13:46	JC/MD	ReRun
15	Q3702-04	VX048654.D	21 Nov 2025 14:07	JC/MD	Ok
16	Q3702-03	VX048655.D	21 Nov 2025 14:44	JC/MD	Ok
17	VSTDCCC050	VX048656.D	21 Nov 2025 15:14	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110725

Review By	John Carlone	Review On	11/10/2025 8:28:24 AM
Supervise By	Amit Patel	Supervise On	11/10/2025 9:54:56 AM
SubDirectory	VX110725	HP Acquire Method	MSVOA_X HP Processing Method 82X110725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136104 VP136097,VP136098,VP135099,VP136100,VP136101,VP135102
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136103

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048515.D	07 Nov 2025 13:06		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX048516.D	07 Nov 2025 13:35		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX048517.D	07 Nov 2025 13:56		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX048518.D	07 Nov 2025 14:16		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX048519.D	07 Nov 2025 14:37	spiked low	JC/MD	Not Ok
6	VSTDIC100	VSTDIC100	VX048520.D	07 Nov 2025 14:58		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX048521.D	07 Nov 2025 15:18		JC/MD	Ok,M
8	VIBLK	VIBLK	VX048522.D	07 Nov 2025 15:39		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX048523.D	07 Nov 2025 16:08	spike error	JC/MD	Not Ok
10	VSTDICCC050	VSTDICCC050	VX048524.D	07 Nov 2025 16:29		JC/MD	Ok,M
11	VSTDICV050	ICVVX110725	VX048525.D	07 Nov 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112125

Review By	John Carlone	Review On	11/24/2025 8:21:25 AM
Supervise By	Amit Patel	Supervise On	11/24/2025 2:13:04 PM
SubDirectory	VX112125	HP Acquire Method	HP Processing Method 82X110725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136361
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136362,VP136363

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048640.D	21 Nov 2025 08:00		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048641.D	21 Nov 2025 09:08	pH#Lot#V12668	JC/MD	Ok,M
3	VX1121MBL01	VX1121MBL01	VX048642.D	21 Nov 2025 09:37		JC/MD	Ok
4	VX1121WBL01	VX1121WBL01	VX048643.D	21 Nov 2025 09:58		JC/MD	Ok
5	VX1121WBS01	VX1121WBS01	VX048644.D	21 Nov 2025 10:26		JC/MD	Ok,M
6	VX1121WBSD01	VX1121WBSD01	VX048645.D	21 Nov 2025 10:55		JC/MD	Ok,M
7	Q3685-01	4197	VX048646.D	21 Nov 2025 11:15	vial A pH<2	JC/MD	Ok
8	Q3685-02	4198	VX048647.D	21 Nov 2025 11:36	vial A pH<2 Surrogate Fail	JC/MD	ReRun
9	PB170646TB	PB170646TB	VX048648.D	21 Nov 2025 11:56		JC/MD	Ok
10	Q3678-02	COMPOSITE-STONE	VX048649.D	21 Nov 2025 12:17	vial A pH#5.0	JC/MD	Ok
11	IBLK	IBLK	VX048650.D	21 Nov 2025 12:38		JC/MD	Ok
12	Q3702-01	STORAGE-BLANK-SAI	VX048651.D	21 Nov 2025 13:04	vial A pH<2	JC/MD	Ok
13	Q3702-02	STORAGE-BLANK-WA	VX048652.D	21 Nov 2025 13:25	vial A pH<2	JC/MD	Ok
14	Q3702-03	STORAGE-BLANK-WA	VX048653.D	21 Nov 2025 13:46	vial A pH<2 Surrogate fail	JC/MD	ReRun
15	Q3702-04	STORAGE-BLANK-SAI	VX048654.D	21 Nov 2025 14:07	vial A pH<2	JC/MD	Ok
16	Q3702-03	STORAGE-BLANK-WA	VX048655.D	21 Nov 2025 14:44	vial B pH<2	JC/MD	Ok
17	VSTDCCC050	VSTDCCC050EC	VX048656.D	21 Nov 2025 15:14		JC/MD	Ok,M

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/24/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:40
In Date: 11/21/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:47
Out Date: 11/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB138016

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3702-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
Q3702-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
Q3702-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
Q3702-08	SVOC-GPC2-BLANK	4	1.00	1.00	2.00	2.00	100.0	
Q3702-09	PEST-GPC2-BLANK	5	1.00	1.00	2.00	2.00	100.0	
Q3702-10	PEST -GPC2-BLANK-SPIKE	6	1.00	1.00	2.00	2.00	100.0	
Q3707-01	V933	14	1.13	10.78	11.91	8.05	64.2	
Q3707-02	V933-E2	15	1.14	11.19	12.33	8.44	65.2	
Q3708-01	112025	16	1.00	1.00	2.00	2.00	100.0	oil sample
Q3709-01	LI-1-112125	17	1.16	10.83	11.99	10.52	86.4	
Q3709-02	LI-1-112125-E2	18	1.14	10.72	11.86	10.53	87.6	
Q3710-01	RBR200028-RT5376-COMP	7	1.15	10.41	11.56	10.38	88.7	
Q3710-03	COMPRESSOR-EXCAVATION	8	1.12	10.93	12.05	10.88	89.3	
Q3710-05	WELD-SHOP-EXCAVATION	9	1.11	10.44	11.55	10.68	91.7	
Q3711-01	VNJ-222	10	1.14	10.40	11.54	10.33	88.4	
Q3711-03	72-12017	11	1.18	10.49	11.67	10.34	87.3	
Q3712-01	JC-03-11212025	13	1.11	10.84	11.95	10.82	89.6	
Q3712-02	JC-03-11212025-E2	12	1.13	10.69	11.82	10.64	89.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

11/21/25 13:30:16

Worklist Name : %1-112125

Worklist ID : 193292

Department : Wet-Chemistry

Date : 11-21-2025 11:36:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3702-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3702-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3702-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3702-08	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3702-09	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3702-10	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3707-01	V933	Solid	Percent Solids	Cool 4 deg C	CHEM02	D41	11/14/2025	Chemtech -SO
Q3707-02	V933-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3708-01	112025	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3709-01	LI-1-112125	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3709-02	LI-1-112125-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3710-01	RBR200028-RT5376-COMP	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3710-03	COMPRESSOR-EXCAVATION	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3710-05	WELD-SHOP-EXCAVATION	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3711-01	VNJ-222	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/21/2025	Chemtech -SO
Q3711-03	72-12017	Solid	Percent Solids	Cool 4 deg C	PSEG03	D11	11/21/2025	Chemtech -SO
Q3712-01	JC-03-11212025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/21/2025	Chemtech -SO
Q3712-02	JC-03-11212025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/21/2025	Chemtech -SO

11/21/25 16:45

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Raw Sample Relinquished by:

Raw Sample Relinquished by:

Date/Time 11/21/25

18:00

Raw Sample Received by:

Raw Sample Relinquished by:

Raw Sample Relinquished by:

Raw Sample Relinquished by:



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 788-9222
www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q3702

COC Number:

CLIENT INFORMATION			PROJECT INFORMATION			BILLING INFORMATION																																																																							
COMPANY: Alliance Technical Group, LLC - Newark			PROJECT NAME: Weekly Storage Blanks			BILL TO:						PO#																																																																	
ADDRESS: 284 Sheffield Street			PROJECT #:			ADDRESS:																																																																							
CITY Mountainside STATE: NJ ZIP: 07092			PROJECT MANAGER:			CITY:						STATE: ZIP:																																																																	
ATTENTION:			E-MAIL:			ATTENTION:						PHONE:																																																																	
PHONE:			FAX:			PHONE:						FAX:																																																																	
DATA TURNAROUND INFORMATION			DATA DELIVERABLE INFORMATION			ANALYSIS																																																																							
FAX: NA DAYS*			☐ RESEULTS ONLY ☐ USEPA CLP			<table border="1"><thead><tr><th>VOC</th><th>SVOC</th><th>PESTICIDES</th><th></th><th></th><th></th><th></th><th></th><th></th><th></th><th></th><th></th><th></th><th></th><th></th></tr><tr><th>1</th><th>2</th><th>3</th><th>4</th><th>5</th><th>6</th><th>7</th><th>8</th><th>9</th><th></th><th></th><th></th><th></th><th></th><th></th></tr></thead><tbody><tr><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr><tr><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr></tbody></table>												VOC	SVOC	PESTICIDES													1	2	3	4	5	6	7	8	9																																				
VOC	SVOC	PESTICIDES																																																																											
1	2	3	4	5	6													7	8	9																																																									
HARD COPY: 10 DAYS*			☐ RESULTS + QC ☐ New York State ASP "B"																																																																										
EDD NA DAYS*			☐ New Jersey REDUCED ☐ New York State ASP "A"																																																																										
* TO BE APPROVED BY ALLIANCE			☐ New Jersey CLP ☐ Other																																																																										
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS			☐ EDD Format																																																																										
CHEMTECH SAMPLE ID		PROJECT SAMPLE IDENTIFICATION		SAMPLE MATRIX		SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles		PRESERVATIVES												COMMENTS																																																					
						COMP GRAB		DATE TIME				1 2 3 4 5 6 7 8 9												← Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other																																																					
1.		STORAGE-BLANK-SOIL-REF-6		AQ				X		2		X																																																																	
2.		STORAGE-BLANK-WATER-REF-5		AQ				X		2		X																																																																	
3.		STORAGE-BLANK-WATER-REF-4		AQ				X		2		X																																																																	
4.		STORAGE-BLANK-SAMPLE-MANAGEMENT		AQ				X		2		X																																																																	
5.		SVOC-GPC-BLANK		SO				X		1		X																																																																	
6.		PEST-GPC-BLANK		SO				X		1		X																																																																	
7.		PEST-GPC-BLANK-SPIKE		SO				X		1		X																																																																	
8.		SVOC-GPC2-BLANK		SO				X		1		X																																																																	
9.		PEST-GPC2-BLANK		SO				X		1		X																																																																	
10.		PEST -GPC2-BLANK-SPIKE		SO				X		1		X																																																																	
SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY																																																																													
RELINQUISHED BY SAMPLER		DATE/TIME		RECEIVED BY		Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp _____ MeOH extraction requires an additional 4oz. Jar for percent solid ☐ Ice in Cooler?: _____																																																																							
1. Sam		1/21/25		1. 7:00		Comments:																																																																							
RELINQUISHED BY		DATE/TIME		RECEIVED BY																																																																									
2.				2.																																																																									
RELINQUISHED BY		DATE/TIME		RECEIVED FOR LAB BY		SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight																																																																							
3.				3.		ALLIANCE: ☐ Picked Up ☐ Overnight																																																																							
Shipment Complete ☐ YES ☐ NO																																																																													
Page _____ of _____																																																																													
WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY																																																																													

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312