

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3702
 Lab File ID: VX048641.D Date Analyzed: 11/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:08
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	165802	5.53	262765	6.74	235899	10.04
UPPER LIMIT	331604	6.031	525530	7.239	471798	10.537
LOWER LIMIT	82901	5.031	131383	6.239	117950	9.537
EPA SAMPLE NO.						
STORAGE-BLANK-SAMPLE-MANAGEMENT	160862	5.53	334385	6.75	357734	10.04
STORAGE-BLANK-WATER-REF-5	170410	5.54	333480	6.75	361276	10.04
STORAGE-BLANK-WATER-REF-4	160298	5.53	322084	6.74	341432	10.04
STORAGE-BLANK-SAMPLE-MANAGEMENT	157000	5.54	308197	6.75	328677	10.04
VX1121WBL01	203389	5.54	383341	6.75	433179	10.04
VX1121WBS01	173428	5.53	281921	6.74	253067	10.04
VX1121WBSD01	169398	5.53	275788	6.74	240043	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.