



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ATG-GREENVILLE AEC

Project Name: AEC-2025-0013-CSC Wheeling IL

Project # N/A

Order ID # Q3703

Test Name: COD,Field pH,TSS

A. Number of Samples and Date of Receipt:

1 Water sample was received on 11/21/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: COD,Field pH,TSS. This data package contains results for COD,Field pH,TSS.

C. Analytical Techniques:

The analysis of TSS was based on method SM2540 D, The analysis of Field pH was based on method SM4500-H B and The analysis of COD was based on method SM5220 D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Sample Q3703-01 was analyzed straight Dilution for COD Parameter because the original samples were reading over range, only 10X has been reported.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____