

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3703

Project: AEC-2025-0013-CSC Wheeling IL

RunNo.: LB138041

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	52.196	50	104	95-105	11/25/2025
Sample ID: CCV2 COD	mg/L	51.184	50	102	95-105	11/25/2025
Sample ID: CCV3 COD	mg/L	51.184	50	102	95-105	11/25/2025