

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3704
Project:	AEC-2025-0013-CSC 3093 Carol Stream	Sample ID:	Q3676-01
Client ID:	OUTFALL-DSN-001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	73.4		26.9		50.0	1	93		11/25/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3704
Project:	AEC-2025-0013-CSC 3093 Carol Stream	Sample ID:	Q3676-01
Client ID:	OUTFALL-DSN-001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	75.5		26.9		50.0	1	97		11/25/2025