

Laboratory Control Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3704
Project:	AEC-2025-0013-CSC 3093 Carol Stream	Run No.:	LB138018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138018BS							
TSS	mg/L	550	591		107	1	90-110	11/21/2025

Laboratory Control Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3704
Project:	AEC-2025-0013-CSC 3093 Carol Stream	Run No.:	LB138041

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138041BS							
COD	mg/L	50	48.2		96	1	90-110	11/25/2025