

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB138033

	NA	NO	YES
1. Calibration Summary met criteria.	_____	_____	_____✓
2. ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓
4. Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
7. Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
8. Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓
9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓

ADDITIONAL COMMENTS:

MS/MSD fail for Q3711-03 (Sb,Cu,K,Ag,Tl) & Q3711-04 (Ba,Mg,Mn,Tl) due to matrix unknown chemical interferences at the time of digestion which shows low recovery for spike added.

Analyst

Date

QA REVIEW

Date