

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3900	4000	98	90 - 110	P	11/24/2025	13:37	LB138033
	Copper	993	1000	99	90 - 110	P	11/24/2025	13:37	LB138033
	Zinc	1970	2000	98	90 - 110	P	11/24/2025	13:37	LB138033

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	20.8	20.0	104	80 - 120	P	11/24/2025	13:41	LB138033
	Copper	22.0	20.0	110	80 - 120	P	11/24/2025	13:41	LB138033
	Zinc	42.7	40.0	107	80 - 120	P	11/24/2025	13:41	LB138033

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4900	5000	98	90 - 110	P	11/24/2025	14:11	LB138033
	Copper	1230	1250	98	90 - 110	P	11/24/2025	14:11	LB138033
	Zinc	2440	2500	98	90 - 110	P	11/24/2025	14:11	LB138033
CCV02	Arsenic	4910	5000	98	90 - 110	P	11/24/2025	15:59	LB138033
	Copper	1240	1250	99	90 - 110	P	11/24/2025	15:59	LB138033
	Zinc	2510	2500	100	90 - 110	P	11/24/2025	15:59	LB138033
CCV03	Arsenic	4830	5000	97	90 - 110	P	11/24/2025	16:49	LB138033
	Copper	1230	1250	98	90 - 110	P	11/24/2025	16:49	LB138033
	Zinc	2460	2500	99	90 - 110	P	11/24/2025	16:49	LB138033
CCV04	Arsenic	4960	5000	99	90 - 110	P	11/24/2025	17:41	LB138033
	Copper	1250	1250	100	90 - 110	P	11/24/2025	17:41	LB138033
	Zinc	2490	2500	100	90 - 110	P	11/24/2025	17:41	LB138033
CCV05	Arsenic	4800	5000	96	90 - 110	P	11/24/2025	18:50	LB138033
	Copper	1210	1250	97	90 - 110	P	11/24/2025	18:50	LB138033
	Zinc	2410	2500	96	90 - 110	P	11/24/2025	18:50	LB138033
CCV06	Arsenic	4910	5000	98	90 - 110	P	11/24/2025	19:04	LB138033
	Copper	1230	1250	98	90 - 110	P	11/24/2025	19:04	LB138033
	Zinc	2480	2500	99	90 - 110	P	11/24/2025	19:04	LB138033