



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	21.6	20.0	108	65 - 135	P	11/24/2025	13:49	LB138033
	Copper	22.7	20.0	114	65 - 135	P	11/24/2025	13:49	LB138033
	Zinc	43.2	40.0	108	65 - 135	P	11/24/2025	13:49	LB138033