



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	13:45	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	13:45	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	13:45	LB138033

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	14:15	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	14:15	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	14:15	LB138033
CCB02	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	16:03	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	16:03	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	16:03	LB138033
CCB03	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	16:53	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	16:53	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	16:53	LB138033
CCB04	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	17:45	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	17:45	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	17:45	LB138033
CCB05	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	18:54	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	18:54	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	18:54	LB138033
CCB06	Arsenic	5.12	+/-10	U	20.0	P	11/24/2025	19:08	LB138033
	Copper	4.60	+/-10	U	20.0	P	11/24/2025	19:08	LB138033
	Zinc	16.7	+/-20	U	40.0	P	11/24/2025	19:08	LB138033