

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	1.18			-20	20	11/24/2025	13:54	LB138033
	Copper	-3.61	2.0	180	-18	22	11/24/2025	13:54	LB138033
	Zinc	8.80			-40	40	11/24/2025	13:54	LB138033
ICSAB01	Arsenic	115	104	111	88.4	120	11/24/2025	13:58	LB138033
	Copper	483	511	94	434	588	11/24/2025	13:58	LB138033
	Zinc	1050	952	110	809	1095	11/24/2025	13:58	LB138033