

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Europastry</u>	level: <u>low</u>	sdg no.: <u>Q3705</u>
contract: <u>EURO03</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3705-01</u>	client id: <u>MH-11212025MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3705-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	384		10.0	U	400	96		P
Copper	ug/L	75 - 125	153		23.3		150	86		P
Zinc	ug/L	75 - 125	672		601		100	71		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Europastry</u>	level:	<u>low</u>	sdg no.:	<u>Q3705</u>
contract:	<u>EURO03</u>			lab code:	<u>ACE</u>
matrix:	<u>Water</u>	sample id:	<u>Q3705-01</u>	client id:	<u>MH-11212025MSD</u>
Percent Solids for Sample:	NA	Spiked ID:	Q3705-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	378		10.0	U	400	94		P
Copper	ug/L	75 - 125	150		23.3		150	85		P
Zinc	ug/L	75 - 125	630		601		100	29		P