

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Europastry</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3705</u>
Contract:	<u>EURO03</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3705-01</u>	Client ID:	<u>MH-11212025DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3705-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Copper	ug/L	20	23.3		22.3		4		P
Zinc	ug/L	20	601		582		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Europastry</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3705</u>
Contract:	<u>EURO03</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3705-01MS</u>	Client ID:	<u>MH-11212025MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3705-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	384		378		2		P
Copper	ug/L	20	153		150		2		P
Zinc	ug/L	20	672		630		6		P