

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Europastry

SDG No.: Q3705

Contract: EURO03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170709BS							
Arsenic	ug/L	400	403		101	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Zinc	ug/L	100	106		106	80 - 120	P