

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN112425

Review By	John Carlone	Review On	11/25/2025 9:36:43 AM
Supervise By	Maresh Dadoda	Supervise On	11/26/2025 1:52:19 PM
SubDirectory	VN112425	HP Acquire Method	MSVOA_N HP Processing Method 82N112425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP136378
Initial Calibration Stds	VP136444,VP136445,VP136446,VP136447,VP136448,VP136449
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136450
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088343.D	24 Nov 2025 08:17		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088344.D	24 Nov 2025 12:38		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088345.D	24 Nov 2025 13:11	COM.#74 Peak not detected in 1 and 5 PPB	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088346.D	24 Nov 2025 13:32		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088347.D	24 Nov 2025 13:53	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088348.D	24 Nov 2025 14:14		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088349.D	24 Nov 2025 14:35		JC\MD	Ok,M
8	IBLK	IBLK	VN088350.D	24 Nov 2025 14:56		JC\MD	Ok
9	VSTDICV050	ICVVN112425	VN088351.D	24 Nov 2025 15:18		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Semsettin Yesilyurt	Review On	12/4/2025 10:27:11 AM
Supervise By	John Carlone	Supervise On	12/4/2025 10:36:07 AM
SubDirectory	VN120125	HP Acquire Method	MSVOA_N HP Processing Method 82N112425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088369.D	01 Dec 2025 08:49		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088370.D	01 Dec 2025 09:58	pH#Lot#V12668	JC\MD	Ok,M
3	VN1201MBL01	VN1201MBL01	VN088371.D	01 Dec 2025 10:33		JC\MD	Ok
4	VN1201WBL01	VN1201WBL01	VN088372.D	01 Dec 2025 10:53		JC\MD	Ok
5	VN1201WBS01	VN1201WBS01	VN088373.D	01 Dec 2025 11:27		JC\MD	Ok,M
6	VN1201WBSD01	VN1201WBSD01	VN088374.D	01 Dec 2025 12:10		JC\MD	Ok,M
7	Q3715-02	MW6	VN088375.D	01 Dec 2025 12:30	vial A pH<2	JC\MD	Ok,M
8	Q3741-07	TB-1	VN088376.D	01 Dec 2025 12:51	vial A pH<2 TB	JC\MD	Ok
9	Q3716-01	MW5	VN088377.D	01 Dec 2025 13:11	vial A pH<2 Need 20X	JC\MD	Dilution
10	Q3716-03	MW4	VN088378.D	01 Dec 2025 13:32	Need Straight Run	JC\MD	Not Ok
11	Q3716-04	MW7	VN088379.D	01 Dec 2025 13:52	Need 200X	JC\MD	Dilution
12	Q3715-01	MW2	VN088380.D	01 Dec 2025 14:13	E flag of previous sample	JC\MD	ReRun
13	Q3716-02	MW3	VN088381.D	01 Dec 2025 14:33	vial A pH<2	JC\MD	Ok
14	Q3738-01	MW4	VN088382.D	01 Dec 2025 14:54	Run @ Lower Dilution	JC\MD	Not Ok
15	Q3738-03	Field	VN088383.D	01 Dec 2025 15:15	vial A pH<2 FB	JC\MD	Ok
16	Q3738-02	MW5	VN088384.D	01 Dec 2025 15:35	vial A pH<2	JC\MD	Ok,M
17	Q3745-01	STORAGE-BLANK-SO	VN088385.D	01 Dec 2025 15:56	vial A pH<2 Tics Contamination	JC\MD	ReRun
18	Q3745-02	STORAGE-BLANK-WA	VN088386.D	01 Dec 2025 16:16	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Semsettin Yesilyurt	Review On	12/4/2025 10:27:11 AM		
Supervise By	John Carlone	Supervise On	12/4/2025 10:36:07 AM		
SubDirectory	VN120125	HP Acquire Method	MSVOA_N	HP Processing Method	82N112425W.M
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136451			
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP136452,VP136453			

19	Q3745-03	STORAGE-BLANK-WA	VN088387.D	01 Dec 2025 16:37	vial A pH<2	JC\MD	Ok
20	Q3745-04	STORAGE-BLANK-SAM	VN088388.D	01 Dec 2025 16:58	vial A pH<2	JC\MD	Ok
21	Q3743-03	TB	VN088389.D	01 Dec 2025 17:18	vial A pH<2 TB	JC\MD	Ok
22	Q3743-04	TB	VN088390.D	01 Dec 2025 17:39	vial A pH<2 TB	JC\MD	Ok
23	Q3743-05	TB	VN088391.D	01 Dec 2025 18:00	SIM Method on login	JC\MD	Not Ok
24	Q3743-01	TW-1125-1	VN088392.D	01 Dec 2025 18:20	vial A pH<2	JC\MD	Ok
25	Q3743-02	TW-1125-2	VN088393.D	01 Dec 2025 18:41	vial A pH<2	JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN088394.D	01 Dec 2025 19:01		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110725

Review By	John Carlone	Review On	11/10/2025 8:28:24 AM
Supervise By	Amit Patel	Supervise On	11/10/2025 9:54:56 AM
SubDirectory	VX110725	HP Acquire Method	MSVOA_X HP Processing Method 82X110725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136104 VP136097,VP136098,VP135099,VP136100,VP136101,VP135102 VP136103

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048515.D	07 Nov 2025 13:06		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX048516.D	07 Nov 2025 13:35		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX048517.D	07 Nov 2025 13:56		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX048518.D	07 Nov 2025 14:16		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX048519.D	07 Nov 2025 14:37	spiked low	JC/MD	Not Ok
6	VSTDIC100	VSTDIC100	VX048520.D	07 Nov 2025 14:58		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX048521.D	07 Nov 2025 15:18		JC/MD	Ok,M
8	VIBLK	VIBLK	VX048522.D	07 Nov 2025 15:39		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX048523.D	07 Nov 2025 16:08	spike error	JC/MD	Not Ok
10	VSTDIC050	VSTDIC050	VX048524.D	07 Nov 2025 16:29		JC/MD	Ok,M
11	VSTDICV050	ICVVX110725	VX048525.D	07 Nov 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112525

Review By	John Carlone	Review On	11/26/2025 8:21:34 AM
Supervise By	Maresh Dadoda	Supervise On	11/26/2025 12:39:09 PM
SubDirectory	VX112525	HP Acquire Method	HP Processing Method 82X110725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136388
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136389,VP136390

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048671.D	25 Nov 2025 08:48		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048672.D	25 Nov 2025 09:48	pH#Lot#V12668	JC/MD	Ok,M
3	VX1125MBL01	VX1125MBL01	VX048673.D	25 Nov 2025 10:17		JC/MD	Ok
4	VX1125WBL01	VX1125WBL01	VX048674.D	25 Nov 2025 10:38		JC/MD	Ok
5	VX1125WBS01	VX1125WBS01	VX048675.D	25 Nov 2025 11:05		JC/MD	Ok,M
6	VX1125WBSD01	VX1125WBSD01	VX048676.D	25 Nov 2025 11:32		JC/MD	Ok,M
7	Q3715-02	MW6	VX048677.D	25 Nov 2025 12:19	Run @ Lower Dilution	JC/MD	Not Ok
8	Q3715-01	MW2	VX048678.D	25 Nov 2025 12:39	vial A pH<2	JC/MD	Ok
9	Q3715-03	MW7	VX048679.D	25 Nov 2025 13:00	vial A pH<2	JC/MD	Ok
10	Q3716-01DL	MW5DL	VX048680.D	25 Nov 2025 13:21	Not match exactly please check	JC/MD	Not Ok
11	Q3716-02	MW3	VX048681.D	25 Nov 2025 13:41	Surrogate Fail;BS,BSD failed high for comp. #52;Run @ Lower Dilution	JC/MD	Not Ok
12	Q3716-03	MW4	VX048682.D	25 Nov 2025 14:02	BS,BSD failed high for comp. #52;Run @ Lower Dilution	JC/MD	Not Ok
13	Q3716-04	MW7	VX048683.D	25 Nov 2025 14:23	vial A pH<2 BS,BSD failed high for comp. #52;Need 200X	JC/MD	Not Ok
14	VSTDCCC050	VSTDCCC050EC	VX048684.D	25 Nov 2025 14:43		JC/MD	Ok,M

M : Manual Integration