

Surrogate Summary

SDG No.: **Q3715**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3715-01	MW2	1,2-Dichloroethane-d4	50	52.6	105		70 (74)	130 (125)
		Dibromofluoromethane	50	47.4	95		70 (75)	130 (124)
		Toluene-d8	50	53.5	107		70 (86)	130 (113)
		4-Bromofluorobenzene	50	63.1	126		70 (77)	130 (121)
Q3715-02	MW6	1,2-Dichloroethane-d4	50	52.1	104		70 (74)	130 (125)
		Dibromofluoromethane	50	49.3	99		70 (75)	130 (124)
		Toluene-d8	50	49.9	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.0	110		70 (77)	130 (121)
Q3715-03	MW7	1,2-Dichloroethane-d4	50	51.4	103		70 (74)	130 (125)
		Dibromofluoromethane	50	47.1	94		70 (75)	130 (124)
		Toluene-d8	50	55.2	110		70 (86)	130 (113)
		4-Bromofluorobenzene	50	60.0	120		70 (77)	130 (121)
VN1201WBL01	VN1201WBL01	1,2-Dichloroethane-d4	50	57.8	116		70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104		70 (75)	130 (124)
		Toluene-d8	50	50.0	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105		70 (77)	130 (121)
VN1201WBS01	VN1201WBS01	1,2-Dichloroethane-d4	50	56.0	112		70 (74)	130 (125)
		Dibromofluoromethane	50	55.1	110		70 (75)	130 (124)
		Toluene-d8	50	54.9	110		70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.3	107		70 (77)	130 (121)
VN1201WBSD0	VN1201WBSD01	1,2-Dichloroethane-d4	50	52.2	104		70 (74)	130 (125)
		Dibromofluoromethane	50	52.3	105		70 (75)	130 (124)
		Toluene-d8	50	53.2	106		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102		70 (77)	130 (121)
VX1125WBL01	VX1125WBL01	1,2-Dichloroethane-d4	50	56.6	113		70 (74)	130 (125)
		Dibromofluoromethane	50	46.3	93		70 (75)	130 (124)
		Toluene-d8	50	48.8	98		70 (86)	130 (113)
		4-Bromofluorobenzene	50	57.4	115		70 (77)	130 (121)
VX1125WBS01	VX1125WBS01	1,2-Dichloroethane-d4	50	51.2	102		70 (74)	130 (125)
		Dibromofluoromethane	50	45.9	92		70 (75)	130 (124)
		Toluene-d8	50	50.1	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98		70 (77)	130 (121)

