

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3715
 Lab File ID: VX048672.D Date Analyzed: 11/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:48
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	150094	5.53	259657	6.74	217303	10.04
UPPER LIMIT	300188	6.025	519314	7.239	434606	10.537
LOWER LIMIT	75047	5.025	129829	6.239	108652	9.537
EPA SAMPLE NO.						
MW2	172027	5.54	315243	6.75	370755	10.04
MW7	166997	5.54	305521	6.75	379381	10.04
VX1125WBL01	175748	5.54	347104	6.75	371931	10.04
VX1125WBS01	152887	5.53	255033	6.75	234446	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.