

## Initial and Continuing Calibration Verification

**Client:** G Environmental

**SDG No.:** Q3716

**Project:** Adoni

**RunNo.:** LB138034

| Analyte                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------|-------|--------|------------|------------|------------------------|---------------|
| <b>Sample ID: ICV1</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 9.7    | 10         | 97         | 90-110                 | 11/03/2025    |
| Chloride               | mg/L  | 2.9    | 3          | 97         | 90-110                 | 11/03/2025    |
| Fluoride               | mg/L  | 2      | 2          | 100        | 90-110                 | 11/03/2025    |
| Nitrite                | mg/L  | 2.9    | 3          | 97         | 90-110                 | 11/03/2025    |
| Nitrate                | mg/L  | 2.4    | 2.5        | 96         | 90-110                 | 11/03/2025    |
| Sulfate                | mg/L  | 14.5   | 15         | 97         | 90-110                 | 11/03/2025    |
| Orthophosphate as P    | mg/L  | 5.2    | 5          | 104        | 90-110                 | 11/03/2025    |
| <b>Sample ID: CCV1</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.4   | 10         | 104        | 90-110                 | 11/25/2025    |
| Chloride               | mg/L  | 3.1    | 3          | 103        | 90-110                 | 11/25/2025    |
| Fluoride               | mg/L  | 2.1    | 2          | 105        | 90-110                 | 11/25/2025    |
| Nitrite                | mg/L  | 3.1    | 3          | 103        | 90-110                 | 11/25/2025    |
| Nitrate                | mg/L  | 2.5    | 2.5        | 100        | 90-110                 | 11/25/2025    |
| Sulfate                | mg/L  | 15.5   | 15         | 103        | 90-110                 | 11/25/2025    |
| Orthophosphate as P    | mg/L  | 4.6    | 5          | 92         | 90-110                 | 11/25/2025    |
| <b>Sample ID: CCV2</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.3   | 10         | 103        | 90-110                 | 11/25/2025    |
| Chloride               | mg/L  | 3.1    | 3          | 103        | 90-110                 | 11/25/2025    |
| Fluoride               | mg/L  | 2.1    | 2          | 105        | 90-110                 | 11/25/2025    |
| Nitrite                | mg/L  | 3.2    | 3          | 107        | 90-110                 | 11/25/2025    |
| Nitrate                | mg/L  | 2.5    | 2.5        | 100        | 90-110                 | 11/25/2025    |
| Sulfate                | mg/L  | 15.4   | 15         | 103        | 90-110                 | 11/25/2025    |
| Orthophosphate as P    | mg/L  | 5      | 5          | 100        | 90-110                 | 11/25/2025    |

## Initial and Continuing Calibration Verification

**Client:** G Environmental

**SDG No.:** Q3716

**Project:** Adoni

**RunNo.:** LB138041

| Analyte           |             | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-------------------|-------------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>COD | <b>ICV</b>  | mg/L  | 50.173 | 50         | 100           | 95-105                    | 09/08/2025       |
| Sample ID:<br>COD | <b>CCV1</b> | mg/L  | 52.196 | 50         | 104           | 95-105                    | 11/25/2025       |
| Sample ID:<br>COD | <b>CCV2</b> | mg/L  | 51.184 | 50         | 102           | 95-105                    | 11/25/2025       |
| Sample ID:<br>COD | <b>CCV3</b> | mg/L  | 51.184 | 50         | 102           | 95-105                    | 11/25/2025       |