

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q3716
Project:	Adoni	Sample ID:	Q3676-01
Client ID:	OUTFALL-DSN-001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	73.4		26.9		50.0	1	93		11/25/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q3716
Project:	Adoni	Sample ID:	Q3676-01
Client ID:	OUTFALL-DSN-001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	75.5		26.9		50.0	1	97		11/25/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q3716
Project:	Adoni	Sample ID:	Q3716-02
Client ID:	MW3MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.0	OR	0.37	U	10	1	100		11/25/2025
Chloride	mg/L	80-120	48.9		47.6	OR	3	1	43	*	11/25/2025
Fluoride	mg/L	80-120	2.90		0.75		2	1	108		11/25/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		11/25/2025
Nitrate	mg/L	80-120	3.00		0.62		2.5	1	95		11/25/2025
Sulfate	mg/L	80-120	28.2		14.0		15	1	95		11/25/2025
Orthophosphate as P	mg/L	80-120	4.90		0.34	U	5	1	98		11/25/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q3716
Project:	Adoni	Sample ID:	Q3716-02
Client ID:	MW3MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	9.60	OR	0.37	U	10	1	96		11/25/2025
Chloride	mg/L	80-120	48.9		47.6	OR	3	1	43	*	11/25/2025
Fluoride	mg/L	80-120	2.60		0.75		2	1	92		11/25/2025
Nitrite	mg/L	80-120	2.90		0.074	U	3	1	97		11/25/2025
Nitrate	mg/L	80-120	2.90		0.62		2.5	1	91		11/25/2025
Sulfate	mg/L	80-120	27.7		14.0		15	1	91		11/25/2025
Orthophosphate as P	mg/L	80-120	5.00		0.34	U	5	1	100		11/25/2025