

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q3716
Project:	Adoni	Sample ID:	Q3676-01
Client ID:	OUTFALL-DSN-001DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	26.9		25.9		1	3.79		11/25/2025

Duplicate Sample Summary

Client: G Environmental Project: Adoni Client ID: OUTFALL-DSN-001MSD	SDG No.: Q3716 Sample ID: Q3676-01 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	73.4		75.5		1	2.82		11/25/2025

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q3716
Project:	Adoni	Sample ID:	Q3716-02
Client ID:	MW3DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	0.20	U	0.20	U	1	0		11/26/2025

Duplicate Sample Summary

Client: G Environmental Project: Adoni Client ID: MW3MSD	SDG No.: Q3716 Sample ID: Q3716-02 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/L	+/-20	48.9	OR	48.9	OR	1	0		11/25/2025
Fluoride	mg/L	+/-20	2.90		2.60		1	11		11/25/2025
Orthophosphate as P	mg/L	+/-20	4.90		5.00		1	2		11/25/2025
Sulfate	mg/L	+/-20	28.2		27.7		1	2		11/25/2025
Nitrate	mg/L	+/-20	3.00		2.90		1	3		11/25/2025
Nitrite	mg/L	+/-20	3.00		2.90		1	3		11/25/2025
Bromide	mg/L	+/-20	10.0		9.60		1	4		11/25/2025